

Name: Suzhou Saibao Calibration Technology Co., Ltd.

Address: No.4, Baotong Road, Wuzhong Economic Development Zone, Suzhou, Jiangsu, China

Registration No. CNAS L2336

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2026-05-21 Expiry Date: 2028-12-23

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
Geometry measuring instruments							
1	*Linear Displacement Sensors	Length	C.S. for Linear Displacement Sensors JJF 1305	(0.1~10)mm	$U=2 \mu m$		
				(10~1000)mm	$U=0.02\%FS$		
				(1000~2000)mm	$U=0.04\%FS$		
2	Tape	Length	V.R. of Steel Measuring Tapes JJG 4	(0~100)m	$U=0.2mm+2 \times 10^{-5}L$	except for Steel sounding tape measure.	
3	Fiber Tapes And Measuring Ropes	Length	V.R. of Fiber Tapes And Measuring Ropes JJG 5	(0~100)m	$U=0.2mm+1 \times 10^{-4}L$		
4	π ruler	Length	π - ruler calibration specification JJF 1423	(9~5000)mm	$U=0.02mm+4 \times 10^{-6}L$		
5	Steel Rule	Length	Verification Regulation of Steel Rule JJG 1	(0~600)mm	$U=0.06mm$		

No. CNAS L2336

第 1 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(600~1500)mm	$U=0.10\text{mm}$		
				(1500~2000)mm	$U=0.15\text{mm}$		
6	*Capacitive grating digital display scale	length	Calibration specification for capacitive grating digital display scale JJF 1280	(0~500)mm	$U=0.01\text{mm}$		
				(500~1000)mm	$U=0.02\text{mm}$		
7	*Cable Length Meter	Length	V.R. of Cable Length Meter JJG 987	(10~1000) m	$U_{\text{rel}}=0.18\%$		
8	*Tool Microscope	Length	V.R. of Universal Measuring Microscopes and Makers Microscopes JJG 56	(0~200)mm	$U=0.5 \mu\text{m}+2 \times 10^{-6}L$		
9	*Reading and measuring microscope	Length	V.R. of Measuring Microscope JJG 571	measuring microscope: (0~50)mm	$U=2 \mu\text{m}$		
		Length		Reading microscope: (0~6)mm	$U=2 \mu\text{m}$		
10	*Metallurgical Microscope	Length	C.S. for Biological Microscopes JJF 1402	(0~10) mm	$U=1.2 \mu\text{m}$		
				(4×~125×)	$U_{\text{rel}}=1.2\%$		
11	*Imaging Probe Measuring Machines	Length	C.S. for Imaging Probe Measuring Machines JJF 1318	(0~1000)mm	$U=0.5 \mu\text{m}+2 \times 10^{-6}L$		
12	*Projector	Length	C.S. for Projectors JJF 1093	(0~200)mm	$U=1.1 \mu\text{m}+5 \times 10^{-6}L$		
13	*Level Rule	Length	Calibration Specification for projectors Detecting The Notch of Tehe Sample JJF(Zhe) 1133	(0~100)mm	$U=6 \mu\text{m}$		
				50×	$U_{\text{rel}}=0.3\%$		
		Angle		(43~47)°	$U=2'$		
14	*Metallurgical Microscope	Length	Calibration Specification for Metallurgical Microscopes	(0~10) mm	$U=1.2 \mu\text{m}$		



No. CNAS L2336

第 2 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			JJF 1914	4×~125×	$U_{rel}=1.2\%$		
15	*Stereomicroscope	Length	C.S. for Stereomicroscope JJF(Jin) 87	(0~10) mm	$U=3 \mu m$		
		Magnification		Total Magnification (5~100) ×	$U_{rel}=1.2\%$		
16	Gauge Block	Length	V.R. of Gauge Blocks JJG 146	(0.5~1000)mm	$U=0.14\mu m+1.4\times 10^{-6}Ln$		
17	Internal Micrometers	Length	V.R. of Internal Micrometers JJG 22	(0~2000)mm	$U=1.5 \mu m+6\times 10^{-6}L$		
18	*Micrometer for measuring internal dimension	Length	C.S. for Micrometers for Measuring Inside Dimension JJF 1411	Internal: (5~200)mm	$U=3 \mu m$		
				three grasp: (6~200) mm	$U=2 \mu m$		
19	*Micrometers	Length	V.R. of Micrometer JJG 21	(0~100)mm	$U=1.3 \mu m+3\times 10^{-6}L$		
				digital: (0~100)mm	$U=0.7 \mu m+4\times 10^{-6}L$		
				(100~500)mm	$U=1.5 \mu m+5\times 10^{-6}L$		
				digital: (100~500)mm	$U=0.8 \mu m+4\times 10^{-6}L$		
				Measuring rod: (25~500)mm	$U=0.5 \mu m+6\times 10^{-6}L$		
20	*Outside micrometer of large size	Length	C.S. for Large Dimension Outside Micrometers JJF 1088	(500~1000)mm	$U=1.5 \mu m+6\times 10^{-6}L$		
				Measuring rod: (500~1000)mm	$U=0.5 \mu m+6\times 10^{-6}L$		
21	*Pointed outside micrometer	length	Calibration specification for pointed outside micrometers JJF (Zhe) 1045	(0~150)mm	$U=2 \mu m$		
22	*Wafer micrometer	length	Calibration specification for micrometers JJF (Zhe) 1090	(0~200)mm	$U=2 \mu m$		

No. CNAS L2336


第 3 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
23	*Micrometer with small probe	length	Calibration specification for micrometer with small probe JJF(Zhe)1131	(0~100)mm	$U=2 \mu m$		
24	*Micrometer with round probe	Length	Calibration specification for micrometers with circular probe JJF(Zhe)1132	(0~200)mm	$U=2 \mu m$		
25	*Depth Micrometer	Length	V.R. of Depth Micrometers JIG 24	(0~300)mm	$U=1.2 \mu m+5 \times 10^{-6}L$		
26	Micrometers with Dial Comparater and Indicating Snap Gauge	Length	V.R. of Micrometers with Dial Comparater and Indicating Snap Gauge JIG 26	(0~200)mm	$U=1.0 \mu m$		
27	*Micrometers with Gauge	Length	V.R. of Micrometers with Gauge JIG 427	(0~100)mm	$U=0.7 \mu m+4 \times 10^{-6}L$		
28	*Indicator table	Length	V. R. of Dial Gauges JIG 34	dial reading in 0.001mm: (0~5)mm	$U=1.8 \mu m+3 \times 10^{-4}L$		
				digital reading in 0.001mm: (0~1)mm	$U=1.4 \mu m$		
				digital reading in 0.001mm: (1~10)mm	$U=2.4 \mu m$		
				dial reading in 0.01mm: (0~10)mm	$U=5 \mu m$		
				digital reading in 0.01mm: (0~10)mm	$U=7 \mu m$		
				dial Large range dial indicator:(10~50)mm	$U=8 \mu m$		
				digital Large range dial indicator:(10~50)mm	$U=0.01mm$		
29	*Dial Test Indicator	Length	V.R. of Dial Test Indicator JIG 35	reading in 0.01mm: (0~1)mm	$U=2.7 \mu m$		



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				reading in 0.001mm: (0~0.4)mm	$U=1.2\mu\text{m}$		
30	*Bore Dial Indicators	Length	C.S. for Bore Dial Indicators JJF 1102	Inner diameter dial gauge:(6~450)mm	$U=3\mu\text{m}$		
				Inside dial indicator:(6~450)mm	$U=1.9\mu\text{m}$		
31	*Depth Dial Gauge	Length	V.R.of Depth Dial Gauge JJG 830	(0~10) mm	$U=1\mu\text{m}$		
				(10~100)mm	$U=2\mu\text{m}$		
32	*Current Calipers	Length	V.R. of Current Calipers JJG 30	(0~500)mm	$U=0.01\text{mm}$		
				(500~1000)mm	$U=0.02\text{mm}$		
				(1000~1500)mm	$U=0.03\text{mm}$		
				(1500~2000)mm	$U=0.04\text{mm}$		
33	*Height Gauge	Length	V. R. of Feeler Gauge JJG 31	(0~500)mm	$U=0.01\text{mm}$		
				(500~1000)mm	$U=0.02\text{mm}$		
34	*Dial Snap Gauges	Length	C.S.for Dial Snap Gauges JJF 1253	(5~100)mm	$U=9\mu\text{m}$		
35	*Thickness Gauges	Length	C.S.for Thickness Gauges JJF 1255	(0~50)mm	$U=2\mu\text{m}$		
36	*Tester for Dial Indicator Gauges	Length	V.R. of Tester for Dial Indicator Gauges JJG 201	(0~50)mm	$U=0.2\mu\text{m}+2\times 10^{-5}L$		
37	*Micrometers with Prismatically Arranged Measuring Faces	Length	V.R.of Micrometers with Prismatically Arranged Measuring Faces JJG 182	(1~50)mm	$U=1.4\mu\text{m}$		




No. CNAS L2336

第 5 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
38	*Rubber, plastic film thickness gauge	Length	C.S. for rubber, plastic film thickness gauge JJF 1488	(0~30) mm	$U=3 \mu m$		
39	Ultrasonic flaw detection test block	Length	C.S. for ultrasonic flaw detection block JJF 1487	(0~300) mm	$U=4 \mu m$		
40	*Inside Calipers	Length	C.S. for Inside Calipers JJF(Zhe) 1091	(0~500) mm	$U=0.01mm$		
41	*Brick Calipers	Length	C.S. for Brick Calipers JJF(Zhe) 1130	(-50~+50) mm	$U=0.01mm$		
42	*Recess Measuring Micrometer	Length	V.R. of Measuring Micromete JJG(JG) 56	(0~50) mm	$U=3 \mu m$		
43	*Wide Rande Electronic Digital Display Dial Indicator	Length	C.S. for Wide Rande Electronic Digital Display Dial Indicator JJF(Zhe) 1135	(0~50) mm	$U=2 \mu m$		
44	Box Plates	Verticality	V.R. of Box Plates JJG 194	(100~400)mm	$U=2.8 \mu m$		
45	Square	Perpendicularity	V.R. of Squares JJG 7	wide-stand square: H(0~630)mm	$U=2 \mu m+3 \times 10^{-6}H$		
				Knife edge square: H(0~200)mm	$U=2 \mu m$		
				Parallel square: H(0~500)mm	$U=3 \mu m$		
				Triangle square: H(0~630)mm	$U=0.5 \mu m+1 \times 10^{-6}H$		
				Linear steel square: (0~500) mm	$U=0.06mm$		
46	Straight Edge	Linearity	V.R. of Straight Edge JJG 63	(75~175) mm	$U=0.4 \mu m$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(175~300)mm	$U=0.8 \mu\text{m}$		
				(300~500)mm	$U=1.4 \mu\text{m}$		
47	Square Gauge	Perpendicularity	V.R. of Square Gauge JJG 1046	H:(100~630)mm	$U=0.2 \mu\text{m}+2 \times 10^{-6}H$		
48	*General Bevel Protractor	Angle	C.S. for General Bevel Protractors JJF 1959	(0~360) $^{\circ}$	$U=1'$		
49	*Composite angle gauge	Angle	C.S. for Composite Type Angle Rules JJF 1132	(0~180) $^{\circ}$	$U=4'$		
		Length		(0~300)mm	$U=0.05\text{mm}$		
50	Frame Levels and Shaft Levels	Angle	C.S. for Frame Levels and Shaft Levels JJF 1084	(0.02~0.10) mm/m	$U_{\text{rel}}=6\%$		
51	*Calibrators for the Level	Angle	V.R. of Calibrators for the Level JJG 191	(0.005~1.5)mm/m	$U_{\text{rel}}=2\%$		
52	clinometers	Angle	Calibration Specification for Clinometers JJF 1915	(0~360) $^{\circ}$	$U=0.01^{\circ}$		
53	Level Rule	Angle	C.S. for Level Rules JJF 1085	(0.5~10) mm/m	$U_{\text{rel}}=4\%$		
		Length		(0~1000)mm	$U=0.1\text{mm}$		
54	*Image contact angle tester	angle	Calibration specification for image contact angle tester JJF (Su) 219	(10~170) $^{\circ}$	$U=0.20^{\circ}$		
55	*Digital Display Height Measuring Instrument	Length	C.S. for Height Measuring Instrument with Digital JJF 1254	(0~1000)mm	$U=0.5 \mu\text{m}+1.5 \times 10^{-6}L$		
56	*Concentricity Testers	Length	C.S. for Concentricity Testers JJF 1109	(20~30) μm	$U=2 \mu\text{m}$		
57	*Tooth thickness caliper	length	Calibration specification for tooth thickness caliper JJF 1072	Modulus: (1~50) mm	$U=0.01\text{mm}$		



No. CNAS L2336

第 7 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
58	*Common Normal Micrometer	Length	V.R. of Common Normal Micrometer JJG 82	(0~200)mm	$U=1.2 \mu m+5.5 \times 10^{-6}L$		
59	*Thread micrometer	length	Verification regulation of thread micrometer JJG25	(0~200)mm	$U=1 \mu m+7 \times 10^{-6}L$		
60	Screw Templates	Length	V.R. of Screw Templates JJG 60	(0.4~6.35)mm	$U=2 \mu m$		
61	Working gauge for petroleum thread	Length	C.S. for OCTG Thread Working Gauges JJF 1108	Pitch:(0.941~3.175)mm	$U=4.0 \mu m$		
				Medium diameter: (5~200)mm	$U=3.4 \mu m+6 \times 10^{-6}L$		
62	*Coordinate Measuring Machine	Length	C.S. for Coordinate Measuring Machines JJF 1064	(0~1000)mm	$U=0.5 \mu m+1 \times 10^{-6}L$	except for the coordinate measuring machine with rotary table as the fourth axis.	
				(1000~7000)mm	$U=2 \mu m+1 \times 10^{-6}L$		
63	*Roundness instrument	Roundness	Measurement Standard Instrument of Roundness and Cylindricity JJG 429	(1~11) μm	$U_{rel}=6\%$		
				(19~20) μm	$U_{rel}=4\%$		
64	*Flat Plate	Flatness	V.R. of Surface Plates JJG 117	160mm×100mm~630mm×400mm	$U=0.9 \mu m$		
				630mm×400mm~1600mm×1000mm:	$U=2.2 \mu m$		
				1600mm×1000mm~3000mm×2000mm:	$U=3.5 \mu m$		
65	Feeler gauge	Length	V. R. of Feeler Gauge JJG 62	(0.02~0.10)mm	$U=2 \mu m$		
				(0.10~3.00)mm	$U=3 \mu m$		
66	Standard Ring Gauge	Length	V.R. of Standard Ring Gauge JJG 894	(2~200) mm	$U=0.7 \mu m+6 \times 10^{-6}L$		


No. CNAS L2336

第 8 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
67	*Snap Gauge Reading in 0.01mm	Length	V.R. of Snap Gauge Reading in 0.01mm JJG 109	(0~200)mm	$U=4 \mu m$		
68	Cylindrical Thread Gauges	Length	C.S. for Cylindrical Thread Gauges JJF 1345	Gauge:(1~500)mm	$U=3.4 \mu m+6 \times 10^{-6}L$		
				Ring:(2~500)mm	$U=3.4 \mu m+6 \times 10^{-6}L$		
69	Sealing pipe thread gauge	Length	Calibration specification for thread gauge of 55° sealing pipe JJF (Su) 223	Medium diameter: (7~165)mm	$U=3.4 \mu m+6 \times 10^{-6}L$		
70	Smooth limit gauge	Length	V.R. of Plain Limit Gauges JJG 343	External dimension:(1~500)mm	$U=0.5 \mu m+4 \times 10^{-6}L$		
				inside dimension:(1~500)mm	$U=1.0 \mu m+3 \times 10^{-6}L$		
71	Pin Gauge, Thread Measuring Wires	Length	C.S. for Cylindrical Measuring Pin JJF 1207	Thread Measuring Wires: (0.118~6.585)mm	$U=0.3 \mu m$		
				Pin Gauge: (0.1~25)mm	$U=0.5 \mu m$		
72	*Optimeter	Length	V.R. of Optimeter JJG 45	(-100~+100) μm	$U=0.12 \mu m$		
73	*Length Measuring Instruments	Length	C.S. for Length Measuring Instruments JJF 1066	(0~3000)mm	$U=0.2 \mu m+2 \times 10^{-6}L$		
74	*Length Measuring Instrument	Length	C.S. for Length Measuring Instrument JJF 1189	(0~50)mm	$U=0.2 \mu m$		
				(50~500)mm	$U=0.2 \mu m+2 \times 10^{-6}L$		
75	*Laser Diameter Measuring Gauges	Length	C.S. for Laser Diameter Measuring Gauges JJF 1250	(0.1~30)mm	$U=0.5 \mu m$		
				Standard gauge: (0.1~30)mm	$U=0.3 \mu m$		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
76	*Microcator	Length	V.R. of Microcator JIG 118	(-300~300) μm	$U=0.14 \mu\text{m}$		
77	*Comparator of Machinery Type	Length	V.R. of Comparator of Machinery Type JIG 39	(-1000~1000) μm	$U=0.12 \mu\text{m}$		
78	*Inductive Micrometers	Length	C.S. for Inductive Micrometers JJF 1331	Resolution 0.01 μm : (-10~+10) μm	$U=0.03 \mu\text{m}$		
				Resolution 0.1 μm : (-100~+100) μm	$U=0.2 \mu\text{m}$		
				Resolution 1 μm : (-1000~+1000) μm	$U=1 \mu\text{m}$		
79	*Grating Micrometer	Length	C.S. for Grating Micrometers JJF 1682	0.2 μm grade; (0~25) mm	$U=0.09 \mu\text{m}$		
				0.5 μm grade; (0~50) mm	$U=0.16 \mu\text{m}$		
				1 μm grade; (0~100) mm	$U=0.3 \mu\text{m}$		
				2 μm grade; (0~100) mm	$U=0.9 \mu\text{m}$		
				5 μm grade; (0~100) mm	$U=2 \mu\text{m}$		
				10 μm grade; (0~100) mm	$U=2 \mu\text{m}$		
80	Roughness Comparison Specimens	Roughness	Calibration specification for surface roughness comparison blocks JJF 1099	Ra(0.012~6.3) μm	$U_{\text{rel}}=7.6\%$		
81	*Contact Instruments of Surface Roughness Meters by the profile	Roughness	C.S. for Contact (Stylus) Instruments of Surface Roughness Measurement by the profile Method JJF 1105	Ra:(0.04~0.1) μm	$U_{\text{rel}}=6.0\%$		




No. CNAS L2336

第 10 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Method			Ra:(1.4~6.9) μm	$U_{rel}=4.0\%$		
82	Fineness of Grind Gauge	Length	V.R.of Fineness of Grind Gage JJG 905	(0~25) μm	$U=0.6 \mu m$		
				(25~50) μm	$U=1.0 \mu m$		
				(50~150) μm	$U=1.3 \mu m$		
83	*Contact (Stylus) Surface Contour Tester	Length	C.S. for Contact (Stylus) Surface Contour Tester JJF(Min) 1043	Z axis: (0~50)mm	$U=0.5 \mu m+3 \times 10^{-5}L$		
				X axis: (0~100)mm	$U=0.8 \mu m$		
				radius: 10mm	$U=0.9 \mu m$		
		Angle	0~90°	$U=40''$			
84	*Tool Presetting and Measuring Instruments	Length	V.R. of Tool Presetting and Measuring Instruments JJG 938	(0~1000)mm	$U=1.7 \mu m+2.2 \times 10^{-6}L$		
85	*Magnetic and Eddy Current Measuring Instrument for Coating Thickness	Length	V.R. of Magnetic and Eddy Current Measuring Instrument for Coating Thickness JJG 818	Magnetic : (0~2000) μm	$U=0.4 \mu m+6 \times 10^{-3}L$		
				(0~2000) μm	$U=0.4 \mu m+6 \times 10^{-3}L$		
				(10~2000) μm	$U_{rel}=0.6\%$		
86	*X-Ray Thickness Gauge	Length	C. S. for X-Ray Fluorescence Coating Thickness Instruments JJF 1306	(0.053~0.2) μm	$U_{rel}=10\%$		
				(0.2~15) μm	$U_{rel}=6\%$		
87	*Ultrasonic Thickness Instrument	Length	C.S. for Ultrasonic Thickness Instrument JJF 1126	(0.5~15) mm	$U=0.02mm$		
				(15~75) mm	$U=0.03mm$		
				(75~200)mm	$U=0.07mm$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
88	*Pneumatic measuring instrument for micrometers	Length	V.R.of Pneumatic Measuring Instrument for Micrometers JJG 356	Buoy-type read in 0.5 μ m: (0~8) μ m	$U=0.2 \mu$ m		
				Buoy-type read in 1 μ m: (0~15) μ m	$U=0.4 \mu$ m		
				Buoy-type read in 2 μ m: (0~40) μ m	$U=0.6 \mu$ m		
				Buoy-type read in 5 μ m: (0~80) μ m	$U=1.3 \mu$ m		
				Electronic column type read in 0.2 μ m: (0~10) μ m	$U=0.2 \mu$ m		
				Electronic column type read in 0.5 μ m: (0~25) μ m	$U=0.4 \mu$ m		
				Electronic column type read in 1 μ m: (0~50) μ m	$U=0.9 \mu$ m		
89	Radius Gauges	Length	C.S.for Radius Gauge JJF 2187	(R1~R25)mm	$U=10 \mu$ m		
90	Wedge-feet Calibrator for Micrometers	Length	V.R.of Wedge-feet Calibrator for Micrometers JJG 525	(0~2000) μ m	$U=0.3 \mu$ m		
91	*Welding Inspection Callipers	Angle	C.S. for Callipers for Welding Inspection Callipers JJF 2161	(0~320) $^{\circ}$	$U=8'$		
		Length		(0~60)mm	$U=0.03$ mm		
92	*Staight Edges	Linearity	C.S.for Staight Edges JJF 1097	(300~1000) mm	$U=0.7 \mu$ m		
				(1000~2000)mm	$U=1.0 \mu$ m		
				(2000~3000)mm	$U=1.4 \mu$ m		



No. CNAS L2336


第 12 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
93	*Surface profile table	Length	Calibration specification for Surface profile gauges JJF 1476	(0~6.5) mm	$U=1.7 \mu\text{m}$		
94	*Tensile stretch tester	elongation	Verification procedure for test equipment of enamelled winding wire- Part3:Elongation tester JB/T4279.3	1%~60%	$U=0.3\%$		
95	*Articulated arm type coordinate measuring machine	length	Calibration specification for articulated arm CMM JJF 1408	(0~1000)mm	$U=6 \mu\text{m}+4 \times 10^{-6}L$		
96	*Laser micrometer	length	Calibration specification for laser micrometer JJF 1663	(0.1~6)mm	$U=0.024\%FS$		
				(6~100)mm	$U=0.015\%FS$		
97	*Solder Paste Inspection Instruments	length	Calibration Specification for Solder Paste Inspection Instruments JJF1965	(75~400) μm	$U=1.5 \mu\text{m}$		
98	*Drop Test Machine for Packages	Length	Drop Test Machine for Packages JJG (Yue) 045	(10~2000)mm	$U=0.7\text{mm}$		
		Angle		0~5°	$U=0.6^\circ$		
99	*Extensometer	Length	Verification Regulation for Extensometer JJG 762	(0~1/3) mm	$U=0.5 \mu\text{m}$		
				(1/3~25) mm	$U_{\text{rel}}=0.18\%$		
100	Microscale	Length	C.S.for Micropattern Standards JJF 1917	line spacing: (0~10)mm	$U=0.5 \mu\text{m}$		
				diameter: (0~10)mm	$U=0.8 \mu\text{m}$		
		Angle		(0~360) °	$U=16''$		
101	Optical instrument	Length	C.S. for Material Measures of Length Measuring Instrument	Eccentric shafte: e(20~30) μm	$U=0.8 \mu\text{m}$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	inspection tool	Length	with Optical Principle JJF1941	Standard mandrel: (0~500)mm	$U=0.5 \mu m$		
102	*Center distance caliper	Length	C.S. for CenterDistance Calipers JJF (Su) 199	(0~500)mm	$U=0.01mm$		
				(500~1000)mm	$U=0.02mm$		
103	*Chamfer caliper	Length	C.S. for Calipers and gauges of Chamfer JJF (Su) 211	(1.2~10)mm	$U=0.02mm$		
104	Gauges for plugs and sockets for household and similar purposes	Length	C.S. for plugs and Socket-outlets gauges forHousehold and Similar Purposes JJF (Zhe) 1119	(0~60)mm	$U=0.004mm$		
		Angle		(119~121)°	$U=2'$		
105	One dimensional and two-dimensional geometric dimensions	Length	C.S. for one-dimensional and two-dimensional geometric dimension measurement JJF(Xin)68	(0~500)mm	$U=0.5 \mu m+5 \times 10^{-6}L$		
				(500~1000)mm	$U=0.7 \mu m+5 \times 10^{-6}L$		
		Angle		(0~360)°	$U=8''$		
106	Hand-held Laser Distance Meters	Length	V.R.of Hand-held Laser Distance Meters JJG 966	(0~30) m	$U=1.0mm+2 \times 10^{-5}D$		
107	*Quick Video Measuring Machines	Length	C.S.for Quick Video Measuring Machines JJF (Zhe) 1194	(0~500)mm	$U=0.5 \mu m+2 \times 10^{-6}L$		
108	Levels	Angle	V.R.of Levels JJG 425	(-25~+25)''	$U=3''$		
109	Optical Theodolites	Angle	V.R.of Optical Theodolites JJG 414	(0~360)°	$U=0.4''$		
110	Protection of person and equipment by enclosures-Probe for verification	Length	C.S. for Protection of person and equipment by enclosures-Probe for verification JJF(Liao) 288	(0~100)mm	$U=0.004mm$		
				(100~200)mm	$U=0.007mm$		
				(200~300)mm	$U=0.04mm$		
				(300~500)mm	$U=0.05mm$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Angle		(0~360)°	U=4'		
111	Laser Alignment Instrument	Length	C.S.for Laser Alignment Instrument JJF(Zhe) 1196	(-20~+20)mm	U=3 μm		
		Angle		(0~360)°	U=0.1°		
112	*Step Gauges	Length	C.S.for Step Gauges JJF(Ji) 154	(-50~+50) mm	U=0.01mm		
113	*Vertical Axis Deviation Measuring Instrument	Length	C.S.for Vertical Axis Deviation Measuring Instrument JJF (Ji) 153	(0~300) mm	U=0.007mm		
114	*Bottom Wall Thickness Measuring Instrument	Length	C.S.for Bottom Wall Thickness Measuring Instrument JJF(Ji) 152	(0~20) mm	U=0.003mm		
115	*Optical 3D Measuring Systems Based on Structured Light Scanning	Length	C.S.for Optical 3D Measuring Systems Based on Structured Light Scanning JJF 1951	(0~1200)mm	U=7 μm+3×10 ⁻⁶ L	No single perspective system type equipment	
116	*Conical Ruler	Length	C.S. of Conical Ruler JJF (Wan) 89	(0~60) mm	U=0.01mm		
117	Taper Thread Gauges	Length	C.S.for Taper Thread Gauges JJF (Xiang) 47	(6~165)mm	U=3.6 μm+6×10 ⁻⁶ L		
118	*Wide Face Caliper	Length	C.S.for Wide Face Caliper JJF(Zhe) 1172	(0~300) mm	U=0.01mm		
119	*Hall Effect Thickness Gauge	Length	V.R.of Hall Effect Thickness Gauge JJG(Yue) 034	(0~20)mm	U=2 μm+4×10 ⁻³ H		
120	Pointer Angle Gauge	Angle	C.S.for Pointer Angle Gauge JJF(Su) 280	(-180~+180) °	U=0.2°		



No. CNAS L2336

第 15 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
121	Coordinate Measuring Spheres	Length	C.S.for Coordinate Measuring Spheres JJF 1422	Φ (10~50)mm	$U=0.4 \mu\text{m}$		
		Roundness		(0~10) μm	$U=0.07 \mu\text{m}$		
122	*Creepage Distance (Clearance) Test Card	Length	C.S.for Creepage Distance (Clearance) Test Card JJF(Shan) 105	(1~10) mm	$U=0.005\text{mm}$		
123	*Micrometer With μm Reading	Length	C.S.for Micrometer With μm Reading JJF(Zhe) 1134	(0~100) mm	$U=1 \mu\text{m}$		
124	*Micrometers for Measuring Gear	Length	V.R.of Micrometers for Measuring Gear JJG(Yue) 074	(0~300) mm	$U=0.7 \mu\text{m} + 9 \times 10^{-6}L$		
125	Phelin Rule	Length	C.S.for Phelin Rule JJF(Meng) 107	Line Spacing: (0~300) mm	$U=7 \mu\text{m}$		
				Line Width: (0.01~10) mm	$U=4 \mu\text{m}$		
				Diameter : (0.01~10) mm	$U=4 \mu\text{m}$		
126	*Linear Displacement Sensors	Length	Calibration Specification for Large Deformation Extensometer JJF(E) 141	(100~1000)mm	$U_{\text{rel}}=0.16\%$		
Thermology measuring instruments							
1	Working Noble Metal Thermocouples	Temperature	V.R.of Working Noble Metal Thermocouples JJG 141	(300~1100) °C	$U=0.8^{\circ}\text{C}$		
				(1100~1500)°C	$U=2.6^{\circ}\text{C}$		
2	Base Metal Thermocouples	Temperature	C. S. for Base Metal Thermocouples JJF 1637	(-40~300)°C	$U=0.3^{\circ}\text{C}$		
				(300~1200)°C	$U=1.1^{\circ}\text{C}$		



No. CNAS L2336

第 16 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
3	Sheathed Thermocouples	Temperature	C. S. for Sheathed Thermocouples JJF1262	(-40~300)°C	U=0.3°C		
				(300~1100)°C	U=1.1°C		
4	Industrial Platinum and Copper Resistance Thermistor	Temperature	V. R. of Industrial Platinum and Copper Resistance Thermistor JJG 229	Platinum Resistance:(-80~300)°C	U=0.09°C		
				Copper Resistance:(0~150)°C	U=0.12°C		
5	Standard Mercury-in-Glass Thermometers	Temperature	V.R.of Standard Mercury-in-Glass Thermometers JJG 161	(-60~300)°C	U=0.04°C		
6	Liquid-in-Glass Thermometers for Working	Temperature	V.R. of Liquid-in-Glass Thermometers for Working JJG 130	(-80~300)°C	U=0.06°C		
7	Electric contact mercury-in-glass thermometers	Temperature	V.R. of Electric Contact Mercury-in-Glass Thermometers JJG 131	(-30~100)°C	U=0.05°C		
				(100~300)°C	U=0.08°C		
8	Clinical Electronic Thermometer	Temperature	V.R. for the Clinical Electronic Thermometer JJG 1162	(35~41)°C	U=0.03°C		
9	Clinical Thermometers	Temperature	V.R. of the Clinical Thermometers JJG 111	(30~43)°C	U=0.03°C		
10	*Bimetallic Thermometers	Temperature	Calibration Specification for Bimetallic Thermometers JJF 1908	(-80~300)°C	U=0.5°C		
11	*Filled System Thermometers	Temperature	Calibration Specification for Filled System Thermometers JJF 1909	(-80~300)°C	U=0.5°C		
12	Surface Thermometers	Temperature	C.S.for the Surface Thermometers JJF 1409	(50~400)°C	U=1.8°C		
13	Surface Platinum	Temperature	C.S. for Surface Platinum Resistance Thermometers JJF	(-30~50) °C	U=0.8°C		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Resistance Thermometers		2137	(50~300) °C	U=1.6°C		
14	*Digital Thermometer	Temperature	C.S.for digital thermometer JJF(Su) 95	(-196~0) °C	U=0.3°C		
				(-80~300)°C	U=0.04°C		
				(300~1000)°C	U=0.90°C		
15	Thermistor Thermometer	Temperature	C. S. for Thermistor Thermometers JJF1379	(15~35) °C, Resolution 0.001 °C	U=0.013 °C		
				(-50~200) °C, Resolution 0.1 °C	U=0.1 °C		
16	*Thermometers of Clinic Autoclave	Temperature	C.S. for Thermometers of Clinic Autoclave JJF 1308	(20~150)°C	U=0.1°C		
17	WBGT-index Meters	Temperature	C.S. for Thermometers of WBGT-index Meters JJF 1407	(5~120)°C	U=0.3°C		
18	Fiber-optic Distributed Thermometers	Temperature	C.S. for Fiber-optic Distributed Thermometers JJF 1630	(-20~100)°C	U=0.4°C		
19	Working Copper/Copper-Nick Thermocouple	Temperature	V.R.of the Working Copper/Copper-Nick Thermocouple JJG368	(-80~300)°C	U=0.2°C		
20	*Temperature Indicators and Simulators by Electrical Simulation and Measurement	Temperature	C.S.for Temperature Indicators and Simulators by Electrical Simulation and Measurement JJF 1309	(-200~1600)°C	U=0.02°C		
21	Recorders for Industrial Process	Temperature	V.R. of the Recorders for Industrial Process	With Platinum thermal resistance:(-200~800)°C	U=0.2°C		



在线扫码获取验证

No. CNAS L2336

第 18 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Measurement		Measurement JJG 74	With Thermocouple: (-200~1600) °C	U=0.8°C		
22	Temperature Data Acquisition Instruments	Temperature	C. S. for Temperature Data Acquisition Instruments JJF 1366	(-80~300)°C	U=0.08°C		
23	Temperature and Humidity Itinerant Detecting Instruments	Temperature	Calibration Specification for Temperature and Humidity Itinerant Detecting Instruments JJF 1171	(-80~300)°C	U=0.04°C		
				(300~1200)°C	U=0.67°C		
		Humidity		10%RH~95%RH	U=1.6%RH		
24	*Analogue Temperature Indicators and Controllers	Temperature	V.R.of Analogue Temperature Indicators and Controllers JJG 951	With Platinum Thermal Resistance:(0~850)°C	U=1.0°C		
				With Thermocouple:(0~1300)°C	U=1.0°C		
25	*Digital Temperature Indicators and Controllers	Temperature	V.R.of Digital Temperature Indicators and Controllers JJG 617	With Platinum Thermal Resistance:(-100~600)°C	U=0.4°C		
				With Thermocouple: (-100~1300) °C	U=0.6°C		
26	*Iron Thermometer	Temperature	C.S. for Soldering Iron Thermometers JJF 1629	(50~600)°C	U=1.5°C		
27	Temperature transmitters	Temperature	C.S.for Temperature Transmitter JJF 1183	With Platinum Thermal Resistance: (-80~300)°C	U=0.2°C		
				Without Platinum Thermal Resistance: (-200~800)°C	U=0.6°C		
				With Thermocouple: (-80~1100)°C	U=1.0°C		
				Without Thermocouple: (-200~1300)°C	U=1.2°C		



No. CNAS L2336

第 19 页 共 187 页


The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
28	*Temperature Indicators	Temperature	C.S. for Temperature Indicators JJF 1664	With Platinum Thermal resistance: (-100~850) °C	U=0.4°C		
				With Thermocouple: (-200~1300) °C	U=0.6°C		
29	*Thermostatic Baths for Temperature Calibration	Temperature	Measurement and Test Norm of Thermostatic Bath's Technological Characteristics JJF 1030	Uniformity: (-80~300)°C	U=0.007°C		
				Volatility: (-80~300)°C	U=0.009°C		
30	*Temperature Block Calibrators	Temperature	Calibration Guideline of the Temperature Block Calibrators JJF 1257	(-60~400)°C	U=0.32°C		
31	*Equipment of the Environmental Testing	Temperature	C. S. for the Equipment of the Environmental Testing for Temperature and Humidity Parameters JJF 1101	(-80~300)°C	U=0.5°C		
		Humidity		20%RH~95%RH	U=1.5%RH		
32	*Equipment of the Chamber Electric Furnace	Temperature	C.S. for the Equipment of the Chamber Electric Furnace JJF 1376	(100~1100)°C	U=1.5°C		
33	*Steam Sterilizer	Temperature	C.S. for Steam Sterilizer's Temperature and Pressure JJF(Su) 96	(40~140)°C	U=0.5°C		
		Pressure		(1~200)kPa	U=0.5kPa		
34	*Salt Mist Testing Chambers	Temperature	Calibration Specification for Salt Mist Testing Chambers JJF 2168	(0~60)°C	U=0.6°C		
		Mist Rate		(1~2)mL/(h•80cm ²)	U=0.4mL/(h•80cm ²)		
35	*Natural Ventilation Thermal Aging Test Oven	Temperature	Verification Procedure for Test Equipment of Rubber Plastic Wire and Cable -Part 6:Natural Ventilation Thermal Aging Test Oven JB/T4278.6	(20~300)°C	U=0.8°C		
		Ventilation		(8~20)Times/Hour	U=1Times/Hour		



在线扫码获取验证


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
36	*Electro-thermal Incubator	Temperature	Calibration Method of Electro-thermal Incubator SL144.4	(0~50)°C	U=0.6°C		
37	*Unsaturated Pressurized Vapour Damp Heat/steady State Testing Equipments	Temperature	V.R.of Unsaturated Pressurized Vapour Damp Heat/steady State Testing Equipments JJG(DZ) 31504	(100~170) °C	U=0.5°C		
		Humidity		50%RH~90%RH	U=1.8%RH		
38	*Electrically-heated Thermostatic Water bath	Temperature	C.S. of Electrically-heated Thermostatic Water bath JJF(Liao) 118	(0~100)°C	U=0.2°C		
39	*Vacuum Drying Chamber	Temperature	C.S. for Temperature and Pressure of Vacuum Drying Chamber JJF(Su) 177	(40~150)°C	U=0.6°C		
		Pressure		(0~101)kPa	U=2.4kPa		
40	*Equipment of Biological Artificial Climate	Temperature	C.S. for Equipment of Biological Artificial Climate JJF(Zhe)1103	(5~40)°C	U=0.6°C		
		Humidity		50%RH~95%RH	U=1.7%RH		
		Illumination		(100~10000)lx	U _{rel} =3.0%		
41	*Dry Bath Incubator for Biological Experiments	Temperature	C.S. for Dry Bath Incubator for Biological Experiments JJF(Zhe)1149	(-10~150) °C (Temperature Deciation)	U=0.13°C		
				(-10~150) °C (Temperature Uniformity)	U=0.07°C		
				(-10~150) °C (Temperature Volatility)	U=0.10°C		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
42	*Temperature and Humidity Standard Chambers	Temperature	C.S. for Temperature and Humidity Standard Chambers JJF1564	(5~50)°C (uniformity)	U=0.05°C		
				(5~50)°C (fluctuation)	U=0.02°C		
		Temperature change rate		(0~5)°C/min	U=0.03°C/min		
		Humidity		20%RH~90%RH (Uniformity)	U=0.6%RH		
				20%RH~90%RH (Fluctuation)	U=0.3%RH		
		Humidity change rate		0RH/min~1%RH/min	U=0.6%RH/min		
43	Working Radiation Thermometers	Temperature	V.R. of the Working Radiation Thermometers JJG 856	(50~400)°C	U=1.3°C		
				(400~800)°C	U=1.7°C		
				(800~1000)°C	U=2.0°C		
				(1000~1200)°C	U=3.0°C		
44	Infrared Thermometers for Measurement of Human Temperature	Temperature	C.S. for Infrared Thermometers for Measurement of Human Temperature JJF 1107	(22.0~40.0) °C	U=0.14°C		
45	Thermal Imagers	Temperature	C. S. for Thermal Imagers JJF 1187	(50~300)°C	U=1.5°C		
				(300~500)°C	U=1.8°C		
				(500~900)°C	U=2.4°C		
				(900~1200)°C	U=3.0°C		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
46	Mechanical Thermo-hygrometers	Temperature	V.R. of Mechanical Thermo-hygrometers JJG 205	(5~50) °C	U=0.6 °C		
		Humidity		30%RH~90%RH	U=1.8%RH		
47	Digital Temperature-hygrometers	Humidity	C.S. for Digital Temperature-hygrometers JJF 1076	10%RH~90%RH	U=0.6%RH~1.5%RH		
		Temperature		(-10~50) °C	U=0.4 °C		
48	Temperature and Humidity Transmitter	Temperature	C.S. for Temperature and Humidity Transmitter JJF(Zhe)1035	(5~50) °C	U=0.4 °C		
		Humidity		20%RH~90%RH	U=1.9%RH		
49	*Dust and Sand Testing Equipments	Temperature	C. S. for Dust and Sand Testing Equipments JJF(JG) 18	(21~75) °C	U=0.4 °C		
		Humidity		20%RH~30%RH	U=1.5%RH		
		Wind Velocity		(0.5~2.5) m/s	U _{rel} =12%		
				(7.7~10.1) m/s	U _{rel} =3.2%		
				(18~29) m/s	U _{rel} =2.8%		
Dust Concentration		(0~20)g/m ³	U=0.76g/m ³				
50	*Rain Testing Equipments	Rainfall Intensity	C. S. for Rain Testing Equipments JJF(JG) 17	(90~110) mm/h	U _{rel} =2.6%		
		Wind Velocity		(16~20) m/s	U _{rel} =4.6%		
51	*Carbon Dioxide Incubator	Temperature	C.S. for Carbon Dioxide Incubator JJF (Liao) 463	(20~55) °C	U=0.2 °C		
		Carbon dioxide concentration		0.1%~10%	U _{rel} =2.4%		
52	*Oxygen Transmission Rate Testers	Oxygen transmittance	C.S. for Oxygen Gas Transmission Rate Testers JJF (Xiang) 66	(0.02~20)cm ³ /(m ² · 24h)	U _{rel} =9%		




№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(20~70) cm ³ /(m ² ·24h)	U _{rel} =7%		
		Temperature		(20~40) °C	U=0.26°C		
53	*Gas Transmittance Meter	Gas transmittance	C. S. for Gas Transmittance Meter JJF(Qian) 42	(0.01~10)[cm ³ /(m ² ·24h·0.1MPa)]	U _{rel} =26%		
				(10~12) [cm ³ /(m ² ·24h·0.1MPa)]	U _{rel} =20%		
		Temperature		(20~40) °C	U=0.4°C		
54	*Water Vapor Transmission Rate Tester	Water vapor transmittance	C.S. of Water Vapor Transmission Rate Tester JJF(Ji) 191	(0.01~2)g/(m ² ·24h)	U _{rel} =17%		
				(2~5) g/(m ² ·24h)	U _{rel} =19%		
		Temperature		(0~50) °C	U=0.14°C		
		Humidity		10%RH~95%RH	U=1.3%RH		
55	*Industrial Temperature Measurement System	Temperature	Calibration Specification for Industrial Temperature Measurement System JJF (Gan) 018	(-90~400) °C	U=0.1°C		
56	*Liquid constant temperature testing equipment	Temperature	Measurement specification for temperature performance of liquid constant temperature testing equipment JJF 2019	Temperature deviation: (-80~300)°C	U=0.2°C		
	Temperature	Temperature uniformity: (-80~300) °C		U=0.02°C			
	Temperature	Temperature fluctuation: (-80~300) °C		U=0.04°C			



No. CNAS L2336


第 24 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
57	*Large steam sterilizers	Temperature	Calibration Specification for Temperature, Pressure and Time Parameters of Large Steam Sterilizers JJF 2088	(0~150) °C	U=0.12°C		
		Pressure		(0~400)kPa	U=1.2kPa		
		Time		(1~3600)s	U=1s		
58	*Medical Cold Storage and Transportation Equipment	Temperature	Calibration Specification of Medical Cold Storage and Transportation Equipment JJF(Liao)310	(2~10) °C	U=0.2°C		
		Humidity		10%RH~90%RH	U=1.6%RH		
		Time		(1~3600)s	U=0.2s		
59	*Temperature Alternating and Temperature Shock Testing Equipments	Temperature	Calibration Specification for Temperature Alternating and Temperature Shock Testing Equipments JJF(Min)1121	(-80~300) °C	U=0.4°C		
		Temperature Rate of Change		(0~30) °C/min	U=0.08°C/min		
		Time		(0~3600) s	U=0.4s		
60	*Ozone Aging Test Chambers	Temperature	Calibration Specification for Ozone Aging Test Chambers JJF 2051	(0~100) °C	U=0.4°C		
		Humidity		10%RH~90%RH	U=1.3%RH		
		Ozone concentration		(0~400) μ mol/mol	(0.019~17) μ mol/mol		
61	*The Constant Temperature Culture Oscillator	Temperature	Calibration Specification for the Constant Temperature Culture Oscillator JJF(Liao) 359	(0~300) °C	U=0.4°C		
		Oscillation frequency		(0~1000) r/min	U=1.0r/min		
		Amplitude		(0~150) mm	U=0.4mm		
62	Mechanical Thermometers of Refrigerator	Temperature	Calibration Specification for Mechanical Thermometers of Refrigerator JJF(Xin)47	(-40~50) °C	U=0.4°C		
63	Digital Thermometers of	Temperature	Calibration Specification for Digital Thermometers of	Built-in: (-40~50) °C	U=0.3°C		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Refrigerator		Refrigerator JJF(Xin)63	External: (-40~50) °C	U=0.2°C		
64	Thermocouple compensation wire	Temperature	Calibration Specification for Base Metal Thermocouples (Appendix C Calibration Method for Thermocouple Compensation Wires) JJF1637	(-40~100) °C	U=0.2°C		
65	*Temperature/Humidity/Vibration Combined Environmental Testing System	Temperature	C. S. for Temperature/Humidity/Vibration Combined Environmental Testing System JJF 1270	(-75~200)°C	U=0.5°C		
		Humidity		20%RH~95%RH	U=1.7%RH		
		Acceleration		(1~980)m/s ²	U _{rel} =3.0%		
		Wind velocity		(0.1~20)m/s	U _{rel} =3.0%		
		Rate of temperature change		(0.5~60)°C/min	U _{rel} =5.0%		
		The test system operates at maximum noise		(30~130)dB	U=0.7dB		
		Indication of acceleration amplitude in temperature cycle		(1~980)m/s ²	U _{rel} =2.6%		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Vibration table additional table acceleration uniformity		0.1%~100%	$U_{rel}=4.0\%$		
66	*Irradiance of Artificial Accelerated Weathering Apparatus of Xenon Arc Lamp	UV Irradiance	C.S. for Irradiance of Artificial Accelerated Weathering Apparatus of Xenon Arc Lamp JJF1525	(0.01~1000)W/m ²	$U_{rel}=14\%$		
67	*Asphalt Aging Oven	Temperature	V.R. of Asphalt Aging Oven JJG(JT)056	(0~300)°C	$U=0.16^{\circ}\text{C}$		
		Length		(0~300)mm	$U=0.04\text{mm}$		
		Rotate speed		(4~20)r/min	$U=0.08\text{r/min}$		
		Time		(0~200)min	$U=5.2\text{s}$		
		Flow		(3.5~4.5)L/min	$U=56\text{mL/min}$		
68	*Constant Temperature Heating Platform	Temperature	Calibration Specification for Constant Temperature Heating Platform JJF (JG) 256	(50~100)°C	$U=1.2^{\circ}\text{C}$		
				(100~300)°C	$U=1.3^{\circ}\text{C}$		
				(300~400)°C	$U=1.5^{\circ}\text{C}$		
69	*Atmospheric Pressure Test Chambers	Temperature	Calibration Specification for Atmospheric Pressure Test Chambers JJF 2119	(-80~200) °C	$U=0.2^{\circ}\text{C}$		
		Time		(0.1~6000) s	$U=0.8\text{s}$		
		Pressure		(0.1~2)kPa	$U=0.06\text{kPa}$		
				(2~25)kPa	$U=0.19\text{kPa}$		




在线扫码获取验证

No. CNAS L2336

第 27 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(25~120)kPa	U=0.70kPa		
		Air pressure rate of change		(1~20) kPa/min	U=0.05kPa/min		
70	*Magnetic Heating Stirrer	Temperature	Calibration Specification for Magnetic Heating Stirrer JJF(Hei) 12	(0~100) °C	U=0.2°C		
		Speed		(50~2000) r/min	U _{rel} =0.2%		
71	*Surface Temperature Source	Temperature	Calibration Specification for the Surface (Appendix D Measurement Methods for Temperature Control Stability and Uniformity of Surface Temperature Sources) JJF1409	Uniformity: (50~400) °C	U=0.3°C		
				Stability: (50~400) °C	U=0.2°C		
72	*Low Temperature Freezers	Temperature	Calibration Specification of Low Temperature Freezers JJF(E) 135	(-80~-25) °C	U=0.6°C		
73	*Ethylene Oxide Sterilization	Temperature	Calibration Specification of Ethylene Oxide Sterilization JJF(E) 68	(37~63) °C	U=0.3°C		
		Humidity		30%RH~80%RH	U=2.4%RH		
		Pressure		(-20~100)kPa	U=1.3kPa		
74	*Washer-disinfection	Temperature	Calibration Specification of Washer-disinfection for Temperature Parameters JJF(Jin) 06	(20~100) °C	U=0.3°C		
Mechanics measuring instruments							
1	Weights	Mass	V. R.of Weights JJG 99	(1~100)mg	U=(0.002~0.005)mg		
				(0.1~10)g	U=(0.005~0.020)mg		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(10~500)g	$U=(0.020\sim 0.25)\text{mg}$		
				(0.5~1)kg	$U=(0.25\sim 1.6)\text{mg}$		
				(1~25)kg	$U=1.6\text{mg}\sim 0.12\text{g}$		
2	*Mechanical Balance	Mass	V. R. of Mechanical Balance JIG 98	1mg~200g	$U=0.03\text{mg}$		
3	*Electronic Balance	Mass	V. R. of Electronic Balance JIG 1036, C. S. for Electronic Balances JJF 1847	(1~100)mg	$U=(0.0020\sim 0.0050)\text{mg}$		
				(0.1~10)g	$U=(0.0050\sim 0.03)\text{mg}$		
				(10~1000)g	$U=(0.03\sim 2)\text{mg}$		
				(1~150)kg	$U=2\text{mg}\sim 0.7\text{g}$		
4	*Torsion Balance	Mass	V. R. of Torsion Balance JIG 46	(5~2500)mg	$U=(0.04\sim 1.3)\text{mg}$		
5	*Table Balance	Mass	V. R. of Table Balances JIG 156	(0.1~10)kg	$U=(0.03\sim 1.6)\text{g}$		
6	*Non-self indication Weighing instruments	Mass	V. R. of Non-self-indication Weighing Instruments JIG 14	(0.02~2)kg	$U=(0.5\sim 1)\text{g}$		
				(2~1000)kg	$U=(0.001\sim 0.1)\text{kg}$		
				(100~2000)kg	$U=(0.1\sim 0.2)\text{kg}$		
7	*Digital Indicating Weighing Instrument	Mass	V. R. of Digital Indicating Weighing Instrument JIG 539	(2~200)g	$U=(0.02\sim 0.2)\text{g}$		
				(0.2~20)kg	$U=(0.2\sim 1.2)\text{g}$		
				(20~2000)kg	$U=1.2\text{g}\sim 0.2\text{kg}$		
				(2~10)t	$U=(0.2\sim 0.7)\text{kg}$		
8	Working Glass Container	Volume	V. R. of Working Glass Container JIG 196	(0~0.1)mL	$U=0.0008\text{mL}$	Only weighing method	


No. CNAS L2336

第 29 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(0.1~0.25)mL	U=0.001mL		
				(0.25~1)mL	U=0.002mL		
				(1~2)mL	U=0.003mL		
				(2~5)mL	U=0.005mL		
				(5~10)mL	U=0.006mL		
				(10~15)mL	U=0.008mL		
				(15~25)mL	U=0.010mL		
				(25~50)mL	U=0.016mL		
				(50~100)mL	U=0.024mL		
				(100~250)mL	U=0.05mL		
				(250~500)mL	U=0.08mL		
				(500~1000)mL	U=0.14mL		
				(1000~2000)mL	U=0.22mL		
9	locomotive pipette	Volume	V. R.of Locomotive Pipette JIG 646	(0.2~1) μ L	U=0.06 μ L		
				(1~10) μ L	U=(0.06~0.20) μ L		
				(10~100) μ L	U=(0.2~0.6) μ L		
				(100~1000) μ L	U=(0.6~3.0) μ L		
				(1000~10000) μ L	U=(3.0~15) μ L		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
10	Float Meter	Flow	V. R. of Float Meter JIG257	(3.3~100)L/min,(gas)	$U_{rel}=0.8\%$	Volumetric method only(grade 1.5 and below)	
11	*Electromagnetic Flowmeters	Flow	C. S. for Electromagnetic Flowmeters JJF(Su) 228	(0.5~8444)m ³ /h,DN:(15~1000)mm	$U_{rel}=1.8\%$		
12	*liquid flowmeter	Flow	One Calibraion Specification for liquid flowmeter JJF(Xin) 99	(0.5~8444)m ³ /h,DN:(15~1000)mm	$U_{rel}=1.8\%$		
13	*Elastic Element Precise Pressure Gaugee and Vacuum Gauges	Pressure	V. R. of Elastic Element Precise Pressure Gaugee and Vacuum Gauges JJG 49	(-0.1~100)MPa	$U=0.09\%FS$	(100 ~ 500) MPa cannot be calibrated on site	
				(100~500)MPa	$U_{rel}=0.03\%$		
14	*Elastic Element Precise Pressure Gauges,Pressure-Vacuum Gauges and Vacuum Gauges for General Use	Pressure	V. R. of Elastic Element Precise Pressure Gauges,Pressure-Vacuum Gauges and Vacuum Gauges for General Use JJG 52	(-0.1~100)MPa	$U=0.6\%FS$	(100 ~ 500) MPa cannot be calibrated on site	
				(100~500)MPa	$U_{rel}=0.4\%$		
15	*Digital Pressure Gauges	Pressure	V. R. of Digital Pressure Gauges JJG 875	(-0.1~100)MPa	$U=0.09\%FS$	Under 100MPa: only grade 0.2 and below; 100MPa and above:only grade 0.02	



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(100~500)MPa	$U_{rel}=0.02\%$	and below; (100 ~ 500) MPa cannot be calibrated on site	
16	*Pressure Transducer(Static)	Pressure	V. R. of the Pressure Transducer(JJG 860)	(-0.1~100)MPa	$U=0.09\%FS$	Under 100MPa: only grade 0.2 and below; 100MPa and above:only grade 0.05 and below; (100 ~ 500) MPa cannot be calibrated on site	
				(100~500)MPa	$U_{rel}=0.02\%$		
17	*Pressure Controller	Pressure	V. R. of Pressure Controller JJG 544	(-0.1~100)MPa	$U=0.09\%FS$	(100~500) MPa cannot be calibrated on site	
				(100~500)MPa	$U_{rel}=0.2\%$		
18	*Pressure Transmitter	Pressure	V. R. of Pressure Transmitter JJG 882	(-0.1~100)MPa	$U=0.07\%FS$	Under 100MPa:only grade 0.2 and below;(100 ~ 500) MPa cannot be calibrated on site	
				(100~500)MPa	$U_{rel}=0.02\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
19	Pressure Regulators with Bourdon Tube Pressure Gauge	Pressure	C. S. for Pressure Regulators with Bourdon Tube Pressure Gauge JJF 1328	(0.001~25)MPa	$U=0.9\%FS$		
20	*Point Type Micro-differential Pressure Gauge	Pressure	V. R. of Point Type Micro-differential Pressure Gauge JJG(Yue) 020	(-30~30)kPa	$U=0.7\%FS$		
21	Compensated Micro-manometer	Pressure	V. R. of Compensated Micro-manomete JJG 158	(-2500~2500) Pa	$U=0.7\%FS$		
22	Tilting Tube Micromanometers	Pressure	V. R. of Tilting Tube Micromanometers JJG 172	(-2~2)kPa	$U=1.0\%FS$		
23	*leak tester of digital difference pressure	Pressure	V. R. of leak tester of digital difference pressure GJB/J 5461	Span: (-1~1) kPa	$U=7.0Pa$		
				(-0.1~2) MPa	$U=0.3\%FS$		
24	Standard Dynamometers	Force	V. R. of Standard Dynamometers JJG 144	(10~1100) N	$U_{rel}=0.03\%$		
25	*Working Dynamometers	Force	V. R. of Working Dynamometers JJG 455	1cN~500N	$U_{rel}=0.1\%$		
				500N~600kN	$U_{rel}=0.15\%$		
				(600~5000)kN	$U_{rel}=0.36\%$		
26	*Working Force Measuring Machines for Special Purposes	Force	C. S. for Working Force Measuring Machines for Special Purposes JJF 1134	1cN~5MN	$U_{rel}=0.4\%$		
27	*Tension, Compression and Universal Testing Machines	Force	V. R. of Universal Testing and Compression Testing Machine JJG 139	0.01N~5MN	$U_{rel}=0.36\%$		
		Displacement		(1~1000)mm	$U_{rel}=0.36\%$		
28	*Electronic Universal Testing Machine	Force	V. R. of Electronic Universal Testing Machine JJG 475	0.01N~5MN	$U_{rel}=0.36\%$		
		Displacement		(1~1000)mm	$U_{rel}=0.36\%$		



在线扫码获取验证

No. CNAS L2336

第 33 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rate		(0.1~1000) mm/min	$U_{rel}=0.36\%$		
29	*Electro-hydraulic Servo Universal Testing Machine	Force	V. R. of Electro-hydraulic Servo Universal Testing Machines JJG 1063	0.01N~1MN	$U_{rel}=0.3\%$		
				(1~3)MN	$U_{rel}=0.5\%$		
30	*Wire Bending Testing Machine	Angle	V. R. of Wire Bending Testing Machine JJG (Yue) 022	0° ~360°	$U=0.2^\circ$		
		Speed		(10~100)times/min	$U_{rel}=0.3\%$		
		Mass		(0.01~2)kg	$U=0.2g$		
31	*Flexure Testing Machine	Force	V. R. of Flexure Testing Machines JJG 476	(0.01~10)kN	$U_{rel}=0.4\%$		
		Torque Speed		(40~60)N/s	$U=2N/s$		
32	*High-Temperature Creep and Stress-Rupture Testing Machines	Length	V. R. of High-Temperature Creep and Stress-Rupture Testing Machines JJG 276	(0.01~15) mm	$U_{rel}=0.08\%$		
		Force		0.01N~1000kN	$U_{rel}=0.4\%$		
		Temperature		(50~900)°C	$U=1.4^\circ C$		
				(900~1000)°C	$U=1.8^\circ C$		
		Alignment		(0.01~25) %	$U=3\%$		
		Time		(0.01~3600) s	$U_{rel}=0.1\%$		
33	*Cupping Testing Machine	Force	V. R. of Cupping Testing Machine JJG 583	(9.5~10.5)kN	$U_{rel}=0.28\%$		
		Length		(15~60) mm	$U=0.02mm$		
34	*Hydraulic Jacks	Force	V. R. of Hydraulic Jacks JJG 621	(1~1000)kN	$U_{rel}=0.6\%$		
				(1~10)MN	$U_{rel}=0.7\%$		



No. CNAS L2336

第 34 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
35	Force Transducers	Force	V. R. of Force Transducers JIG 391	(100~1100)N	$U_{rel}=0.3\%$	Grade 0.3 and below only	
36	Screen Tension Meters	Force	C. S. for Screen Tension Meters JJF 1465	(5~50) N/cm	$U=1.5\%FS$		
37	*Pendulum Impact Testing Machines	Energy	V. R. of Pendulum Impact Testing Machines JIG 145	(1~30)J	$U_{rel}=2.0\%$		
				(30~110)J	$U_{rel}=1.9\%$		
				(110~300)J	$U_{rel}=1.8\%$		
		Length		(0.1~1000) mm	$U=0.02mm$		
38	*Cantilever-Beam(Izod-Type)Impact Testing Machine	Energy	V. R. of Cantilever-Beam(Izod-Type) Impact Testing Machine JIG 608	(1~150)J	$U_{rel}=0.5\%$		
		Length		(0.01~25) mm	$U=0.01mm$		
				(50~1000) mm	$U=0.06mm$		
39	*Calibration Instrument for Torque Wrenches	Torque	V. R. of Calibration Instrument for Torque Wrenches JIG 797	(0.2~100)Nm	$U_{rel}=0.18\%$		
40	*Working Torque-meters	Torque	V. R. of Working Torque-meters JIG1146	(0.1~2000)Nm	$U_{rel}=0.4\%$		
41	Torque Wrenches	Torque	V. R. of Torque Wrenches JIG 707	(0.2~500)Nm	$U_{rel}=0.5\%$		
				(500~2000)Nm	$U_{rel}=0.4\%$		
42	*Electric and pneumatic torque wrenches	Torque	C. S. for Electric and Pneumatic Torque Wrenches JJF 1610	(0.5~180)Nm	$U_{rel}=1.2\%$		
43	*Equipment of Power Measuring	Torque	V. R. of Equipment of Power Measuring JIG 653	(0.01~500)Nm	$U_{rel}=0.5\%$	Non-contact type only(grade B、grade C)	
		Rotate speed		(1~30000)r/min	$U_{rel}=0.2\%$		

No. CNAS L2336


第 35 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
44	*Metallic Brinell Hardness Testers	Hardness	V. R.of Metallic Brinell Hardness Testers JJG 150	(90~125)HBW	$U_{rel}=2.0\%$		
				(125~225)HBW	$U_{rel}=1.9\%$		
				(225~350)HBW	$U_{rel}=2.0\%$		
45	*Portable Brinell Hardness Tester	Hardness	C. S. for Portable Brinell Hardness Tester JJF 1595	(70~300)HBW	$U_{rel}=1.7\%$		
		Force		(0.06~30) kN	$U_{rel}=0.6\%$		
46	*Portable Rockwell Hardness Tester	Hardness	C. S. for Portable Rockwell Hardness Tester JJF 1594	(20~70)HRC	$U=1.0HRC$		
				(80~100)HRBW	$U=1.2HRBW$		
				(75~88)HRA	$U=1.0HRA$		
		Force		(0.08~1.5) kN	$U_{rel}=0.6\%$		
47	*Metallic Rockwell Hardness Testing Machines	Hardness	V. R.of Metallic Rockwell Hardness Testing Machines(Scales A, B, C, D, E, F, G, H, K, N, T) JJG 112	(20~30)HRC	$U=0.7HRC$		
				(35~55)HRC	$U=0.6HRC$		
				(60~70)HRC	$U=0.5HRC$		
				(45~88) HRA	$U=0.6HRA$		
				(85~100) HRBW	$U=0.7HRBW$		
				(89~91) HR15N	$U=0.9HR15N$		
				(42~54) HR30N	$U=1.0HR30N$		
				(74~80) HR30N	$U=0.9HR30N$		
				(32~61) HR45N	$U=1.1HR45N$		
				(88~93) HR15TW	$U=1.0HR15TW$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			C. S. for Ultrasonic Hardness Testers JJF 1436	(70~82) HR30TW	$U=0.9HR30TW$		
		Time		(0.1~20) s	$U=0.4s$		
		Length		(0~2) mm	$U=0.6 \mu m$		
		Angle		(119~121) °	$U=0.08^\circ$		
48	*Ultrasonic Hardness Testers	Hardness	C. S. for Ultrasonic Hardness Testers JJF 1436	(175~200)HV	$U_{rel}=5.0\%$		
				(200~400)HV	$U_{rel}=3.0\%$		
				(400~600)HV	$U_{rel}=1.7\%$		
				(600~800)HV	$U_{rel}=1.4\%$		
				(20~70)HRC	$U=0.5HRC$		
				(85~650)HBW	$U_{rel}=3.0\%$		
		Force		(1~98)N	$U_{rel}=1.2\%$		
49	*Plastic Rock Well Hardness Testing Machine	Hardness	V. R. of Plastic Rock Well Hardness Testing Machine JJG 884	(70~94)HRE	$U=0.9HRE$		
				(100~120)HRL	$U=0.8HRL$		
				(114~125)HRR	$U=0.9HRR$		
50	*Metallic Webster Hardness Testing Machines	Hardness	V. R. of Metallic Webster Hardness Testing Machines JJG 944	(8~17)HWA	$U=0.3HWA$		
				(4~6)HWB	$U=0.3HWB$		
51	*Equotip Hardness Tester	Mass	V. R. of Equotip Hardness Tester JJG 747	(0~30)g	$U=0.08g$		
		Hardness		(490~570)HLD	$U=5HLD$		
				(590~670)HLD	$U=6HLD$		




No. CNAS L2336

第 37 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(750~830)HLD	U=6HLD		
				(460~540)HLG	U=6HLG		
				(550~630)HLG	U=6HLG		
52	*Shore Hardness Tester	Hardness	V. R.of Shore Hardness Tester JJG 346	(25~140)HSD	U=0.8HSD		
53	*Metal Vickers Hardness Testers	Hardness	V. R.of Metal Vickers Hardness Testers JJG 151	(700~800) HV5	U _{rel} =2.7%		
				(175~225) HV5	U _{rel} =3.9%		
				(400~600) HV10	U _{rel} =2.6%		
				(400~600) HV30	U _{rel} =2.6%		
				(700~800) HV1	U _{rel} =3.5%		
				(700~800) HV0.5	U _{rel} =4.5%		
				(700~800) HV0.2	U _{rel} =5.0%		
				(400~600) HV0.1	U _{rel} =5.5%		
				(175~225) HV0.05	U _{rel} =6.0%		
54	Shore A Durometers	Hardness	V. R.of Shore A Durometers JJG 304	(0~100)HA	U=0.2HA		
		Length		(0.5~2) mm	U=0.01mm		
		Angle		(34~36) °	U=0.08°		
		Force		(0.5~8.5) N	U=0.03N		
55	Shore D Durometer	Hardness	V. R.of Shore D Durometer JJG 1039	(0~100)HD	U=0.2HD		
		Length		(0.05~4) mm	U=0.005mm		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Angle		(28~32)°	U=0.08°		
		Force		(0.05~45) N	U=0.03N		
56	*Type A Barcol Hardness Testers	Hardness	V. R.of Type A Barcol Hardness Testers JJG 610	(42~52)HBa	U=0.7HBa		
				(82~88)HBa	U=1.2HBa		
57	*Shore Durometer Calibrating Apparatus	Force	C. S. for Shore Durometer Calibrating Apparatus SBJZ-ZB-L002	1959mN	U=2mN	g=9.7946m/s ²	
				4897mN	U=2mN		
				9795mN	U=3mN		
				19589mN	U=5mN		
				48973mN	U=8mN		
		Length		(0~3)mm	U=0.5 μm		
58	Vibration Displacement Transducer	Displacement	V. R.of Vibration Displacement Transducer JJG 644	(0.1~5)mm, (3~2000)Hz	U _{rel} =1.8%		
59	Electromagnetic Velocity Transducer	Velocity	V. R.of Electromagnetic Velocity Transducer JJG 134	(0.1~100)cm/s, (3~2000)Hz	U _{rel} =1.6%		
		Velocity sensitivity		(1~30)mV/(mm·s ⁻¹), (3~2000)Hz	U _{rel} =1.6%		
60	Piezoelectric accelerometer	Acceleration	V. R.of Piezoelectric Accelerometer JJG 233	reference point:100m/s ² ,160Hz	U _{rel} =0.9%	Comparative method only	
				Passband: (1~441) m/s ² ,(2~52000)Hz	U _{rel} =1.2%		
61	Portable Vibration Calibrator	Acceleration	V. R.of Portable Vibration Calibrator JJG1062	(1~980)m/s ² ,(3~3570)Hz	U _{rel} =1.0%		



No. CNAS L2336

第 39 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-M	(1~980)m/s ² , (3570~5300)Hz	$U_{rel}=1.5\%$		
		Frequency		(3~5300)Hz	$U_{rel}=0.01\%$		
		Distortion factor		1%~20%	$U_{rel}=1\%$		
62	*Standard Vibrators	Frequency	V. R. of Standard Vibrators JJG 298	(3~5000)Hz	$U_{rel}=0.02\%$		
		Acceleration		(1~980)m/s ² , (3~5000)Hz	$U_{rel}=2.3\%$		
		Distortion factor		1%~50%	$U_{rel}=1\%$		
		Transverse vibration ratio		0.1%~50%	$U_{rel}=4.0\%$		
		Voltage		10mv~10v, (1~10)kHz	$U_{rel}=0.1\%$		
63	*Electrodynamic Vibration Testing Systems	Frequency	V. R. of Electrodynamic Vibration Testing Systems JJG 948	(1~5000)Hz	$U_{rel}=0.02\%$		
		Acceleration		(1~980)m/s ² , (1~5000)Hz	$U_{rel}=2.6\%$		
		Distortion factor		0.1%~50%	$U_{rel}=1\%$		
		Velocity		Sinusoidal: (0.001~10)m/s	$U_{rel}=2.1\%$		
				Change of impact velocity: (0.1~30)m/s	$U_{rel}=3.2\%$		
		Displacement		(0.001~200)mm	$U_{rel}=2.1\%$		
		PSD		(0.001~100)m ² /s ³	$U_{rel}=4.7\%$		
		Impact acceleration amplitude		(30~3000)m/s ²	$U_{rel}=4.7\%$		




No. CNAS L2336

第 40 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		pulse duration	ilac-MRA SCHEDULE OF ACCREDITATION CERTIFICATE	(3~50)ms	$U_{rel}=3.2\%$		
		Random dynamic range		(20~100)dB	$U=0.3\text{dB}$		
		Sound pressure level		(30~130)dB	$U=0.7\text{dB}$		
		Vibration control channel consistency		(0.01~3)dB	$U=0.01\text{dB}$		
		Magnetic Induction		(0.01~20)mT	$U_{rel}=2.0\%$		
		Force		(1.0~700)kN	$U_{rel}=2.8\%$		
64	*Mechanical Vibration Generator for Tsetting	Frequency	V. R.of Mechanical Vibration Generator for Tsetting JJG 189	(3~2000)Hz	$U_{rel}=0.2\%$		
		Displacement		(0.1~50)mm	$U_{rel}=2.6\%$		
		Distortion factor		1%~50%	$U_{rel}=1\%$		
		Acceleration		(1~200)m/s ²	$U_{rel}=2.6\%$		
		Sound Pressure Level		(30~130)dB	$U=0.7\text{dB}$		
65	*Hydraulic Vibration Testing System	Frequency	V. R.of Hydraulic Vibration Testing System JJG 638	(1~5000)Hz	$U_{rel}=0.02\%$		
		Acceleration		(1~200)m/s ²	$U_{rel}=3.0\%$		
		Distortion factor		0.1%~50%	$U_{rel}=3\%$		
		Velocity		(0.001~10)m/s	$U_{rel}=2.1\%$		



No. CNAS L2336

第 41 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Displacement		(0.1~500)mm	$U_{rel}=2.1\%$		
		PSD		(0.001~100)m ² /s ³	$U_{rel}=4.7\%$		
		Noise		(30~130)dB	$U=0.7dB$		
66	*Shock and Bump Testing Machines	Acceleration	V. R.of Shock and Bump Testing Machines JJG 1174	(30~1000) m/s ²	$U_{rel}=5\%$		
				(1000~50000) m/s ²	$U_{rel}=7\%$		
		Time		(0.3~60)ms	$U_{rel}=5\%$		
		Velocity		(0.01~20)m/s	$U_{rel}=4.2\%$		
		Repeat frequency of shock		(2~100) min ⁻¹	$U_{rel}=1.0\%$		
67	*Electrodynamic Horizontal Vibration Generator for Testing	Frequency	V. R.of Electrodynamic Horizontal Vibration Generator for Testing JJG 1000	(1~4000)Hz	$U_{rel}=0.05\%$		
		Acceleration		(1~980)m/s ² , (1~4000)Hz	$U_{rel}=2.6\%$		
		Displacement		(0.001~300)mm	$U_{rel}=2.6\%$		
		Distortion factor		0.1%~50%	$U_{rel}=1\%$		
		Sound pressure level		(30~130)dB	$U=0.7dB$		
68	Vibration Meters	Acceleration	V. R.of Vibration Meters JJG 676	(1~441)m/s ² , (2~6000)Hz	$U_{rel}=2.0\%$	Comparative method only	
		Speed		(0.1~1500)mm/s, (2~1000)Hz	$U_{rel}=2.0\%$		
		Displacement		(0.01~16)mm ,(2~500)Hz	$U_{rel}=2.0\%$		
		Frequency		(2~6000)Hz	$U_{rel}=0.012\%$		



No. CNAS L2336

第 42 页 共 187 页


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
69	*Spring Hammers	Force	C. S. for Spring Hammers JJF 1475	10N	$U_{rel}=0.28\%$		
		Length		(9.8~10.2)mm	$U=0.06\text{mm}$		
		Energy		(0.18~0.22)J	$U_{rel}=2.0\%$		
				(0.32~0.38)J	$U_{rel}=2.0\%$		
				(0.46~0.54)J	$U_{rel}=2.5\%$		
				(0.65~0.75)J	$U_{rel}=2.3\%$		
				(0.95~1.05)J	$U_{rel}=2.2\%$		
(1.95~2.05)J	$U_{rel}=2.1\%$						
70	Tachometers	Rotate speed	V. R.of Tachometers JJG 105	(20~30000)r/min	$U_{rel}=0.02\%$		
71	*Table Friability Surveymeters	Rotate speed	C. S. for Table Friability Surveymeters JJF(Lu) 131	(1~200)r/min	$U_{rel}=1.5\%$		
		Count		(1~100)round	$U=1$ round		
		Length		(150~160)mm	$U=0.4\text{mm}$		
72	*Flow Integration Meters	DC Electric current	V. R.of Flow Integration Meters JJG 1003	(0.1~20)mA	$U_{rel}=0.04\%$		
		DC Voltage		(0.01~5)V	$U_{rel}=0.04\%$		
		Resistance		(1~10000)Ω	$U_{rel}=0.04\%$		
		Frequency		(0.1~10)kHz	$U_{rel}=0.02\%$		
73	*Type Pressure Gauges	Pressure	V. R.of Type Pressure Gauges JJG 927	(0~2.5)MPa	$U=0.7\%FS$		
74	*Simulation Test-bed for Road Transportation	Frequency	C. S. for Simulation Test-bed for Road Transportation JJF 1271	(3~400)Hz	$U_{rel}=2.0\%$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Acceleration		(0.1~100)m/s ²	$U_{rel}=3.0\%$		
75	*Tablet Hardness Tester Needle Tube	Force	C.S. for Tablet Hardness Tester Needle Tube JJF(E) 46	1N~500N	$U_{rel}=0.1\%$		
76	*Analogue Indicating Weighing Instruments	Mass	V. R. of Analogue Indicating Weighing Instruments JJG 13	(0.02~100) kg	$U=1.2g\sim 0.5kg$		
77	*Stature and Weight Instrument	Mass	C. S for Stature and Weight Instrument JJF (E) 96	Analogue Indicating: (0.1~300) kg	$U= (0.012\sim 1.2) kg$		
				Digital Indicating: (0.1~30) kg	$U= (0.006\sim 0.29) kg$		
		Length		(20~200) cm	$U=0.12cm$		
78	Liquid Level Gauges	Length	V.R of Liquid Level Gauges JJG 971	Vertical method: (0~2000)mm	$U=0.7mm$		
				Horizontal Simulation method: (0.1~50)m	$U_{rel}=0.013\%$		
79	*Constant Revolution Speed Sources	Rotate speed	C. S. for Constant Revolution Speed Sources JJF(Ji) 146	(100~300)r/min	$U_{rel}=0.1\%$		
80	Plastic Containers	Volume	C. S. for Plastic Containers JJF(Ji) 108	(0.1~1)mL	$U= (0.001\sim 0.003) mL$		
				(1~10)mL	$U= (0.003\sim 0.010) mL$		
				(10~100)mL	$U= (0.010\sim 0.03) mL$		
				(100~500)mL	$U= (0.03\sim 0.08) mL$		
				(500~1000)mL	$U= (0.08\sim 0.14) mL$		
				(1000~2000)mL	$U= (0.14\sim 0.22) mL$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(2000~5000)mL	$U = (0.22 \sim 1.4) \text{ mL}$		
81	Bottle Top Dispenser	Volume	C. S. for Bottle Top Dispenser JJF(Ji) 181	(0.1~10)mL	$U=0.007\text{mL}$		
				(10~50)mL	$U=0.03\text{mL}$		
				(50~100)mL	$U=0.05\text{mL}$		
				(100~200)mL	$U=0.08\text{mL}$		
82	Micro sampling syringe	Volume	V. R. of Micro sampling syringe JJG(Ji) 166	(0.1~1) μL	$U=0.016 \mu\text{L}$		
				(1~25) μL	$U=0.08 \mu\text{L}$		
				(25~100) μL	$U=0.2 \mu\text{L}$		
				(100~1000) μL	$U=1.0 \mu\text{L}$		
83	Soil meter	Density	V. R. of Working Glass Hydrometers JJG 42	Type I: (-5~50) s°	$U=0.2s^\circ$	Only use direct comparison method.	
				Type II: 0.995~1.030	$U=0.0005$		
84	*Abrasion Resistance Instruments	Length	C. S. for Abrasion Resistance Instruments JJF(Zhe)1070	(0.01~25) mm	$U=0.01\text{mm}$		
		Angle		(80~100) $^\circ$	$U=0.2^\circ$		
		Rotational speed		(100~300) r/min	$U_{\text{rel}}=0.2\%$		
		Mass		(500~2000) g	$U=0.1\text{g}$		
		Time		(0.1~1200) s	$U=0.3\text{s}$		
85	Pile Dynamic Measuring Instrument	Acceleration	V. R. of Pile Dynamic Measuring Instrument JJG 930	(2~450)m/s ² , (2~5000)Hz	$U_{\text{rel}}=2.0\%$		




No. CNAS L2336

第 45 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
86	*Concrete Batching Scales	Mass	V. R.of Concrete Batching Scales JJG 1171	(0.2~3000) kg	U= (0.01~0.7) kg		
87	*Asphalt Mixture and Cement Concrete Mixing Device Measure System	Mass	V.R. of Asphalt Mixture and Cement Concrete Mixing Device Measure System JJG (JT) 071	(2~3000) kg	U _{rel} =0.08%	Only for (static) of the indication of the measuring device for the preparation of component materials	
88	*Density of Abrasive Grains Testers	Volume	C. S. for Density of Abrasive Grains Testers JJF (JX) 1076	(195~205) cm ³	U=0.08cm ³	Only for ordinary abrasive bulk density measuring instrument	
		Length		(10~300) mm	U=0.03mm		
89	*laboratory oscillation-type liquid density meters	Density	Calibration Specification for Laboratory Oscillation-type Liquid Density Meters JJF 2165	(650~2000)kg/m ³	U=0.06kg/m ³		
90	*AutomaticPipetting Workstation	Volume	Calibration Specification for AutomaticPipetting Workstation JJF(Ji) 233	(1~5) μ L	U=(0.06~0.15) μ L		
				(5~10) μ L	U=(0.15~0.24) μ L		
				(10~100) μ L	U=(0.24~0.6) μ L		
				(100~1000) μ L	U=(0.6~2.6) μ L		
91	*Automated Dilution and	Volume	Calibration Specification for Automated Dilution and	(1~10) μ L	U=(0.06~0.32) μ L		




在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Standard Solution Preparation Instrument		Standard Solution Preparation Instrument JJF(Ji) 189	(10~100) μ L	$U=(0.32\sim 0.7)\mu L$		
				(100~1000) μ L	$U=(0.7\sim 3.0)\mu L$		
				(1000~10000) μ L	$U=(3.0\sim 12)\mu L$		
92	*Elevator Overspeed Governor Testers	Speed	Calibration Specification for Elevator Overspeed Governor Testers JJF 1374	(0.1~10)m/s	$U_{rel}=0.18\%$		
93	*Helium Mass Spectrometer Leak Detector	Leak rate	Calibration Specification of Helium Mass Spectrometer Leak Detector JJF(JG) 186	$(5\times 10^{-10}\sim 5\times 10^{-5})Pa\cdot m^3/s$	$U_{rel}=13\%$	Using Only the Vacuum Standard Leakage Set Method	
		Calibration coefficient		0.5~1.5	$U_{rel}=13\%$		
94	LX-C Micropore Material Durometers	Length	Calibration Specification for LX-C Micropore Material Durometers JJF(Su) 285	(0.02~2.52)mm	$U=0.006mm$		
		Force		(0.5~8.5)N	$U=0.03N$		
95	AO Durometers	Hardness	Calibration Specification for Shore AO Durometers JJF 1312	(0~100)HAO	$U=0.3HAO$		
		Length		(0.02~5.60)mm	$U=(0.006\sim 0.02)mm$		
		Force		(0.5~8.1)N	$U=0.03N$		
96	*Filter Integrity Testers	Pressure	Calibration Specification for Filter Integrity Testers JJF 2142	(0~1000)kPa	$U=0.15\%FS$		
		Flow		(0.8~100)mL/min	$U_{rel}=1.4\%$		
97	*Screen Tension Meter Calibration Instruments	Force	Calibration Specification of Screen Tension Meter Calibration Instruments JJF(Su) 304	(100~1500)N	$U_{rel}=0.15\%$		
		Length		(100~300)mm	$U=0.15mm$		
		Tension		(7~50)N/cm	$U=0.15\%FS$		
98	*Mass Comparators	Weight	Calibration Specification for Mass Comparators JJF 1326	Repeatability: 1g~100g	$U=(0.1\sim 1.6)\mu g$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Repeatability: 1mg~1000g	$U=(1.6\sim 23)\mu g$		
				Load unbalance error: 1mg~1000g	$U=(1.0\sim 6.0)\mu g$		
				Load unbalance error: 100g~1kg	$U=(6.0\sim 46)\mu g$		
				Local indication error: 1g~100g	$U=(2.1\sim 18)\mu g$		
				Local indication error: 100g~1kg	$U=(18\sim 60)\mu g$		
99	Aneroid Barometers and Aneroid Barographs	Pressure	Verification Regulation of Aneroid Barometers and Aneroid Barographs JJG 272	(500~1060)hPa	$U=0.7\text{hPa}$		
Electromagnetism measuring instruments(including magnetism)							
1	*Charge Amplifiers	Voltage	V.R.of Charge Amplifiers JJG338	10mV~10V	$U_{\text{rel}}=0.1\%$		
		Amplitude Frequency Characteristic		10mV~10V,(0.01~20)kHz	$U_{\text{rel}}=0.1\%$		
		Distortion factor		0.01%~100%	$U_{\text{rel}}=5\%$		
2	*Instantaneous Discontinuity Testing Instruments	Resistance	Calibration Specification of Instantaneous Discontinuity Testing Instruments JJF(DZ)0042	10mΩ~99.99Ω	$U_{\text{rel}}=1.0\%$		
		Time		(0.01~99.99)μs	$U_{\text{rel}}=0.2\%$		
3	*Process Calibrators	IN-DC resistor	Calibration specification for process calibrators JJF 1472	(10~400)Ω	$U_{\text{rel}}=0.012\%$		
				(400~4000)Ω	$U_{\text{rel}}=0.010\%$		
		IN-DC		(0.2~2)mA	$U_{\text{rel}}=0.013\%$		




No. CNAS L2336

第 48 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		current		(2~100)mA	$U_{rel}=0.015\%$		
		IN-DC voltage		(10~1000)mV	$U_{rel}=0.007\%$		
				(1~50)V	$U_{rel}=0.006\%$		
		IN-AC voltage		(10~1000)mV (10Hz~20kHz)	$U_{rel}=0.02\%$		
				(1~20)V (10Hz~20kHz)	$U_{rel}=0.01\%$		
		IN-Frequency		10Hz~50kHz	$U_{rel}=0.009\%$		
		IN-Temperature(TC)		(-100~1300)°C	$U=0.4^{\circ}\text{C}$		
		IN-Temperature(RTD)		(-100~800)°C	$U=0.3^{\circ}\text{C}$		
		OUT-DC resistor		(10~400) Ω	$U_{rel}=0.007\%$		
				(400~10000) Ω	$U_{rel}=0.006\%$		
		OUT-DC current		(0.2~2)mA	$U_{rel}=0.008\%$		
				(2~24)mA	$U_{rel}=0.009\%$		
		OUT-DC voltage		(10~1000)mV	$U_{rel}=0.006\%$		
				(1~100)V	$U_{rel}=0.005\%$		
		IN-Temperature(TC)		(-100~1300)°C	$U=0.2^{\circ}\text{C}$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		IN-Temperature (RTD)		(-100~800)°C	$U=0.2^{\circ}\text{C}$		
4	*DC Standard Voltage Source	DC Voltage	Verification regulation of DC Standard Voltage Source JIG (JG) 193	10mV~330mV	$U_{\text{rel}}=0.0019\%$		
				0.33V~3.3V	$U_{\text{rel}}=0.0015\%$		
				3.3V~330V	$U_{\text{rel}}=0.0015\%$		
				330V~1000V	$U_{\text{rel}}=0.0016\%$		
5	*DC Voltage Stable Source	DC Voltage	DC Voltage Stable Source JIG(JG) 77	0.1V~100V	$U_{\text{rel}}=0.003\%$		
				100V~1000V	$U_{\text{rel}}=0.002\%$		
		DC Current		0.01A~20A	$U_{\text{rel}}=0.02\%$		
				20A~200A	$U_{\text{rel}}=0.03\%$		
		Regulation rate of regulated power supply		0.01%~10%	$U_{\text{rel}}=0.006\%$		
				0.01%~10%	$U_{\text{rel}}=0.01\%$		
		Ripple voltage		0.1mV~1000mV	$U_{\text{rel}}=2\%$		
6	*Precise AC Voltage Calibration Source	AC Voltage	V.R.of Precise AC Voltage Calibration Source JIG 410	2V~400V(45Hz~1kHz)	$U_{\text{rel}}=0.2\%$		
		Frequency		40Hz~1kHz	$U_{\text{rel}}=0.015\%$		
		Distortion factor		0.01%~10%	$U=0.3\%$		




No. CNAS L2336

第 50 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
7	*DC Digital Voltmeter	DC Voltage	Verification regulation of DC Digital Voltmeter JJG (JG) 198	10mV~100mV	$U_{rel}=0.009\%$		
				0.1V~10V	$U_{rel}=0.0015\%$		
				10V~100V	$U_{rel}=0.0016\%$		
				100V~1000V	$U_{rel}=0.0017\%$		
8	*AC Digital Voltmeter	AC Voltage	V.R.of AC Digital Voltmeter JJG(JG)72	10mV~100mV (50Hz)	$U_{rel}=0.11\%$		
				10mV~100mV (1kHz)	$U_{rel}=0.11\%$		
				0.1V~100V (50Hz)	$U_{rel}=0.032\%$		
				0.1V~100V (1kHz)	$U_{rel}=0.032\%$		
				100V~1000V (50Hz)	$U_{rel}=0.038\%$		
				100V~1000V (1kHz)	$U_{rel}=0.038\%$		
9	*D.C.Potentiometers	DC Voltage	Verification Regulation Of D.C.Potentiometers JJG 123	10mV~200mV	$U_{rel}=0.01\%$		
				0.2V~2V	$U_{rel}=0.003\%$		
10	*Digital High Voltmeter	DC Voltage	Verification Regulation of Digital High Voltmeter DL/T973	0.1kV~50kV	$U_{rel}=0.1\%$		
		AC Voltage		0.1kV~50kV (50Hz,60Hz)	$U_{rel}=0.2\%$		
11	High Voltage Electrostatic Voltmeters	DC Voltage	V.R.of High Voltage Electrostatic Voltmeters JJG 494	0.1kV~50kV	$U_{rel}=0.1\%$		
		AC Voltage		0.1kV~50kV (50Hz,60Hz)	$U_{rel}=0.2\%$		
12	*DC Resistive Volt Ratio Box	DC Voltage	V.R.of the DC Resistive Volt Ratio Box JJG 531	(0.01~1000)V	$U_{rel}=0.005\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
13	*High-voltage Divider At Power Frequency	AC Voltage	Verification Regulation of High-voltage Divider At Power Frequency JJG 496	0.1kV~50kV(50Hz)	$U_{rel}=0.2\%$		
14	DC High voltage Divider	DC Voltage	Verification Regulation of DC High voltage Dividers JJG 1007	0.1kV~50kV	$U_{rel}=0.1\%$		
15	*AC Voltage Stable Source	Frequency	Calibration specification for Stable Character of AC Voltage Stable Source JJF(JG) 85	10Hz~40Hz	$U_{rel}=0.03\%$		
				40Hz~1kHz	$U_{rel}=0.012\%$		
		AC Voltage		0.1V~300V (40Hz~1kHz)	$U_{rel}=0.12\%$		
		Rate of adjustment of supply voltage		0.01%~10%	$U_{rel}=0.12\%$		
		Load adjustment rate		0.01%~10%	$U_{rel}=0.12\%$		
		Degree of distortion		0.01%~10%	$U=1.2dB$		
		Efficiency		0.1%~100%	$U=0.5\%$		
16	contactless electrostatic voltage measuring instrument	Voltage	Calibration specification for contactless electrostatic voltage measuring instrument JJF 1517	0.1kV~20kV	$U_{rel}=3\%$		
17	*Withstanding Voltage Testers	DC Voltage	V.R. of Withstanding Voltage Testers JJG 795	0.1kV~15kV	$U_{rel}=0.6\%$		
		AC voltage		0.1kV~15kV(50Hz, 60Hz)	$U_{rel}=0.6\%$		
		DC Current		0.01mA~200mA	$U_{rel}=0.6\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC current		0.01mA ~ 200mA(50Hz,60Hz)	$U_{rel}=0.6\%$		
		Time		(0.01 ~ 999.99)s	$U=0.8\%T+0.03s$		
18	*DC Standard Current Source	DC Current	V.R.for Direct Current Standard Source JJG(JG)-69	0.01A ~ 20A	$U_{rel}=0.05\%$		
				20A ~ 200A	$U_{rel}=0.07\%$		
				200A ~ 2000A	$U_{rel}=0.12\%$		
19	*DC Digital Amperemeter	DC Current	Verification regulation of DC Digital Amperemete JJG (JG) 197	10 μ A ~ 100 μ A	$U_{rel}=0.11\%$		
				0.1mA ~ 100mA	$U_{rel}=0.012\%$		
				0.1A ~ 1A	$U_{rel}=0.025\%$		
				1A ~ 3A	$U_{rel}=0.031\%$		
				3A ~ 10A	$U_{rel}=0.058\%$		
				10A ~ 20A	$U_{rel}=0.077\%$		
20	*AC Digital Amperemeter	AC Current	V.R.of AC Digital Ampere-meter JJG(JG)68	29 μ A ~ 100 μ A (50Hz)	$U_{rel}=0.43\%$		
				29 μ A ~ 100 μ A (1kHz)	$U_{rel}=0.38\%$		
				0.1mA ~ 1mA (50Hz)	$U_{rel}=0.27\%$		
				0.1mA ~ 1mA (1kHz)	$U_{rel}=0.26\%$		
				1mA ~ 100mA (50Hz)	$U_{rel}=0.071\%$		
				1mA ~ 100mA (1kHz)	$U_{rel}=0.070\%$		
				100mA ~ 1A (50Hz)	$U_{rel}=0.075\%$		
				100mA ~ 1A (1kHz)	$U_{rel}=0.074\%$		




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION FOR COMPETENCY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATION	1A~3A (50Hz)	$U_{rel}=0.078\%$		
				1A~3A (1kHz)	$U_{rel}=0.077\%$		
				3A~10A (50Hz)	$U_{rel}=0.16\%$		
				3A~10A (1kHz)	$U_{rel}=0.15\%$		
				10A~20A (50Hz)	$U_{rel}=0.19\%$		
				10A~20A (1kHz)	$U_{rel}=0.17\%$		
21	*Clamp Ammeters	DC Current	C. S. of Clamp Ammeters JJF 1075	analog:1A~20A	$U_{rel}=0.4\%$		
				analog:2A~1000A	$U_{rel}=0.6\%$		
				digitale: 1A~20A	$U_{rel}=0.3\%$		
				digitale: 20A~1000A	$U_{rel}=0.6\%$		
		AC Current		analog:1A~20A (45Hz~60Hz)	$U_{rel}=0.5\%$		
				analog:20A~1000A (45Hz~60Hz)	$U_{rel}=0.7\%$		
				digitale: 1A~ 20A(45Hz~60Hz)	$U_{rel}=0.3\%$		
				analog:20A~1000A (45Hz~60Hz)	$U_{rel}=0.6\%$		
22	*Magnetic Yoke Detectors	DC current	Calibration specification of Magnetic Yoke Detectors JJF 1458	(0~20)A	$U=0.3A$		
		Lifting Power		(1~300)N	$U_{rel}=1\%$		
23	*instrument Current Transformers	AC Current	Verification Regulation of instrument Current Transformers JJG 313	(5~5000)/5A (5%~ 120%)(50Hz)	$U_{rel}=0.07\%$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		phase		(0~900)', 5%~120%(50Hz)	U=2'		
24	Leakage Current Meter	AC Current	V.R. of Leakage Current Testers JJG 843	(0.1~2)mA (45Hz~60Hz)	U=0.004mA		
		AC Voltage		(2~20)mA (45Hz~60Hz)	U=0.04mA		
		DC Current		(0.1~250)V (45Hz~60Hz)	U=0.8V		
		DC Voltage		(0.1~2)mA	U=0.004mA		
		Input Resistance		(2~20)mA	U=0.04mA		
				(0.1~250)V	U=0.8V		
				100 Ω ~ 1M Ω	U _{rel} =0.3%		
25	*Electrolytic Capacitor Leakage Current Meter	DC Voltage	Electrolytic Capacitor Leakage Current Meter JJG(DZ) 306003	10V~1000V	U _{rel} =0.08%		
		DC Current		0.1μA~200μA	U _{rel} =0.6%		
				200μA~30mA	U _{rel} =0.4%		
26	Hall effect current sensor	DC Current	Calibration specification of Hoare current sensor JJF(machinery) 1067	(1 ~100)A	U _{rel} =0.02%		
				(100 ~2000)A	U _{rel} =0.03%		
		AC Current		(1 ~100)A, (45~60)Hz	U _{rel} =0.07%		
				(100 ~2000)A, (45~60)Hz	U _{rel} =0.08%		
				(2000 ~6000)A, (45~60)Hz	U _{rel} =0.6%		
27	*DC Standard Resistors	Resistance	V.R. of DC Standard Resistors JJG 166	0.01 Ω ~ 0.1 Ω	U _{rel} =0.2%		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.1 Ω ~ 1 Ω	$U_{rel}=0.05\%$		
				1 Ω ~ 10 Ω	$U_{rel}=0.01\%$		
				10 Ω ~ 100 Ω	$U_{rel}=0.004\%$		
				100 Ω ~ 100k Ω	$U_{rel}=0.002\%$		
				100k Ω ~ 1M Ω	$U_{rel}=0.005\%$		
				1M Ω ~ 10M Ω	$U_{rel}=0.01\%$		
28	*DC Resistance Box	Resistance	V. R.of D.C. Resistance Boxes JJG 982	0.01 Ω ~ 0.1 Ω	$U_{rel}=0.2\%$		
				0.1 Ω ~ 1 Ω	$U_{rel}=0.05\%$		
				1 Ω ~ 10 Ω	$U_{rel}=0.01\%$		
				10 Ω ~ 100 Ω	$U_{rel}=0.005\%$		
				100 Ω ~ 100000 Ω	$U_{rel}=0.002\%$		
29	*DC Digital Ohmmeter	Resistance	Verification regulation of DC Digital Ohmmeter JJG (JG) 199	1 Ω ~ 10 Ω	$U=0.0045\%R_x+0.01 \Omega$		
				10 Ω ~ 100 Ω	$U=0.0035\%R_x+0.015 \Omega$		
				100 Ω ~ 100k Ω	$U=0.0035\%R_x+1 \Omega$		
				100k Ω ~ 1M Ω	$U=0.0035\%R_x+10 \Omega$		
				1M Ω ~ 10M Ω	$U=0.0015\%R_x+0.15k \Omega$		
				10M Ω ~ 100M Ω	$U=0.0055\%R_x+3k \Omega$		
30	*DC Low Resistance Meters	Resistance	V.R.of DC Low Resistance Meters JJG 837	(1~20) μ Ω	$U_{rel}=1.3\%$		
				(20~200) μ Ω	$U_{rel}=0.6\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(0.2~2)mΩ	$U_{rel}=0.13\%$		
				(2~20) mΩ	$U_{rel}=0.06\%$		
				20mΩ ~20kΩ	$U_{rel}=0.013\%$		
				20kΩ ~1MΩ	$U_{rel}=0.024\%$		
				1mΩ , 10mΩ , 100mΩ , 1Ω , 10Ω , 100Ω , 1kΩ , 10kΩ , 100kΩ	$U_{rel}=0.012\%$		
31	*D.C Bridges	Resistance	V.R.of D.C Bridges JIG 125	(0.01~0.1)Ω	$U_{rel}=6\%$		
				(0.1~1)Ω	$U_{rel}=1.2\%$		
				(1~10)Ω	$U_{rel}=0.12\%$		
				(10~100)Ω	$U_{rel}=0.04\%$		
				100Ω ~100kΩ	$U_{rel}=0.03\%$		
32	*Insulation Resistance	Insulation Resistance	V.R. of Megohmmeter JIG 622	0.1MΩ ~10MΩ	$U_{rel}=0.3\%$		
				10MΩ ~100MΩ	$U_{rel}=0.6\%$		
				100MΩ ~1GΩ	$U_{rel}=1.2\%$		
				1GΩ ~10GΩ	$U_{rel}=3\%$		
		Terminal voltage	50V~10000V	$U_{rel}=0.3\%$			
33	*Electronic Insulating Resistance Meters	Insulation Resistance	V.R. of Electronic Insulating Resistance meters JIG 1005	0.1MΩ ~10MΩ	$U_{rel}=0.3\%$		
				10MΩ ~100MΩ	$U_{rel}=0.6\%$		



No. CNAS L2336

第 57 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
34	*High Insulation Resistance Meters	Charging Voltage	C.S. for High Insulation Resistance Meters JJF 2225	100M Ω ~ 1G Ω	$U_{rel}=1.2\%$		
				1G Ω ~ 1000G Ω	$U_{rel}=3\%$		
				10V ~ 10000V	$U_{rel}=0.6\%$		
		Resistance		1M Ω ~ 100M Ω	$U_{rel}=0.6\%$		
				100M Ω ~ 1G Ω	$U_{rel}=1.2\%$		
				1G Ω ~ 1000G Ω	$U_{rel}=3\%$		
Terminal voltage	25V ~ 1kV	$U_{rel}=0.3\%$					
35	*Earth Resistance Meters	Resistance	V.R. of Earth Resistance Meters JJG 366	0.01 Ω ~ 0.1 Ω	$U_{rel}=7\%$		
				0.1 Ω ~ 1 Ω	$U_{rel}=1.2\%$		
				1 Ω ~ 1000 Ω	$U_{rel}=0.7\%$		
36	*Earth-Continuity Testers	Resistance	V.R. of Earth-Continuity Testers JJG 984	(0.01 ~ 0.1)m Ω	$U_{rel}=12\%$		
				(0.1 ~ 1)m Ω	$U_{rel}=5\%$		
				(1 ~ 10)m Ω	$U_{rel}=0.8\%$		
				(10 ~ 1000)m Ω	$U_{rel}=0.7\%$		
		Current		(0.1 ~ 60)A	$U_{rel}=0.4\%$		
37	*Clamp Earth Resistance Meters	Resistance	V.R. of Clamp Earth Resistance Meters JJG 1054	0.100 Ω ~ 9.999 Ω	$U_{rel}=0.6\%$		
				10.00 Ω ~ 99.99 Ω	$U_{rel}=0.5\%$		
				100.0 Ω ~ 199.9 Ω	$U_{rel}=0.7\%$		



No. CNAS L2336

第 58 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				200 Ω ~ 400 Ω	$U_{rel}=1\%$		
				400 Ω ~ 600 Ω	$U_{rel}=3\%$		
				600 Ω ~ 1111 Ω	$U_{rel}=5\%$		
38	*Wrist Strap Tester	Insulation Resistance	C.S. for Wrist Strap Tester JJF(DZ) 31502	0.01M Ω ~ 10M Ω	$U_{rel}=1.5\%$		
				10M Ω ~ 100M Ω	$U_{rel}=2\%$		
				100M Ω ~ 1G Ω	$U_{rel}=3\%$		
39	*Resistance	Resistance	C.S. for Surface Resistance Tester JJF 1285	0.001M Ω ~ 10M Ω	$U_{rel}=0.3\%$		
				10M Ω ~ 100M Ω	$U_{rel}=0.6\%$		
				100M Ω ~ 1G Ω	$U_{rel}=1.2\%$		
				1G Ω ~ 1000G Ω	$U_{rel}=3\%$		
		Open circuit voltage		10V ~ 250V	$U_{rel}=0.3\%$		
40	*Loop Resistance Meters	Resistance	V.R. of Loop Resistance Testers JJG 1052	10 μ Ω ~ 2m Ω	$U_{rel}=0.4\%$		
				2m Ω ~ 200m Ω	$U_{rel}=0.2\%$		
		Current		(10~100)A	$U_{rel}=0.3\%$		
41	DC Shunt	Resistance	V. R. of DC Shunts JJG 1069	(0.001 ~ 10) Ω	$U_{rel}=0.08\%$		
42	*Air Ionizer	Voltage	Periodic Verification of Air Ionizers	(1 ~ 1000)V	$U=3V$		
		Time	ANSI/ESDSP3.3, Calibration Specification for Ionizer JJF(DZ) 0110	(0.1 ~ 999.9)s	$U=0.2s$		




No. CNAS L2336

第 59 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
43	*Charged Plate Monitor	Electrostatic voltage	Calibration Regulation of Charged Plate Monitor	(0.1~1000)V	$U_{rel}=0.3\%$		
		decay time	JJF(DZ) 31003	(0.01~60)s	$U_{rel}=0.1\%$		
44	*Standard Digital Power Meters	AC Digital Powermeter	Calibration Specification for Standard Digital Power Meters JJF 2226	0.1W~20W (45Hz~65Hz)	$U_{rel}=0.10\%$	Only calibrate AC Digital Powermeter	
				20W~500W(45Hz~65Hz)	$U_{rel}=0.12\%$		
				500W~20kW(45Hz~65Hz)	$U_{rel}=0.15\%$		
		AC Digital Voltmeter		(1~600)V(45Hz~65Hz)	$U_{rel}=0.08\%$		
		AC Digital Ampere-meter		10mA~20A (45Hz~65Hz)	$U_{rel}=0.09\%$		
45	*Digital electric parameters test meter	Frequency	Calibration Specification for Digital AC Electrical Parameters Meter JJF 1491	40Hz~1000Hz	$U_{rel}=0.03\%$		
		AC Voltage		(1~600)V (40Hz~1kHz)	$U_{rel}=0.08\%$		
		AC Current		10mA~20A (40Hz~1kHz)	$U_{rel}=0.09\%$		
		AC Powermeter		500W~20kW (40Hz~1kHz)	$U_{rel}=0.15\%$		
46	*Industry Frequency Singlephase Phase Meter	Phase	V.R.of Industry Frequency Singlephase Phase Meter JJG 440	Phase: (0~360)° (50Hz)	$U=0.15^\circ$		
				Power factor: (0~1.000)(50Hz)	$U=0.001$		
47	*AC Impedance Parameter Testers	AC Voltage	C. S. of AC Impedance Parameter Testers JJF(Zhe) 1083	(10~1000)V (45~65)Hz	$U_{rel}=0.10\%$		
		AC Current		(0.1~20)A (45~65)Hz	$U_{rel}=0.12\%$		
		AC Power		1W~10kW(45~65)Hz	$U_{rel}=0.15\%$		




No. CNAS L2336

第 60 页 共 187 页


The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Frequency		(45~65)Hz	$U_{rel}=0.02\%$		
		Phase		$0^{\circ} \sim 359.99^{\circ}$	$U=0.02^{\circ}$		
48	*Motor Stator Testing System	DC Resistance	Calibration Specification for Motor Stator Testing System JJF(Min)1060	$1m\Omega \sim 2m\Omega$	$U_{rel}=0.13\%$		
				$(2\sim 20)m\Omega$	$U_{rel}=0.06\%$		
				$20m\Omega \sim 20k\Omega$	$U_{rel}=0.02\%$		
		AC Voltage		$(0.1\sim 3)kV(45\sim 65)Hz$	$U_{rel}=0.7\%$		
		AC Current		$(0.1\sim 100)mA(45\sim 65)Hz$	$U_{rel}=0.8\%$		
		Insulation Resistance		$(0.1\sim 10)M\Omega$	$U_{rel}=0.3\%$		
				$(10\sim 100)M\Omega$	$U_{rel}=0.7\%$		
				$(100\sim 1000)M\Omega$	$U_{rel}=0.8\%$		
		DC Voltage		$(500\sim 1000)V$	$U_{rel}=0.8\%$		
		Pulse voltage		$(0.1\sim 3)kV(45\sim 65)Hz$	$U_{rel}=1.2\%$		
Front time	$(0.1\sim 1.2)\mu s$	$U_{rel}=1.0\%$					
49	*Multifunction Calibrator	DC Voltage	Calibration Specification for Multifunction Standard Sources JJF 1638	$10mV\sim 200mV$	$U=0.0011\%V_x+1.0\mu V$		
				$200mV\sim 2V$	$U=0.0010\%V_x+1.4\mu V$		
				$2V\sim 10V$	$U=0.00096\%V_x+4.8\mu V$		
				$10V\sim 20V$	$U=0.0010\%V_x+10\mu V$		
				$20V\sim 200V$	$U=0.0011\%V_x+0.12mV$		
				$200V\sim 1000V$	$U=0.0013\%V_x+0.7mV$		




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Voltage	ilac-M	10mV~200mV(40Hz~20kHz)	$U=0.013\%V_x+12\mu V$		
				10mV~200mV(20kHz~50kHz)	$U=0.043\%V_x+12\mu V$		
				10mV~200mV(50kHz~100kHz)	$U=0.11\%V_x+36\mu V$		
				10mV~200mV(100kHz~300kHz)	$U=0.13\%V_x+36\mu V$		
				10mV~200mV(300kHz~500kHz)	$U=0.22\%V_x+48\mu V$		
				0.2V~2V(40Hz~20kHz)	$U=0.01\%V_x+8\mu V$		
				0.2V~2V(20kHz~50kHz)	$U=0.017\%V_x+24\mu V$		
				0.2V~2V(50kHz~100kHz)	$U=0.034\%V_x+96\mu V$		
				0.2V~2V(100kHz~300kHz)	$U=0.058\%V_x+0.18mV$		
				0.2V~2V(300kHz~500kHz)	$U=0.14\%V_x+0.5mV$		
				2V~20V(40Hz~20kHz)	$U=0.01\%V_x+84\mu V$		
				2V~20V(20kHz~50kHz)	$U=0.017\%V_x+0.24mV$		
				2V~20V(50kHz~100kHz)	$U=0.034\%V_x+0.5mV$		
				2V~20V(100kHz~300kHz)	$U=0.072\%V_x+1.7mV$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
				2V~20V(300kHz~500kHz)	$U=0.17\%V_x+6mV$			
					20V~200V(40Hz~20kHz)	$U=0.011\%V_x+1.2mV$		
					20V~200V(20kHz~50kHz)	$U=0.03\%V_x+5mV$		
					20V~200V(50kHz~100kHz)	$U=0.072\%V_x+12mV$		
					200V~1000V(40Hz~10kHz)	$U=0.0119\%V_x+5mV$		
		DC Current		10 μ A~200 μ A	$U=0.007\%I_x+11nA$			
					200 μ A~2mA	$U=0.007\%I_x+11nA$		
					2mA~20mA	$U=0.007\%I_x+0.11 μ A$		
					20mA~200mA	$U=0.007\%I_x+11 μ A$		
					200mA~2A	$U=0.01\%I_x+33 μ A$		
		AC Current		2A~10A	$U=0.04\%I_x+0.53mA$			
					10A~20A	$U=0.11\%I_x+0.83mA$		
					(10~20)mA 40Hz~1kHz	$U=0.018\%I_x+0.4 μ A$		
					(10~20)mA (1~5)kHz	$U=0.077\%I_x+6 μ A$		
					(10~20)mA 5kHz~10kHz	$U=0.2\%I_x+11 μ A$		
				(20~200)mA 40Hz~1kHz	$U=0.02\%I_x+5 μ A$			



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Current	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(20~200)mA 1kHz~5kHz	$U=0.08\%Ix+60 \mu A$		
				(20~200)mA 5kHz~10kHz	$U=0.2\%Ix+0.11mA$		
				(0.20~2)A 40Hz~1kHz	$U=0.083\%Ix+44 \mu A$		
				(0.2~2)A 1kHz~5kHz	$U=0.094\%Ix+0.11mA$		
				(0.2~2)A 5kHz~10kHz	$U=1.1\%Ix+0.22mA$		
				(2~10)A 40Hz~1kHz	$U=0.051\%Ix+0.19mA$		
				(10~20)A 45Hz~100Hz	$U=0.13\%Ix+6mA$		
				(10~20)A 100Hz~1kHz	$U=0.17\%Ix+6mA$		
				(10~20)A 1kHz~5kHz	$U=0.33\%Ix+6mA$		
		1 Ω ~ 10 Ω		$U_{rel}=0.022\%$			
		10 Ω ~ 100 Ω		$U_{rel}=0.0054\%$			
		100 Ω ~ 10k Ω		$U_{rel}=0.0014\%$			
		10k Ω ~ 100k Ω		$U_{rel}=0.016\%$			
		100k Ω ~ 1M Ω		$U_{rel}=0.0028\%$			
		1M Ω ~ 10M Ω		$U_{rel}=0.005\%$			
10M Ω ~ 100M Ω	$U_{rel}=0.015\%$						
50	*Calibrators for Electrical Meters	DC voltage	Calibration Sepecification of Calibrators for Electrical Meters JJF 1284	(10~200)mV	$U_{rel}=0.003\%$	合格评定 国家认可 委员会 认可证书专用章	
				(0.2~1000)V	$U_{rel}=0.002\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC voltage	ilac-M	10mV~1000V, (40Hz~2kHz)	$U_{rel}=0.04\%$		
		DC current		10 μ A~20A	$U_{rel}=0.05\%$		
		AC current		(20~200) μ A(40Hz~2kHz)	$U_{rel}=0.1\%$		
				(0.2~200)mA,(40Hz~2kHz)	$U_{rel}=0.09\%$		
				(0.2~2)A(40Hz~2kHz)	$U_{rel}=0.15\%$		
				(2~20)A(40Hz~2kHz)	$U_{rel}=0.4\%$		
		resistance		(1~20) Ω	$U_{rel}=0.01\%$		
				20 Ω ~2M Ω	$U_{rel}=0.003\%$		
				(2~20)M Ω	$U_{rel}=0.004\%$		
				(20~200)M Ω	$U_{rel}=0.1\%$		
		Frequency		40Hz~10kHz	$U_{rel}=6 \times 10^{-6}$		
		51		*Digital Multimeter	DC Voltage		
220mV~2.2V	$U=0.0008\%V_x+1.2 \mu V$						
2.2V~11V	$U=0.0008\%V_x+4 \mu V$						
11V~22V	$U=0.0008\%V_x+8 \mu V$						
22V~220V	$U=0.0009\%V_x+100 \mu V$						
220V~1000V	$U=0.0011\%V_x+0.6mV$						
AC Voltage	(10~220)mV 40Hz~20kHz		$U=0.011\%V_x+10 \mu V$				




No. CNAS L2336

第 65 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(10~220)mV 20kHz~50kHz	$U=0.036\%V_x+10\mu V$		
				(10~220)mV 50kHz~100kHz	$U=0.090\%V_x+30\mu V$		
				(10~220)mV 100kHz~300kHz	$U=0.11\%V_x+30\mu V$		
				(10~220)mV 300kHz~500kHz	$U=0.18\%V_x+40\mu V$		
				(10~220)mV 500kHz~1MHz	$U=0.36\%V_x+0.1mV$		
				(0.22~2.2)V 40Hz~20kHz	$U=0.0085\%V_x+7\mu V$		
				(0.22~2.2)V 20kHz~50kHz	$U=0.014\%V_x+20\mu V$		
				(0.22~2.2)V 50kHz~100kHz	$U=0.028\%V_x+80\mu V$		
				(0.22~2.2)V 100kHz~300kHz	$U=0.048\%V_x+0.15mV$		
				(0.22~2.2)V 300kHz~500kHz	$U=0.12\%V_x+0.4mV$		
				(0.22~2.2)V 500kHz~1MHz	$U=0.24\%V_x+1mV$		
				(2.2~22)V 40Hz~20kHz	$U=0.0085\%V_x+70\mu V$		
				(2.2~22)V 20kHz~50kHz	$U=0.014\%V_x+0.2mV$		
				(2.2~22)V 50kHz~100kHz	$U=0.028\%V_x+0.4mV$		
				100V~1000V(45Hz~10kHz)	$U_{rel}=2.0\times 10^{-4}\sim 9\times 10^{-4}$		




在线扫码获取验证

No. CNAS L2336

第 66 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Current	ilac-M	(2.2~22)V 100kHz~300kHz	$U=0.060\%V_x+1.7mV$		
				(2.2~22)V 300kHz~500kHz	$U=0.14\%V_x+5mV$		
				(2.2~22)V 500kHz~1MHz	$U=0.30\%V_x+9mV$		
				(22~220)V 40Hz~20kHz	$U=0.0090\%V_x+1mV$		
				(22~220)V 20kHz~50kHz	$U=0.025\%V_x+4mV$		
				(22~220)V 50kHz~100kHz	$U=0.060\%V_x+10mV$		
				(22~220)V 100kHz~300kHz	$U=0.16\%V_x+0.11V$		
				(22~220)V 300kHz~500kHz	$U=0.54\%V_x+0.11V$		
				(22~220)V 500kHz~1MHz	$U=1.3\%V_x+0.22V$		
				(220~1000)V 40Hz~1kHz	$U=0.009\%V_x+4mV$		
				10 μ A~220 μ A	$U=0.006\%I_x+10nA$		
				220 μ A~2.2mA	$U=0.0060\%I_x+10nA$		
				2.2mA~22mA	$U=0.0060\%I_x+100nA$		
				22mA~220mA	$U=0.0070\%I_x+1 μ A$		
				220mA~2.2A	$U=0.0095\%I_x+30 μ A$		
				2.2A~11A	$U=0.036\%I_x+0.48mA$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Current		11A~20A	$U=0.1\%Ix+0.75mA$		
				(1~22)mA 40Hz~1kHz	$U=0.016\%Ix+0.4 \mu A$		
				(1~22)mA 1kHz~5kHz	$U=0.070\%Ix+5 \mu A$		
				(1~22)mA 5kHz~10kHz	$U=0.18\%Ix+10 \mu A$		
				(22~220)mA 40Hz~1kHz	$U=0.018\%Ix+4 \mu A$		
				(22~220)mA 1kHz~5kHz	$U=0.070\%Ix+50 \mu A$		
				(22~220)mA 5kHz~10kHz	$U=0.18\%Ix+0.1mA$		
				(0.22~2.2)A 40Hz~1kHz	$U=0.075\%Ix+40 \mu A$		
				(0.22~2.2)A 1kHz~5kHz	$U=0.085\%Ix+0.1mA$		
				(0.22~2.2)A 5kHz~10kHz	$U=1.0\%Ix+0.2mA$		
				(2.2~11)A 40Hz~1kHz	$U=0.046\%Ix+0.17mA$		
				(2.2~11)A 1kHz~5kHz	$U=0.095\%Ix+0.38mA$		
				(2.2~11)A 5kHz~10kHz	$U=0.36\%Ix+0.75mA$		
				(11~20)A 45Hz~100Hz	$U=0.12\%Ix+5mA$		
				(11~20)A 100Hz~1kHz	$U=0.15\%Ix+5mA$		
		(11~20)A 1kHz~5kHz	$U=0.3\%Ix+5mA$				




No. CNAS L2336

第 68 页 共 187 页


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Resistance	ilac-MRA CNAS CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMANCE ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATION	1 Ω	$U_{rel}=0.011\%$		
				10 Ω	$U_{rel}=0.0027\%$		
				100 Ω	$U_{rel}=0.0012\%$		
				1k Ω	$U_{rel}=0.0010\%$		
				10k Ω	$U_{rel}=0.0010\%$		
				100k Ω	$U_{rel}=0.0013\%$		
				1M Ω	$U_{rel}=0.0023\%$		
				10M Ω	$U_{rel}=0.0046\%$		
				100M Ω	$U_{rel}=0.012\%$		
				52	*Amperemeter, Voltmeter, Wattmeter And Ohmmeter		
DC Current	1mA~1A	$U_{rel}=8 \times 10^{-4}$					
	1A~10A	$U_{rel}=1 \times 10^{-3}$					
	10A~20A	$U_{rel}=2 \times 10^{-3}$					
	AC Voltage	10mV~1000V(45Hz~1kHz)	$U_{rel}=8 \times 10^{-4}$				
AC Current	1mA~10mA(45Hz~1kHz)	$U_{rel}=2 \times 10^{-3}$					
	10mA~1A(45Hz~1kHz)	$U_{rel}=9 \times 10^{-4}$					
	1A~10A(45Hz~100Hz)	$U_{rel}=1 \times 10^{-3}$					
	10A~20A(45Hz~100Hz)	$U_{rel}=2 \times 10^{-3}$					



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Resistance		100 Ω ~ 1M Ω	$U_{rel}=7 \times 10^{-4}$		
		DC Powermeter		(1~3000)W	$U_{rel}=0.09\%$		
		AC Powermeter		(1~3000)W(45Hz~65Hz)	$U_{rel}=0.10\%$		
				1M Ω ~ 10M Ω	$U_{rel}=1 \times 10^{-3}$		
53	*DC Voltage Stable Source	DC Voltage	Calibration specification for DC Stabilized Power Supplies JJF 1597	0.1V~100V	$U_{rel}=0.007\%$		
				100V~1000V	$U_{rel}=0.008\%$		
		DC Current		10 μ A~200A	$U_{rel}=0.07\%$		
				200A~1000A	$U_{rel}=0.09\%$		
		Regulated power supply adjustment rate		0.01%~10%	$U_{rel}=0.008\%$		
		Regulation rate of steady current power supply		0.01%~10%	$U_{rel}=0.09\%$		
		Regulated load adjustment rate		0.01%~10%	$U_{rel}=0.008\%$		
		Steady current load regulation rate		0.01%~10%	$U_{rel}=0.09\%$		
Ripple voltage	0.1mV~1000mV	$U_{rel}=2\%$					



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Ripple current		0.1mA~10A	$U_{rel}=2.4\%$		
54	Resistance strain gauge indicators	Strain Gauge	Resistance strain gauge indicators JJG 623	(10~100000) $\mu \epsilon$	$U_{rel}=0.09\%$		
		Indication Stability		(10~100000) $\mu \epsilon$	$U_{rel}=0.09\%$		
55	*High-voltage Withstanding Voltage Tester	DC Voltage	Verification Regulation of High-voltage Withstanding Voltage Tester JJG(JG) 18	(1~15)kV	$U_{rel}=0.9\%$		
				(15~50)kV	$U_{rel}=1.0\%$		
		AC Voltage		(1~15)kV(45~65)Hz	$U_{rel}=0.9\%$		
				(15~50)kV(45~65)Hz	$U_{rel}=1.0\%$		
		DC Current		(0.1~100) mA	$U_{rel}=1.0\%$		
		AC Current		(0.1~100) mA(45~65)Hz	$U_{rel}=1.2\%$		
Time	(0.1~999.99)s	$U=0.8\%T+0.03s$					
56	*Strength Detector of Insulating Oils	AC Voltage	V. R. of Dielectric Strength Detector of Insulating Oils JJG (Ji)112	(1~50)kV(45Hz~65Hz)	$U_{rel}=1.5\%$		
		Speed of pressure Rise		(0.2~10)kV/s	$U_{rel}=2\%$		
57	*DC High Voltage Test Device	DC Voltage	Calibration Specification for DC High Voltage Test Device JJF(Zhe)1146	(10~50)kV	$U_{rel}=0.9\%$		
				(50~100)kV	$U_{rel}=1.2\%$		
		ripple voltage		(0.1~50)V	$U_{rel}=0.9\%$		
		DC Current		(0.1~100) mA	$U_{rel}=1.0\%$		
		Short Term Stability		(10~100)kV	$U_{rel}=0.9\%$		
58	*AC High Voltage Test Device	AC Voltage	Calibration Specification for AC High Voltage Test Device	(10~50)kV(45Hz~65Hz)	$U_{rel}=0.9\%$		



No. CNAS L2336

第 71 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			JJF(Zhe)1144	(50~100)kV(45Hz~65Hz)	$U_{rel}=1.2\%$		
		Frequency		(45~65)Hz	$U_{rel}=0.02\%$		
		Total Harmonic Distortion		(0.1~100)%	$U_{rel}=0.9\%$		
		Short Term Stability		(10~100)kV	$U_{rel}=0.9\%$		
59	*Magnetic Partical Detector	DC current	Calibration specification for Magnetic Particle Flaw Detectors JJF 1273	(100~2000)A	$U_{rel}=1.2\%$		
		AC current		(200~10000)A, (50Hz~400Hz)	$U_{rel}=1.2\%$		
		Residual magnetic induction intensity		(0.01~0.2)mT	$U=0.05mT$		
		Magnetization time		(0.4~20)s	$U=0.03s$		
60	Fluxgate Magnetometer	magnetism	Calibration Specification for Fluxgate Magnetometer JJF 1519	-250 μ T~250 μ T	$U=1\%B+1 \mu$ T		
61	Magnetometer	magnetism	Calibration Specification for Magnetometers JJF 1656	(0.05~10) mT	$U_{rel}=1.6\%$		
62	Pointer Magnetometer	magnetism	Calibration Specification for Pointer Magnetometer JJF (JG) 75	50 μ T~5mT	$U=1\%B+1 \mu$ T		
63	Tesla meter	magnetism	Calibration Specification for(1 mT~2.5 T) Magnetometers JJF 1832	1mT~2.5T	$U_{rel}=0.3\%$		




No. CNAS L2336

第 72 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
64	Alternating magnetometer	magnetism	Verification regulation of weak magnetic field alternating magnetometer JJG 1049	1pT~1mT	$U_{rel}=1\%$		
65	*Eddy current conductivity meter	Eddy current conductivity	C. S. fHr Eddy current conductivity meter JJF 1692	(0.58~58)MS/m	$U=(0.11\sim0.36) MS/m$		
66	*source meter	DC voltage (output)	Calibration specification for source meter JJF(DZ)0024	(0.5~1)mV	$U_{rel}=0.5\%$		
				(1~10)mV	$U_{rel}=0.2\%$		
				(10~200)mV	$U_{rel}=0.0018\%$		
				(0.2~200)V	$U_{rel}=0.0017\%$		
				(200~1000)V	$U_{rel}=0.0018\%$		
		DC current (output)		(10~20) μ A	$U_{rel}=0.16\%$		
				20 μ A~200mA	$U_{rel}=0.12\%$		
				(0.2~20)A	$U_{rel}=0.07\%$		
		DC voltage (measured)		(10~200)mV	$U_{rel}=0.0091\%$		
				(0.2~200)V	$U_{rel}=0.0022\%$		
				(200~1000)V	$U_{rel}=0.0025\%$		
		DC current (measurement)		(10~200) μ A	$U_{rel}=0.11\%$		
				(0.2~200)mA	$U_{rel}=0.012\%$		
(0.2~2)A	$U_{rel}=0.019\%$						
(2~10)A	$U_{rel}=0.061\%$						




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC resistance (measurement)		(100 Ω ~ 1M Ω)	$U_{rel}=0.007\%$		
				(1M Ω ~ 10M Ω)	$U_{rel}=0.023\%$		
				(10M Ω ~ 100M Ω)	$U_{rel}=0.07\%$		
67	*Magnetic FLUX Meter	Magnetic FLUX	Calibration Specification for Magnetic Flux Meters JJF 1905	0.1mWb~1mWb	$U_{rel}=0.4\%$		
				1mWb~10Wb	$U_{rel}=0.07\%$		
68	*hand - held digital Multimeters	DC voltage	Calibration Specification for hand - held digital Multimeters JJF(DZ) 0023	(10~330)mV	$U=0.005\%V_x+0.78 \mu V$		
				(0.33~3.3)V	$U=0.006\%V_x +1.6 \mu V$		
				(3.3~33)V	$U=0.006\%V_x +16 \mu V$		
				(33~330)V	$U=0.007\%V_x +0.12mV$		
				(330~1000)V	$U=0.007\%V_x +1.2mV$		
		AC voltage		(10~33)mV(20~45)Hz	$U=0.10\%V_x+7 \mu V$		
				(10~33)mV(45Hz~10kHz)	$U=0.018\%V_x +7 \mu V$		
				(33~330)mV(20Hz~45Hz)	$U=0.040\%V_x +7 \mu V$		
				(33~330)mV(45Hz~10kHz)	$U=0.018\%V_x +9 \mu V$		
				(0.33~3.3)V(20Hz~45Hz)	$U=0.04\%V_x +38 \mu V$		
				(0.33~3.3)V(45Hz~10kHz)	$U=0.014\%V_x +46 \mu V$		
		(3.3~33)V(20Hz~45Hz)	$U=0.04\%V_x +0.7mV$				



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC current		(3.3~33)V(45Hz~10kHz)	$U=0.014\%V_x + 0.5mV$		
				(33~330)V(45Hz~1kHz)	$U=0.022\%V_x + 5mV$		
				(33~330)V(1kHz~10kHz)	$U=0.03\%V_x + 12mV$		
				(330~1000)V(40Hz~1kHz)	$U=0.04\%V_x + 12mV$		
				(330~1000)V(1kHz~5kHz)	$U=0.03\%V_x + 12mV$		
				(10~330) μ A	$U=0.033\%I_x + 0.04 μ A$		
				(0.33~3.3)mA	$U=0.020\%I_x + 0.039 μ A$		
				(3.3~33)mA	$U=0.030\%I_x + 1.9 μ A$		
				(33~330)mA	$U=0.030\%I_x + 1.9 μ A$		
				(0.33~1.1)A	$U=0.03\%I_x + 31 μ A$		
				(1.1~3)A	$U=0.03\%I_x + 32 μ A$		
				(3~11)A	$U=0.04\%I_x + 0.26mA$		
		(11~20)A	$U=0.008\%I_x + 0.9mA$				
		AC current		(30~330) μ A(20~45Hz)	$U=0.22\%I_x + 0.2 μ A$		
				(30~330) μ A(45Hz~1kHz)	$U=0.16\%I_x + 0.2 μ A$		
				(0.33~3.3)mA(45Hz~1kHz)	$U=0.15\%I_x + 0.3 μ A$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(0.33~3.3mA)(100Hz~1kHz)	$U=0.12\%I_x + 0.3 \mu A$		
				(3.3~33)mA(45Hz~1kHz)	$U=0.11\%I_x + 3 \mu A$		
				(3.3~33)mA(45Hz~1kHz)	$U=0.06\%I_x + 3 \mu A$		
				(33~330)mA(20~45)Hz	$U=0.11\%I_x + 30 \mu A$		
				(33~330)mA(45Hz~1kHz)	$U=0.06\%I_x + 30 \mu A$		
				(0.33~1.1)A(20~45)Hz	$U=0.2\%I_x + 0.3mA$		
				(0.33~1.1)A(45Hz~1kHz)	$U=0.07\%I_x + 0.3mA$		
				1.1A~3A(20~45)Hz	$U=0.2\%I_x + 0.6mA$		
				1.1A~3A(45Hz~1kHz)	$U=0.07\%I_x + 0.6mA$		
				3A~11A(20~100)Hz	$U=0.07\%I_x + 3mA$		
				3A~11A(100Hz~1kHz)	$U=0.12\%I_x + 3mA$		
				11A~20A(20~45)Hz	$U=0.14\%I_x + 6mA$		
				11A~20A(45Hz~1kHz)	$U=0.17\%I_x + 6mA$		
		Resistance		1 Ω ~ 11 Ω	$U=0.07\%R_x + 0.04 \Omega$		
				11 Ω ~ 33 Ω	$U=0.06\%R_x + 0.06 \Omega$		
				33 Ω ~ 1.1M Ω	$U=0.06\%R_x + 0.08 \Omega$		
				1.1M Ω ~ 3.3M Ω	$U=0.08\%R_x + 0.1 \Omega$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				3.3MΩ ~ 33MΩ	$U=0.3\%R_x + 0.3\Omega$		
				33MΩ ~ 100MΩ	$U=2.4\%R_x + 0.4\Omega$		
		Frequency		10Hz~100kHz	$U=0.0003\%R_x + 6\mu\text{Hz}$		
		Capacitors		1nF ~ 10nF	$U=0.6\%C_x + 0.02\text{nF}$		
				10nF ~ 100nF	$U=0.3\%C_x + 0.1\text{nF}$		
				100nF ~ 1000nF	$U=0.3\%C_x + 0.7\text{nF}$		
				1μF ~ 10μF	$U=0.3\%C_x + 6\text{nF}$		
				10μF ~ 100μF	$U=0.5\%C_x + 60\text{nF}$		
				100μF ~ 1000μF	$U=0.5\%C_x + 0.6\mu\text{F}$		
				1mF ~ 10mF	$U=0.5\%C_x + 6\mu\text{F}$		
				10mF ~ 100mF	$U=1.3\%C_x + 60\mu\text{F}$		
		Temperature		With Thermocouple: (-200~1200)°C	$U=0.5^\circ\text{C}$		
		Temperature		With Platinum thermal resistance:(-200~800)°C	$U=0.3^\circ\text{C}$		
		69		*Medical leakage current testers	DC leakage current		
AC leakage current	(0.001~20)mA(50Hz/60Hz)		$U_{\text{rel}}=0.2\%$				
Test voltage	100mV~600V(40Hz~60Hz)		$U_{\text{rel}}=0.2\%$				
DC input resistance	(0.3~3)kΩ		$U_{\text{rel}}=0.2\%$				



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Input impedance	ilac-M	(0.3~3)k Ω (10Hz~100kHz)	$U_{rel}=0.4\%$		
				(0.3~3)k Ω (100kHz~1MHz)	$U_{rel}=0.7\%$		
		Frequency response characteristics of transmission impedance		(-70~-0.01)dB(10Hz~1kHz)	$U=0.05dB$		
				(-70~-0.01)dB(1kHz~1MHz)	$U=0.1dB$		
70	*Touch Current Testers	AC leakage current	Verification Regulation for Touch Current Testers JIG(Yue) 027	(0.001~20)mA(10Hz~3000Hz)	$U_{rel}=0.2\%$		
				(0.001~20)mA(3kHz~1MHz)	$U_{rel}=0.6\%$		
		DC input resistance		(0.3~3)k Ω	$U_{rel}=0.2\%$		
		Input impedance		(0.3~3)k Ω (10Hz~1kHz)	$U_{rel}=0.4\%$		
				(0.3~3)k Ω (1kHz~1MHz)	$U_{rel}=0.7\%$		
71	*DC High Current Meters	DC Current	Verification Regulation for DC High Current Meters JIG(JG)195	100A~2000A	$U_{rel}=0.007\%$		
72	*The power supply of AC/DC electric welder	DC Voltage	Calibration Specification for the power supply of AC/DC electric welder JJF (Jin) 02	(15~500)V	$U_{rel}=0.08\%$		
		DC Current		(50~2000)A	$U_{rel}=0.12\%$		
		AC Voltage		(15~500)V/50Hz	$U_{rel}=0.08\%$		
		AC Current		(50~2000)A/50Hz	$U_{rel}=0.12\%$		
73	*Arc welding and	DC Voltage	Calibration Specification for	15V~500V	$U_{rel}=0.08\%$		

No. CNAS L2336


第 78 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	resistance welding	DC Current	Arc welding and resistance welding JJF(JX) 1025	50A~2000A	$U_{rel}=0.12\%$		
		AC Voltage		(15~500)V/50Hz	$U_{rel}=0.08\%$		
		AC Current		(50~2000)A/50Hz	$U_{rel}=0.12\%$		
		DC Power		10W~20kW	$U_{rel}=0.3\%$		
		AC Power		(10W~3.9kW)/50Hz	$U_{rel}=0.3\%$		
		Power factor		0.01~1.0	$U=0.02$		
		Frequency		40Hz~200Hz	$U_{rel}=0.3\%$		
		Conversion time		10ms~10s	$U_{rel}=0.3\%$		
		Duration		10s~600s	$U_{rel}=0.03\%$		
		Electrode Force		1N~50kN	$U_{rel}=1\%$		
74	*Uninterruptible power supply	No-load output voltage	Calibration Specification for Uninterruptible Power Supply JJF (DZ) 0027	(10~400)V (50Hz, 60Hz)	$U_{rel}=0.14\%$		
		Frequency		20Hz~1kHz	$U_{rel}=0.03\%$		
		Distortion		0.01%~10%	$U_{rel}=7\%$		
		Rated Output Power		(0.1W~6000W)/50Hz	$U_{rel}=0.3\%$		
		Efficiency		0.1%~100%	$U_{rel}=0.4\%$		
		Load adjustment rate		0.01% ~10%	$U_{rel}=0.2\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Source voltage adjustment rate	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	0.01% ~10%	$U_{rel}=0.2\%$		
		Conversion time		0.1ms ~10ms	$U_{rel}=0.3\%$		
		Duration		1s~1000s	$U=0.02\%T_x+0.07s$		
75	*DC High Current Sources	DC Current	Verification Regulation of DC High Current Sources JJG(JG) 196	(100~2000)A	$U_{rel}=0.007\%$		
76	*DC Low Current Reference Sources	DC Current	Verification Regulation of DC Low Current Reference Sources JJG(JG) 19	5pA~200pA	$U_{rel}=1.2\%$		
				0.2nA~200nA	$U_{rel}=0.24\%$		
				200nA~20 μ A	$U_{rel}=0.12\%$		
				20 μ A~200 μ A	$U_{rel}=0.003\%$		
77	Hall Current (Voltage) Transducer	AC Current	Verification Regulation of Hall Current (Voltage) Transducer JJG(Chuan) 136	1A~20A,(40Hz~400Hz)	$U_{rel}=0.03\%$		
				1A~20A,(400Hz~1000Hz)	$U_{rel}=0.06\%$		
				20A~2000A,(50Hz)	$U_{rel}=0.03\%$		
		DC Current		1A~2000A	$U_{rel}=0.007\%$		
		AC Voltage		5V~300V,(45Hz~10kHz)	$U_{rel}=0.03\%$		
				300V~1000V,(45Hz~10kHz)	$U_{rel}=0.04\%$		
		DC Voltage		5V~1000V	$U_{rel}=0.003\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
78	*Standard AC power source	AC Voltage	Verification regulation for standard AC power source JJG(JG) 6	10mV~200mV,(40Hz~20kHz)	$U=0.004\%V_x+0.002mV$		
				0.2V~1000V,(40Hz~20kHz)	$U_{rel}=0.003\%$		
		AC Current		1mA~20A,(40Hz~10kHz)	$U_{rel}=0.007\%$		
				20A~80A,(40Hz~1kHz)	$U_{rel}=0.03\%$		
		AC power		(1~1000)V,(0.001~80)A,(45Hz~60Hz)	$U_{rel}=0.03\%$		
		Phase		-150° ~ 150°	$U=0.009^\circ$		
		Frequency		40Hz~10kHz	$U_{rel}=0.002\%$		
		Voltage Distortion		0.1%~30%,(40Hz~1kHz)	$U_{rel}=7\%$		
Current distortion	0.1%~30%,(40Hz~1kHz)	$U_{rel}=7\%$					
79	*Current Transformer Volt-Ampere Characteristic Meters	AC Voltage	Calibration Specification for Current Transformer Volt-Ampere Characteristic JJF 1584	(1~200)V, (50Hz)	$U_{rel}=0.08\%$		
		AC Current		(0.1~5)A, (50 Hz)	$U_{rel}=0.12\%$		
80	*Mutual Inductor Error and Volt-ampere characteristic tester	voltage	Calibration Specification of Mutual Inductor Error and Volt-ampere characteristic tester JJF(Min) 1108	(1~200)V, (50Hz)	$U_{rel}=0.08\%$		
		current		(0.1~5)A, (50 Hz)	$U_{rel}=0.12\%$		
		Current Transformer Ratio		(5A~2000A)/(5A)	$U_{rel}=0.06\%$		




No. CNAS L2336

第 81 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Voltage Transformer Ratio	ilac-M	(0.1~10)kV, (0.1~35)kV, (0.1~50)kV/(0.1、0.1/√3)kV	$U_{rel}=0.06\%$		
		Ratio difference		0.01%~10%	$U=0.06\%$		
		phase difference		(0.1~900)	$U=1.4$		
		Secondary circuit load		(0.01~10000) Ω	$U_{rel}=0.03\%$		
81	*High Current Generator	AC Current	Calibration Specification of High Current Generator JJF(JX) 1037	(20~2000)A 50Hz	$U_{rel}=0.03\%$		
82	*Instrument Voltage Transformers	Ratio difference	Verification Regulation of Instrument Voltage Transformers JJG 314	(0.1kV~35kV)/100V, (0~35kV)/(100/√3) V20%Rated voltage	$U_{rel}=0.06\%$		
				(0.1kV~35kV)/100V, (0~35kV)/(100/√3) V50%Rated voltage	$U_{rel}=0.05\%$		
				(0.1kV~35kV)/100V, (0~35kV)/(100/√3) V80%~120%Rated voltage	$U_{rel}=0.03\%$		
				(0.1kV~10kV)/100V, (0~35kV)/(100/√3) V20%Rated voltage	$U_{rel}=0.06\%$		
				(0.1kV~10kV)/100V, (0~35kV)/(100/√3) V50%Rated voltage	$U_{rel}=0.05\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Phase difference	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITED CERTIFICATION	(0.1kV~10kV) /100V, (0~35kV)/(100/ √3) V80%~ 120%Rated voltage	$U_{rel}=0.03\%$		
				(0.001~900)' 20% Rated voltage	$U=1.4'$		
				(0.001~900)' 50% Rated voltage	$U=1.2'$		
				(0.001~900)' 80%~ 120%Rated voltage	$U=0.9'$		
83	*Resistive Current Testers for Zinc-oxide Surge Arrester	AC Voltage	Calibration Specification for Zinc-oxide Arrester Testers Surge Arrester JJF2194	(1~300)V, (45Hz~ 65Hz)	$U_{rel}=0.1\%$		
		AC current		(0.1~30)mA, (45Hz~ 65Hz)	$U_{rel}=0.2\%$		
		Phase		0° ~270°	$U=0.1^\circ$		
84	*Power Analyzer	DC Voltage	Calibration Specification for Power Analyzers JJF 2040	10mV~220mV	$U=9 \times 10^{-6}U_x+0.8 \mu V$		
				0.22V~2.2V	$U=8 \times 10^{-6}U_x+1.2 \mu V$		
				2.2V~22V	$U=8 \times 10^{-6}U_x+8 \mu V$		
				22V~220V	$U=9 \times 10^{-6}U_x+100 \mu V$		
				220V~1000V	$U=11 \times 10^{-6}U_x+0.6mV$		
		AC voltage		10mV~220mV,(10Hz~ 20Hz)	$U=0.06\%U_x+16 \mu V$		
				10mV~220mV(20Hz~ 40Hz)	$U=0.024\%U_x+10 \mu V$		
				10mV~220mV(40Hz~ 20kHz)	$U=0.011\%U_x+10 \mu V$		




No. CNAS L2336

第 83 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
				0.22V~2.2V,(10Hz~20Hz)	$U=0.06\%U_x+100\ \mu\text{V}$			
				220mV~2.2V,(20Hz~40Hz)	$U=0.018\%U_x+30\ \mu\text{V}$			
				220mV~2.2V,(40Hz~20kHz)	$U=0.0085\%U_x+7\ \mu\text{V}$			
				2.2V~22V,(10Hz~20Hz)	$U=0.06\%U_x+0.1\text{mV}$			
				2.2V~22V,(20Hz~40Hz)	$U=0.018\%U_x+0.3\text{mV}$			
				2.2V~22V,(40Hz~20kHz)	$U=0.0085\%U_x+70\ \mu\text{V}$			
				22V~220V,(10Hz~20Hz)	$U=0.06\%U_x+10\text{mV}$			
				22V~220V,(20Hz~40Hz)	$U=0.018\%U_x+3\text{mV}$			
				22V~220V,(40Hz~20kHz)	$U=0.009\%U_x+1\text{mV}$			
				220V~1000V,(15Hz~50Hz)	$U=0.05\%U_x+20\text{mV}$			
				220V~1000V,(50Hz~1kHz)	$U=0.009\%U_x+4\text{mV}$			
				DC Current				
					10 μA ~220 μA	$U=0.005\%I_x+7\text{nA}$		
			1mA~2.2mA	$U=0.006\%I_x+10\text{nA}$				
			2.2mA~22mA	$U=0.006\%I_x+100\text{nA}$				
			22mA~220mA	$U=0.007\%I_x+1\ \mu\text{A}$				
			0.22A~2.2A	$U=0.009\%I_x+30\ \mu\text{A}$				



在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC current		2.2A~10A	$U=0.02\%I_x+3mA$		
				10A~20A	$U=0.03\%I_x+6mA$		
				0.1mA~2.2mA,(10Hz~20Hz)	$U=0.08\%I_x+30nA$		
				0.1mA~2.2mA,(20Hz~40Hz)	$U=0.042\%I_x+25nA$		
				0.1mA~2.2mA,(40Hz~1kHz)	$U=0.016\%I_x+20nA$		
				0.1mA~2.2mA,(1kHz~5kHz)	$U=0.07\%I_x+0.05 \mu A$		
				0.1mA~2.2mA,(5kHz~10kHz)	$U=0.18\%I_x+0.1 \mu A$		
				2.2mA~22mA,(10Hz~20Hz)	$U=0.08\%I_x+0.5 \mu A$		
				2.2mA~22mA,(20Hz~40Hz)	$U=0.042\%I_x+0.4 \mu A$		
				2.2mA~22mA,(40Hz~1kHz)	$U=0.016\%I_x+0.4 \mu A$		
				2.2mA~22mA,(1kHz~5kHz)	$U=0.07\%I_x+5 \mu A$		
				2.2mA~22mA,(5kHz~10kHz)	$U=0.18\%I_x+10 \mu A$		
				22mA~220mA,(10Hz~20Hz)	$U=0.08\%I_x+5 \mu A$		
				22mA~220mA,(20Hz~40Hz)	$U=0.042\%I_x+4 \mu A$		
				22mA~220mA,(40Hz~1kHz)	$U=0.018\%I_x+4 \mu A$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				22mA~220mA,(1kHz~5kHz)	$U=0.07\%I_x+50\mu A$		
				22mA~220mA,(5kHz~10kHz)	$U=0.18\%I_x+100\mu A$		
				0.22A~2A,(16Hz~850Hz)	$U=0.02\%I_x+48\mu A$		
				0.22A~2A,(850Hz~6kHz)	$U=0.06\%I_x+60\mu A$		
				0.22A~2.2A,(6kHz~10kHz)	$U=0.1\%I_x+0.2mA$		
				2A~10A,(16Hz~450Hz)	$U=0.02\%I_x+0.3mA$		
				2A~10A,(450Hz~850Hz)	$U=0.03\%I_x+0.3mA$		
				2A~10A,(850Hz~6kHz)	$U=0.06\%I_x+0.4mA$		
				10A~20A,(16Hz~450Hz)	$U=0.024\%I_x+0.8mA$		
				10A~20A,(450Hz~850Hz)	$U=0.03\%I_x+0.9mA$		
				10A~20A,(850Hz~6kHz)	$U=0.07\%I_x+1mA$		
				20A~80A,(40Hz~450Hz)	$U=0.03\%I_x+3mA$		
				20A~80A,(450Hz~850Hz)	$U=0.04\%I_x+3mA$		
				20A~80A,(850Hz~3kHz)	$U=0.07\%I_x+3mA$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC power		33mV~ 1000V,0.33mA~ 2.9999A	$U_{rel}=0.03\%$		
				33mV~1000V,3A~20A	$U_{rel}=0.08\%$		
		AC power		0.1A~20A,6.4V~ 1000V,(16Hz~ 69Hz,0.75≤PF≤1)	$U_{rel}=0.03\%$		
				0.1A~20A,6.4V~ 1000V,(16Hz~69Hz,0.5 ≤PF≤0.75)	$U_{rel}=0.032\%$		
				0.1A~20A,6.4V~ 1000V,(16Hz~ 69Hz,0.25≤PF≤0.5)	$U_{rel}=0.04\%$		
				0.1A~20A,6.4V~ 1000V,(69Hz~ 180Hz,0.75≤PF≤1)	$U_{rel}=0.032\%$		
				0.1A~20A,6.4V~ 1000V,(69Hz~ 180Hz,0.5≤PF≤0.75)	$U_{rel}=0.04\%$		
				0.1A~20A,6.4V~ 1000V,(69Hz~ 180Hz,0.25≤PF≤0.5)	$U_{rel}=0.06\%$		
				0.1A~20A,6.4V~ 1000V,(180Hz~ 450Hz,0.75≤PF≤1)	$U_{rel}=0.05\%$		
				0.1A~20A,6.4V~ 1000V,(180Hz~ 450Hz,0.5≤PF≤0.75)	$U_{rel}=0.07\%$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.1A~20A,6.4V~1000V,180Hz~450Hz,0.25≤PF≤0.5)	$U_{rel}=0.14\%$		
				20A~80A,6.4V~1000V,(16Hz~69Hz,0.75≤PF≤1)	$U_{rel}=0.041\%$		
				20A~80A,6.4V~1000V,(16Hz~69Hz,0.5≤PF≤0.75)	$U_{rel}=0.043\%$		
				20A~80A,6.4V~1000V,(16Hz~69Hz,0.25≤PF≤0.5)	$U_{rel}=0.05\%$		
				20A~80A,6.4V~1000V,(69Hz~180Hz,0.75≤PF≤1)	$U_{rel}=0.042\%$		
				20A~80A,6.4V~1000V,(69Hz~180Hz,0.5≤PF≤0.75)	$U_{rel}=0.05\%$		
				20A~80A,6.4V~1000V,(69Hz~180Hz,0.25≤PF≤0.5)	$U_{rel}=0.07\%$		
				20A~80A,6.4V~1000V,(180Hz~450Hz,0.75≤PF≤1)	$U_{rel}=0.06\%$		
				20A~80A,6.4V~1000V,(180Hz~450Hz,0.5≤PF≤0.75)	$U_{rel}=0.09\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-M	20A~80A,6.4V~1000V,180Hz~450Hz,0.25≤PF≤0.5)	$U_{rel}=0.18\%$		
		Frequency		10Hz~2MHz	$U=0.0003\%fx+5 \mu Hz$		
		Phase		(1~360)°, (16~850)Hz	$U=0.007\%$		
85	*DC ground fault detectors	DC Voltage	Calibration Specification for DC ground fault detectors JJF (DZ) 0094	24V~300V	$U_{rel}=0.06\%$		
		AC Voltage		(10~300)V,(45Hz~55Hz)	$U_{rel}=0.1\%$		
		Resistance		(1~999)k Ω	$U_{rel}=0.2\%$		
		Capacitor		(0.1~100) μ F,(45Hz~55Hz)	$U_{rel}=0.3\%$		
86	*Clamp Meter Calibrators	DC Current	Calibration Specification for Clamp Meter Calibrators JJF (Min) 1031	0.01A~20A	$U_{rel}=0.003\%$		
				20A~2000A	$U_{rel}=0.007\%$		
		AC Current		0.01A~20A(40Hz~1kHz)	$U_{rel}=0.005\%$		
				20A~1000A(40Hz~400Hz)	$U_{rel}=0.03\%$		
		Distortion		0.03%~10%	$U_{rel}=12\%$		
		Ripple Coefficient		0.01%~10%	$U_{rel}=2\%$		
87	*Frequency series resonance device	AC Voltage	Calibration Specification for Frequency series resonance device JJF(Gan) 0081	(10~100)kV, 50Hz	$U_{rel}=1.2\%$		
		Frequency		(30~300) Hz	$U_{rel}=0.02\%$		
		Distortion Degree		0.1%~10%, (30Hz~300Hz)	$U_{rel}=12\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
88	*Transformer Core Grounding Current Tester	AC current	Calibration Specification for Transformer Core Grounding Current Tester JJF (Zhe) 1143	1mA~10A,(45Hz~65Hz)	$U_{rel}=0.5\%$		
89	*Clamp Digital Phase voltammeters	AC Voltage	Calibration Specification for Clamp Digital Phase voltammeters JJF (Ji) 210	(0.1~600)V,(45Hz~65Hz)	$U_{rel}=0.1\%$		
		AC current		(0.01~20)A,(45Hz~65Hz)	$U_{rel}=0.1\%$		
				(20~600)A,(45Hz~65Hz)	$U_{rel}=0.5\%$		
		AC power		1mW~12kW,(45Hz~65Hz)	$U_{rel}=0.2\%$		
	Phase		0° ~360° ,(45Hz~65Hz)	$U=0.2^\circ$			
90	*AC/DC High Current Measuring Instruments	DC Current	Calibration Specification of AC/DC High Current Measuring Instruments JJF(HuSuZheWan) 4007	100A~2000A	$U_{rel}=0.007\%$		
		AC Current		100A~2000A, (50Hz)	$U_{rel}=0.03\%$		
91	*Power Frequency A.C. Electrical Quantities Measuring Transducers	AC voltage	Power Frequency A.C. Electrical Quantities Measuring Transducers JJG 126	10mV~1000V,(45Hz~65Hz)	$U_{rel}=0.02\%$		
		AC current		1mA~80A,(45Hz~65Hz)	$U_{rel}=0.02\%$		
		AC power		1V~600V,1mA~80A,(45Hz~65Hz)	$U_{rel}=0.03\%$		
		Power factor		0.5C~1~0.5L	$U_{rel}=0.03\%$		
		Frequency		45Hz~65Hz	$U_{rel}=0.02\%$		
		Ripple content		0.01%~1%	$U_{rel}=0.05\%$		
92	*Clamp Ammeters for Measurement	DC current	Verification Regulation of Clamp Ammeters	1mA~60A	$U_{rel}=0.3\%$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	of Leakage Currents	AC current	for Measurement of Leakage Currents JJG(ji) 3007	1mA~60A(40Hz~400Hz)	$U_{rel}=0.3\%$		
93	AC shunts	AC Resistance	Calibration specification for AC shunts JJF (DZ) 0020	0.01 Ω ~ 10 Ω ,(0.02A~20A, 50Hz)	$U_{rel}=0.015\%$		
				0.01m Ω ~ 0.01 Ω ,(20A~2000A, 50Hz)	$U_{rel}=0.03\%$		
94	AC & DC Resistance Simulators	DC Resistance	Calibration Specification for AC & DC Resistance Simulators JJF 1723	1 $\mu \Omega$ ~ 10 $\mu \Omega$	$U_{rel}=0.2\%$		
				10 $\mu \Omega$ ~ 100 $\mu \Omega$	$U_{rel}=0.03\%$		
				0.1m Ω ~ 1 Ω	$U_{rel}=0.012\%$		
				1 Ω ~ 200 Ω	$U_{rel}=0.01\%$		
		AC Resistance		0.1m Ω ~ 1m Ω ,(45Hz~65Hz)	$U_{rel}=0.05\%$		
				1m Ω ~ 10m Ω ,(45Hz~65Hz)	$U_{rel}=0.03\%$		
				10m Ω ~ 200 Ω ,(45Hz~65Hz)	$U_{rel}=0.02\%$		
		DC current		0.1A ~ 60A	$U_{rel}=0.012\%$		
AC current	0.1A ~ 80A,(45Hz~65Hz)	$U_{rel}=0.03\%$					
95	DC Low-resistance Meters	DC Resistance	Calibration Specification for Calibrators of DC Low-resistance Meters (JJF (CHUAN) 173-2020) JJF (CHUAN) 173	5 $\mu \Omega$ ~ 10 $\mu \Omega$	$U_{rel}=0.1\%$		
				10 $\mu \Omega$ ~ 100 $\mu \Omega$	$U_{rel}=0.03\%$		
				100 $\mu \Omega$ ~ 1m Ω	$U_{rel}=0.02\%$		
				1m Ω ~ 100 Ω	$U_{rel}=0.003\%$		
				100 Ω ~ 20k Ω	$U_{rel}=0.0024\%$		



No. CNAS L2336

第 91 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
96	Transformer Capacity Tester	AC Voltage	Calibration Specification for Transformer Capacity Tester JJF (shan) 087	40V~660V(45Hz~65Hz)	$U_{rel}=0.02\%$		
		AC Current		0.05A~80A(45Hz~65Hz)	$U_{rel}=0.02\%$		
		AC Power		0.1kW~140kW, (45Hz~65Hz)	$U_{rel}=0.03\%$		
		Frequency		45Hz~65Hz	$U_{rel}=0.006\%$		
		Phase		0° ~360° (45Hz~65Hz)	$U=0.005^\circ$		
97	*Cable Fault Flashover Tester	Test Distance	Calibration Specification for Cable Fault Flashover Tester JJF(zhe)1164	5m~100km	$U_{rel}=0.1\%$		
98	*Welding Power Sources of DC Electric Welding Machines	DC Current	Calibration Specification for Welding Power Sources of DC Electric Welding Machines JJF 1985	1A~2000A	$U_{rel}=0.12\%$		
		DC Voltage		1V~150V	$U_{rel}=0.08\%$		
99	*Equipotential Testers	Resistance	Calibration Specification for Equipotential Testers JJF(JI)3022	0.001 Ω ~ 0.1 Ω	$U_{rel}=0.3\%$		
				0.1 Ω ~ 1 Ω	$U_{rel}=0.2\%$		
				1 Ω ~ 10 Ω	$U_{rel}=0.12\%$		
				10 Ω ~ 3000 Ω	$U_{rel}=0.3\%$		
100	*Voltage Monitor	AC Voltage	Calibration Specification of Voltage Monitor JJF (zhe) 1098	(5~456)V,(45Hz~65Hz)	$U_{rel}=0.08\%$		
		Time		(0.5~120)min	$U=0.4s$		
101	*DC Digital Power Meters	DC Power	Calibration Specification for Dc Power Meters JJF (e) 163	1W~330W	$U_{rel}=0.029\%$		
				330W~3300W	$U_{rel}=0.027\%$		




No. CNAS L2336

第 92 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				3.3kW~20kW	$U_{rel}=0.083\%$		
102	*DC High Voltage Generator	DC Voltage	Calibration Specification of DC High Voltage Generator JJF (JX) 1036	(0.1~50)kV	$U_{rel}=0.07\%$		
		DC Current		(0.1~50)mA	$U_{rel}=0.07\%$		
103	*Standard DC Power Sources	DC Voltage	Verification Regulation for Standard DC Power Sources JJG (Yue) 064	(1~200)mV	$U=0.0005\%V_x+0.1\mu V$		
				(0.2~2)V	$U=0.0004\%V_x+0.4\mu V$		
				(2~20)V	$U=0.0004\%V_x+4\mu V$		
				(20~200)V	$U=0.0005\%V_x+0.04mV$		
				(200~1000)V	$U=0.0005\%V_x+0.5mV$		
				DC Current	(10~200) μA	$U=0.0014\%I_x+0.4nA$	
		(0.2~2)mA			$U=0.0014\%I_x+4nA$		
		(2~20)mA			$U=0.0015\%I_x+40nA$		
		(20~200)mA			$U=0.0037\%I_x+0.8\mu A$		
		(0.2~20)A			$U_{rel}=0.008\%$		
		(20~100)A			$U_{rel}=0.015\%$		
		DC Power		1W~1kW	$U_{rel}=0.006\%$		
1kW~10kW	$U_{rel}=0.02\%$						
104	*Current Coils	AC Current-Voltage Conversion Value	Calibration Specification of Current Coils JJF(Wan) 170	1A~1000A/1mV~5V(50Hz)	$U_{rel}=0.1\%$		




No. CNAS L2336

第 93 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Current-Voltage Conversion Value		1A~1000A/1mV~5V	$U_{rel}=0.1\%$		
105	*Measuring Transducers for Covering DC Electrical Quantities	DC voltage DC current	Verification Regulation of Measuring Transducers for Covering DC Electrical Quantities JJG (JG) 191	10mV~1000V 10 μ A ~ 2000A	$U_{rel}=0.01\%$ $U_{rel}=0.01\%$		
Radio measuring instrument							
1	*Passive Intermodulation Analyzers	Carrier Output Frequency Carrier output level The Receiver Displays the Average Noise Level Passive Intermodulation System residual intermodulation	Calibration Specification for Passive Intermodulation Analyzers JJF 1463	800MHz~3GHz 20dBm~46dBm(800MHz~3GHz) -120dBm~-60dBm(800MHz~3GHz) -170dBc~-90dBc(800MHz~3GHz) -170dBc~-90dBc	$U_{rel}=1.2 \times 10^{-7}$ $U=0.5\text{dB}$ $U=0.7\text{dB}$ $U=0.5\text{dB}$ $U=0.7\text{dB}$		
2	*RF and Microwave Power Meters	power	Calibration Specification for RF and Microwave Power Meters JJF 1885	(0.01~25) mW, (0.01~2)GHz (0.01~25) mW, (2.1~12.4) GHz	$U_{rel}=1.9\%$ $U_{rel}=2.6\%$	JJG 1024: except Instrument Accuracy;JJ	




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(0.01~25) mW, (12.75 ~18)GHz	$U_{rel}=3.3\%$	G (ELECTR ONIC) 30202:Instru ment Accuracy	
		VSWR		1.0~2.0, (10MHz~ 18GHz)	$U=0.07$		
		Instrument Accuracy		3.16 μ W~100mW	$U_{rel}=0.5\%$		
		Power Reference		1mW(50MHz)	$U_{rel}=1.9\%$		
		Source Frequency Accuracy		50MHz	$U_{rel}=1 \times 10^{-4}$		
3	*Function Generator	Frequency	V.R. of Function Generator JJG 840	1Hz~225MHz	$U_{rel}=4.6 \times 10^{-7}$		
Voltage	10mV~20V(p-p), (1Hz~225MHz)	$U_{rel}=1.2\%$					
DC Offset	0.01V~5V	$U_{rel}=0.1\%$					
Attenuation	(0~80)dB, 1kHz	$U=0.10$ dB					
Distortion	(0.03~10)%, (20Hz~ 20kHz)	$U_{rel}=12\%$					
Sine wave amplitude flatness	100mV~10V (1kHz~ 225MHz)	$U=0.10$ dB					
Pulse duty cycle	1%~99%	$U_{rel}=0.14\%$					
Overshoot	0.001%~20%	$U_{rel}=1.8\%$					
rise time	1ns~10 μ s	$U_{rel}=5.5\%$					
4	*Low-frequency Volt-meters	Voltage	Calibration Specification for Low-frequency Volt-meters JJF 1925	10mV ~300V, (10Hz~500kHz)	$U_{rel}=0.6\% \sim 1.0\%$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Frequency Response		10Hz ~ 500kHz	$U_{rel}=0.9\%$		
5	*RF Power sensor	power	Calibration Specification for RF & Microwave Power Sensors JJF 1887	(0.01~25) mW, (0.01~2)GHz	$U_{rel}=1.9\%$		
				(0.01~25) mW, (2.1~12.4) GHz	$U_{rel}=2.6\%$		
				(0.01~25) mW, (12.75 ~18)GHz	$U_{rel}=3.3\%$		
		VSWR		1.0~2.0, (10MHz~18GHz)	$U=0.07$		
6	*Attenuators from Radio Frequency to Microwave Frequency	Attenuation	C.S. for Attenuators from Radio Frequency to Microwave Frequency JJF 2092	0dB~110dB, (10MHz~18GHz)	$U=0.3dB$		
		VSWR		1.0~2.0, (10MHz~18GHz)	$U=0.07$		
7	*Spectrum Analyzer	Frequency	C.S. for Spectrum Analyzers JJF 1396	9kHz~18GHz	$U_{rel}=1.2 \times 10^{-7} \sim 3.2 \times 10^{-5}$		
		Electrical Level		(-120~30)dBm	$U=0.3dB$		
		Frequency Response		(-120~30)dBm, 9kHz~18GHz	$U=0.38dB$		
		Sweep frequency		100Hz~18GHz	$U_{rel}=0.02\%$		
		Resolution bandwidth		1Hz~30MHz	$U_{rel}=0.9\%$		
		influence of resolution bandwidth conversion on amplitude measurement		0dB~2dB(1Hz~30MHz)	$U=0.02dB$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Reference level	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(-90~10)dBm	U=0.22dB		
		Vertical display scale		1dB/div~10dB/div (9kHz~18GHz)	U=0.22dB		
		Display average noise		(-160~0)dBm	U=1.4dB		
		Calibration signal level		(-40~0)dBm	U=0.2dB		
		Calibration signal frequency		150kHz~1000MHz	U _{rel} =1.2×10 ⁻⁷		
8	*Vector Network Analyzers	Frequency	Calibration Specification for Vector Network Analyzers JJF 1495	9kHz~40GHz	U _{rel} =1.2×10 ⁻⁷		
		Electrical Level		(-55 ~ 10)dBm , (9kHz~18GHz)	U=0.20dB		
				(-25 ~ 10)dBm, (18GHz~26.5GHz)	U=0.22dB		
				(-25 ~ 10)dBm, (26.5GHz~40GHz)	U=0.26dB		
		Transmission Amplitude		20dB, 50dB, (50MHz~18GHz)	U=0.54dB		
		VSWR		1.0~2.0, (10MHz~2GHz)	U _{rel} =4%		
				1.0~2.0, (2GHz~12GHz)	U _{rel} =4.2%		
				1.0~2.0, (12GHz~18GHz)	U _{rel} =5.2%		
Modular dynamic accuracy	(0~-100)dB,1MHz~18GHz	U=(0.06~0.40)dB					




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Sweep noise (modulus)		0~1(9KHz~18GHz)	U=0.02dB		
		Sweep noise (phase)		(0~1)° (9kHz~18GHz)	U=0.02°		
		Background noise		-140dBm~-60dBm(9kHz~18GHz)	U=2.3dB		
		Crosstalk		-140dBm~-60dBm(9kHz~18GHz)	U=2.0dB		
		Calibration piece reflection coefficient (modulus)		0~1,300kHz~18GHz	U=0.004~0.014		
		Calibration piece reflection coefficient (phase)		0~1,300kHz~18GHz	U=0.020~0.044		
		Scattering parameters(S21/S12 modulus)		(-180~180)°,300kHz~18GHz	U=(1.1~2.5)°		
		Scattering parameters(S21/S12 phase)		(-20~-50)dB,10MHz~18GHz	U=0.26dB		
		Scattering parameters(modulus)		(-180~180)°,10MHz~18GHz	U=1.5°		
		Scattering parameters(phase)		0~1,10MHz~18GHz	U=0.03		
				(-180~180)°,10MHz~18GHz	U=1.5°		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
9	*543-1 noise signal generator	Sound pressure level	543-1 noise signal generator calibration wi SBJZ-ZB-D001	(75~118)dB	$U=2.4\text{dB}$		
		AC Voltage		(0.05V~1.1V), 14kHz	$U=0.01\text{V}$		
10	*Pulse Generator	Output Level	Verification Regulation of Pulse Generator JJG 490	10mV~200V, (1kHz)	$U_{\text{rel}}=0.15\%$		
		Rise/Fall time		(1~100) ns	$U_{\text{rel}}=4.2\%$		
		Pulse Width		(0.5~100) ns	$U_{\text{rel}}=4.2\%$		
		Output Frequency		5Hz~300MHz	$U_{\text{rel}}=4.2 \times 10^{-6}$		
		DC Offset		0.1V~20V	$U_{\text{rel}}=0.12\%$		
11	*Low frequency Signal Generator	Frequency	V.R. of Low Frequency Signal Generator JJG 602	1Hz~1MHz	$U_{\text{rel}}=6 \times 10^{-6}$		
		Voltage		10mV~100V, (1Hz~1MHz)	$U_{\text{rel}}=0.1\% \sim 1.2\%$		
		Attenuation		(0~60)dB, (10Hz~300kHz)	$U=0.10\text{dB} \sim 0.12\text{dB}$		
		Distortion		(0.001~10)%, (20Hz~20kHz)	$U_{\text{rel}}=12\%$		
12	*High frequency Signal Generator	Frequency	C.S. for Signal Generators JJF 1931	150kHz~18GHz	$U_{\text{rel}}=6 \times 10^{-9} \sim 1.8 \times 10^{-6}$		
		Electrical Level		(30~-110)dBm(2.5MHz~18GHz)	$U=0.38\text{dB}$		
		AM		5%~99%, (150kHz~18GHz), (0.01~100) Hz	$U_{\text{rel}}=2.6\% \sim 1.5\%$		
		FM		(0.4 ~ 400)kHz, (150kHz~18GHz), (0.01~200) kHz	$U_{\text{rel}}=2.7\% \sim 1.2\%$		




No. CNAS L2336

第 99 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		PM		(0.4 ~ 400)rad, (150kHz~18GHz)	$U_{rel}=3.5\% \sim 5.0\%$		
13	*Distortion Meter	Distortion	Calibration Specification for Distortion Meters JJF 1852	0.03%~100%(10Hz~ 10kHz)	$U_{rel}=2.5\% \sim 1.2\%$		
		Voltage		10mV ~ 300V(10Hz~ 500kHz)	$U_{rel}=0.6\% \sim 1.0\%$		
		Frequency Response		10Hz ~ 500kHz	$U_{rel}=0.9\%$		
14	*Audio Analyzer	Output Frequency	Calibration Specification for Audio Analyzer JJF 1395	10Hz~200kHz	$U_{rel}=2.0 \times 10^{-6}$		
		Output Level		10mV~20V (10Hz~ 20kHz)	$U_{rel}=0.10\%$		
				10mV~20V (20kHz~ 50kHz)	$U_{rel}=0.11\%$		
				10mV~20V (50kHz~ 100kHz)	$U_{rel}=0.14\%$		
				10mV~20V (100kHz~ 200kHz)	$U_{rel}=0.36\%$		
				Total Harmonic Distortion	(0.03%~0.3%) (20Hz~20kHz)	$U_{rel}=12\%$	
		Frequency Measurement level Measurement		10Hz~200kHz	$U_{rel}=6 \times 10^{-4}$		
		10mV~300V (10Hz~ 45Hz)		$U_{rel}=0.22\%$			
		10mV~300V (45Hz~ 10kHz)		$U_{rel}=0.10\%$			
		10mV~300V (10kHz~ 20kHz)		$U_{rel}=0.14\%$			




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Distortion Measurement	ilac-M	10mV~300V (20kHz~50kHz)	$U_{rel}=0.22\%$		
				10mV~300V (50kHz~100kHz)	$U_{rel}=0.5\%$		
				10mV~300V (100kHz~200kHz)	$U_{rel}=0.9\%$		
				0.03%~0.05% (10Hz~200kHz)	$U_{rel}=2.3\%$		
				0.05%~0.1% (10Hz~200kHz)	$U_{rel}=1.6\%$		
				0.1%~0.3% (10Hz~200kHz)	$U_{rel}=0.8\%$		
				0.3%~100% (10Hz~200kHz)	$U_{rel}=0.6\%$		
15	*Oscilloscopes Recorder	Input Impedance Voltage	Calibration standards on Oscilloscopes Recorder JJF(Min)1061	50 Ω ,1M Ω	$U_{rel}=0.12\%$		
			6mV~0.3V(1M Ω)	$U_{rel}=1.0\%$			
			0.6V~120V(1M Ω)	$U_{rel}=0.37\%$			
			6mV~ 6V(50 Ω)	$U_{rel}=1.1\%$			
	Time		2ns ~ 200ms	$U_{rel}=0.16\%$			
	RiseTime		400ms ~ 5s	$U_{rel}=0.5\%$			
	Bandwidth		500ps ~ 50ns	$U_{rel}=4.2\%$			
			10MHz ~ 100MHz	$U_{rel}=3.2\%$			
			100MHz ~300MHz	$U_{rel}=3.6\%$			




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				300MHz ~ 600M	$U_{rel}=5.4\%$		
16	*Digital Oscilloscope	Voltage	Verification regulation for digital oscilloscope GJB 7691,C.S. of Digital Storage Oscilloscope JJF 1057	6mV~0.3V(1MΩ)	$U_{rel}=0.97\%~0.38\%$		
				0.6V~120V(1MΩ)	$U_{rel}=0.37\%$		
				6mV~6V(50Ω)	$U_{rel}=1.1\%~0.45\%$		
		Time		2ns ~ 200ms	$U_{rel}=0.16\%$		
				400ms ~ 5s	$U_{rel}=0.17\%~0.49\%$		
		RiseTime		30ps ~ 500ns	$U_{rel}=4.2\%$		
		Bandwidth		10MHz ~ 100MHz	$U_{rel}=3.2\%$		
				100MHz ~ 300MHz	$U_{rel}=3.6\%$		
				300MHz ~ 600M	$U_{rel}=5.4\%$		
				600MHz ~ 2GHz	$U_{rel}=5.0\%$		
				2GHz ~ 6GHz	$U_{rel}=5.3\%$		
		Input Impedance		50Ω, 1MΩ	$U_{rel}=0.12\%$		
		Calibration signal frequency		5Hz ~ 300kHz	$U_{rel}=1.2 \times 10^{-7}$		
Calibration signal amplitude	10mV~10V	$U_{rel}=0.5\%$					
17	*Analogue Oscilloscope	Voltage	V.R. of Analogue Oscilloscope JJG 262	12mV~ 0.3V(1MΩ)	$U_{rel}=0.56\%~0.28\%$		
				0.6V~30V(1MΩ)	$U_{rel}=0.27\%$		




在线扫码获取验证

No. CNAS L2336

第 102 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time	ilac-MRA CHINA NATIONAL ACCREDITATION FOR CALIBRATION COFFICE SCHEDULE OF ACCREDITATION CERTIFICATE	20ns ~ 500ms	$U_{rel}=0.25\%$		
		RiseTime		500ps ~ 50ns	$U_{rel}=4.2\%$		
		Bandwidth		10MHz ~ 100MHz	$U_{rel}=3.2\%$		
				100MHz ~ 300MHz	$U_{rel}=3.6\%$		
				300MHz ~ 600M	$U_{rel}=5.4\%$		
		Input Impedance		50 Ω ,1M Ω	$U_{rel}=0.12\%$		
18	*Television video Signal Generator	level	C.S. for Television video Signal Generator JJF 1235	Luminance :(1~1000)mV; Chroma:(1~1000)mV	$U=4.1mV$		
		Time		1ns~100ms	$U=0.06 \mu s$		
		phase		(0~360)°	$U=0.6^\circ$		
		Drive signal frequency		10Hz~10MHz	$U_{rel}=1.6 \times 10^{-7}$		
		Drive signal amplitude		(0.1~10) V	$U_{rel}=2.1\%$		
19	*Impedance/Material Analyzers	DC voltage	Calibration Specification for RF Impedance/Material Analyzers JJF 1127	(0.01~40)V	$U_{rel}=0.10\%$		
		signal level		-40dBm~20dBm	$U=0.3dB$		
		Test signal frequency		100kHz~3GHz	$U_{rel}=4 \times 10^{-7}$		
		Admittance		(1MHz~3GHz)1 μ S~13mS	$U_{rel}=2.2\%$		
		Impedance		(1MHz~3GHz)1m Ω ~22k Ω	$U_{rel}=2.2\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		phase		(1MHz~3GHz)(-1.57~1.57)rad	$U=16\text{mrad}$		
20	*Standard Capacitors	Capacitors	Verification Regulation of Standard Capacitors JJG 183	1pF~10 μ F (100Hz~10kHz)	$U_{\text{rel}}=0.03\%$		
				10 μ F~10mF (1kHz)	$U_{\text{rel}}=0.05\%$		
21	*Standard Inductors	Inductors	Verification Regulation of Standard inductors JJG 726	1 μ H~10 μ H (1kHz)	$U_{\text{rel}}=0.6\%$		
				10 μ H~100 μ H (1kHz)	$U_{\text{rel}}=0.10\%$		
				100 μ H~100H (1kHz)	$U_{\text{rel}}=0.04\%$		
22	*Digital LCR meter	Resistance	Verification Regulation for wide digital RLC meter GJB/J5412, Alternating Current Bridge JJG 441	1 Ω ~10 Ω (1kHz)	$U_{\text{rel}}=0.06\%$		
				10 Ω ~100 Ω (1kHz)	$U_{\text{rel}}=0.04\%$		
				100 Ω ~100k Ω (1kHz)	$U_{\text{rel}}=0.03\%$		
				1m Ω (100Hz~1kHz)	$U_{\text{rel}}=0.8\%$		
				10m Ω (100Hz~10kHz)	$U_{\text{rel}}=0.6\%$		
				100m Ω (100Hz~1000kHz)	$U_{\text{rel}}=0.3\%$		
				1 Ω (100Hz~1kHz)	$U_{\text{rel}}=0.03\%$		
				1 Ω (1kHz~1MHz)	$U_{\text{rel}}=0.06\%$		
				10 Ω ,100 Ω ,1k Ω (100Hz~1kHz)	$U_{\text{rel}}=0.03\%$		
				10 Ω ,100 Ω ,1k Ω (1kHz~1MHz)	$U_{\text{rel}}=0.06\%$		
				10 Ω ,100 Ω ,1k Ω (1MHz~3MHz)	$U_{\text{rel}}=0.17\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Capacitance		10 Ω, 100 Ω, 1k Ω (3MHz~4MHz)	$U_{rel}=0.23\%$		
				10 Ω, 100 Ω, 1k Ω (4MHz~5MHz)	$U_{rel}=0.29\%$		
				10 Ω, 100 Ω, 1k Ω (5MHz~10MHz)	$U_{rel}=0.8\%$		
				10 Ω, 100 Ω, 1k Ω (10MHz~13MHz)	$U_{rel}=1.2\%$		
				10k Ω, 100k Ω (100Hz~1kHz)	$U_{rel}=0.03\%$		
				10k Ω, 100k Ω (1kHz~100kHz)	$U_{rel}=0.06\%$		
				10k Ω, 100k Ω (100kHz~1MHz)	$U_{rel}=0.12\%$		
				100pF~1 μ F (1kHz)	$U_{rel}=0.06\%$		
				1 μ F~100 μ F(1kHz)	$U_{rel}=0.07\%$		
				1pF, 10pF, 100pF, 1000pF (100Hz~1MHz)	$U_{rel}=0.12\%$		
				10nF, 100nF(100Hz~100kHz)	$U_{rel}=0.12\%$		
				1000nF(100Hz~10kHz)	$U_{rel}=0.12\%$		
				Inductance	1 μ H~10 μ H (1kHz)	$U_{rel}=0.6\%~6\%$	
		10 μ H~100 μ H (1kHz)	$U=0.09\%~0.6\%$				
		100 μ H~1H (1kHz)	$U_{rel}=0.07\%~0.09\%$				
		1 μ H (100Hz~1MHz)	$U_{rel}=1\%$				




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
			ilac-M	5 μ H, 10 μ H (100Hz~5MHz)	$U_{rel}=1\%$			
				50 μ H, 100 μ H (100Hz~5MHz)	$U_{rel}=0.5\%$			
				500 μ H, 1000 μ H (100Hz~1MHz)	$U_{rel}=0.3\%$			
				5mH (100Hz~500kHz)	$U_{rel}=0.3\%$			
				Frequency	10Hz~1MHz			$U_{rel}=6 \times 10^{-6}$
				Voltage	10mV~100V(10Hz~20kHz)			$U_{rel}=0.2\%$
					10mV~100V(20kHz~100kHz)			$U_{rel}=0.8\%$
					10mV~100V(100kHz~2MHz)			$U_{rel}=1.4\%$
loss	$1 \sim 10^{-5}$	$U=0.4\%R+0.0002$						
23	*r A. C. Resistance Boxes	AC Resistance	C. S.for AC Resistance Boxes JJF 1636	(0.01~0.1) Ω (100Hz~1kHz)	$U_{rel}=0.15\%$			
				(0.1~1000) Ω (60Hz~10kHz)	$U_{rel}=0.08\%$			
				(1~100)k Ω (60Hz~10kHz)	$U_{rel}=0.03\%$			
				(100~1000)k Ω (1kHz)	$U_{rel}=0.08\%$			
24	*RF Coaxial Impedance Standard	Impedance(short,load)	Verification Regulation of RF Coaxial Impedance Standard JJG(DZ) 306001	(0.01~34) Ω (1MHz~3GHz)	$U=0.5 \Omega$			
				50 Ω (1MHz~3GHz)	$U=0.8 \Omega$			




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		admittance(open)		10 μ S~71 μ S(1MHz~3GHz)	U=6 μ S		
25	*Transistor Character Graphic Meter	Step Current	C.S.for Semiconductor Device Curve Tracers JJF 1236	100 μ A~10A	U _{rel} =1.4%		
		Step Voltage		20mV~10V	U _{rel} =1.4%		
		Collector Voltage		20mV~1000V	U _{rel} =1.4%		
		Base Voltage		20mV~100V	U _{rel} =1.6%		
		Collector Current		100 μ A~20A	U _{rel} =1.5%		
		Calibration Voltage		20mV~100V	U _{rel} =0.12%		
26	*Communication Testers	RF SignalSource Frequency	C.S. for TDMA-GSM Radio Communication Testers JJF 1131,C.S. for CDMA Digital Radio Communication Testers JJF 1177,C.S. for LTE Digital Radio Communication Testers JJF 1443,C.S. for TD-SCDMA Digital Radio Communication Testers JJF 1204,C.S. for WCDMA Digital Radio Communication Testers JJF 1276,C.S. for RF Communication Test Set JJF 1065	30MHz~3.8GHz	U _{rel} =1.2×10 ⁻⁷		
		RF SignalSource Electrical Level		(-20~ 0)dBm(30MHz~3.8GHz)	U=0.20dB		
		RF SignalSource SpectrumPurity		(-20~ -110)dBm(30MHz~3.8GHz)	U=0.38dB		
		RF SignalSource Phase Noise		(-80~-25)dBc	U=1.5dB~2.1dB		
		RF AnalyzerFrequency		(-110~-25) dBc/Hz	U=3dB		
				Bias frequency(20~1990)kHz	U _{rel} =6×10 ⁻⁸		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RF Analyzer Electrical Level		(-50dBm ~ 20dBm)(30MHz ~ 3.8GHz)	$U=0.20\text{dB}$		
		AF Signal Source Frequency		10Hz ~ 40Hz	$U_{\text{rel}}=0.035\%$		
		AF Signal Source Voltage		40Hz ~ 20kHz	$U_{\text{rel}}=0.012\%$		
				10mV ~ 100mV(40Hz ~ 20kHz)	$U_{\text{rel}}=0.054\% \sim 0.12\%$		
				100mV ~ 1V(40Hz ~ 20kHz)	$U_{\text{rel}}=0.042\% \sim 0.11\%$		
				1V ~ 5V(40Hz ~ 20kHz)	$U_{\text{rel}}=0.042\% \sim 0.07\%$		
		AF Signal Source Distortion		(0.001 ~ 0.3)%(20Hz ~ 20kHz)	$U_{\text{rel}}=12\%$		
		AF Analyzer Frequency		20Hz ~ 2kHz	$U_{\text{rel}}=0.015\% \sim 0.003\%$		
				2kHz ~ 40kHz	$U_{\text{rel}}=0.003\%$		
		AF Analyzer Voltage		20mV ~ 33mV(10Hz ~ 45Hz)	$U_{\text{rel}}=5\% \sim 3.2\%$		
				20mV ~ 33mV(45Hz ~ 10kHz)	$U_{\text{rel}}=5\% \sim 3.2\%$		
				20mV ~ 33mV(10kHz ~ 20kHz)	$U_{\text{rel}}=5\% \sim 3.2\%$		
				20mV ~ 33mV(20kHz ~ 40kHz)	$U_{\text{rel}}=5\% \sim 3.2\%$		
				33mV ~ 330mV(10Hz ~ 45Hz)	$U_{\text{rel}}=3.2\% \sim 0.44\%$		
				33mV ~ 330mV(45Hz ~ 10kHz)	$U_{\text{rel}}=3.2\% \sim 0.32\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				33mV~330mV(10kHz~20kHz)	$U_{rel}=3.2\% \sim 0.34\%$		
				33mV~330mV(20kHz~40kHz)	$U_{rel}=3.2\% \sim 0.38\%$		
				0.33V~3.3V(10Hz~45Hz)	$U_{rel}=0.42\% \sim 0.18\%$		
				0.33V~3.3V(45Hz~10kHz)	$U_{rel}=0.32\% \sim 0.05\%$		
				0.33V~3.3V(10kHz~20kHz)	$U_{rel}=0.34\% \sim 0.10\%$		
				0.33V~3.3V(20kHz~40kHz)	$U_{rel}=0.42\% \sim 0.18\%$		
				3.3V~5V(10Hz~45Hz)	$U_{rel}=0.26\% \sim 0.23\%$		
				3.3V~5V(45Hz~10kHz)	$U_{rel}=0.08\% \sim 0.07\%$		
				3.3V~5V(10kHz~20kHz)	$U_{rel}=0.19\% \sim 0.16\%$		
				3.3V~5V(20kHz~40kHz)	$U_{rel}=0.40\% \sim 0.34\%$		
		Channel power		(-70~20)dBm	$U=0.6\text{dB}$		
		Adjacent channel power rate		(0.01~90)dBc	$U=0.5\text{dB}$		
		Frequency error		(-1~1)MHz	$U=9\text{Hz}$		
		EVM		(0.01~10)%	$U=0.6\%$		
				(0.01~10%)(CDMA)	$U=1.2\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date					
		Mag error	ilac-MRA	(0.01~10)%	$U=0.6\%$							
		Phase error		(0.01~15)°	$U=0.6^\circ$							
		FSK error		(0.01~10)%	$U=2.0\%$							
		Waveform quality factor		(0.8~1.0)	$U_{rel}=0.18\%$							
		Frequency Measurement error		(-1~1)MHz	$U=9\text{Hz}$							
		Channel power Measurement		(-70~10)dBm	$U=0.6\text{dB}$							
		Adjacent channel power rate Measurement		(0.01~90)dBc	$U=0.5\text{dB}$							
		EVM Measurement		(0.01~10)%	$U=0.6\%$							
				(0.01~10)% (CDMA)	$U=1.2\%$							
		Phase error Measurement		(0.01~15)°	$U=0.6^\circ$							
		Mag error Measurement		(0.01~10)%	$U=0.6\%$							
		27		*Telephone analyzer	Feed voltage			C.S. for Telephone analyzer JJF(DZ) 30702	(10~100)V	$U_{rel}=0.08\%$		
				Feed current				(10~150)mA	$U_{rel}=0.2\%$			
	level		(-50~0)dBm	$U=0.2\text{dB}$								
	signal frequency		1Hz~20kHz	$U_{rel}=0.05\%$								




在线扫码获取验证

No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
28	*Bluetooth Test Set	Output Frequency	Calibration Specification for Bluetooth Test Set JJF 1278	2.402GHz~2.480GHz	$U_{rel}=1.2 \times 10^{-7}$		
		Output Level		(0~-20)dBm(2.402GHz~2.480GHz)	$U=0.20$ dB		
				(-20~-90)dBm(2.402GHz~2.480GHz)	$U=0.38$ dB		
		Harmonic		(-80~-30)dBc(2.402GHz~2.480GHz)	$U=1.8$ dB		
		Phase Noise		(-110~-30)dBc/Hz(2.402GHz~2.480GHz)	$U=1.1$ dB		
		Frequency Measurement level		(2402~2480) MHz	$U_{rel}=6 \times 10^{-8}$		
		Measurement level		(-70~13) dBm(2.402GHz~2.480GHz)	$U=0.20$ dB		
		Mag error		(0.01~10)%	$U=1.2$ %		
		Modulation level		(-50~-10) dBm	$U=0.25$ dB		
		EVM		(0.01~10)%	$U=1.2$ %		
		Frequency error Measurement level		(-1~1)MHz	$U=10$ Hz		
		Modulation level Measurement		(-50~-10) dBm	$U=0.4$ dB		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		EVM Measurement		(0.01~10)%	$U=1.2\%$		
		Frequency error Measurement		(-1~1)MHz	$U=10\text{Hz}$		
29	*WLAN Test Set	Output Frequency	Calibration Specification for WLAN Test Set JJF 1277	2.412GHz~5.945GHz	$U_{\text{rel}}=1.2 \times 10^{-7}$		
		Output Level		(0~ -20)dBm(2.412GHz~5.945GHz)	$U=0.20\text{dB}$		
				(-20~ -90)dBm(2.412GHz~5.945GHz)	$U=0.38\text{dB}$		
		Harmonic		(-80~-25)dBc (2412~2484) MHz	$U=1.8\text{dB}$		
				(-80~-25)dBc (4900~5945) MHz	$U=2.4\text{dB}$		
		Phase Noise		(-110~-30)dBc/Hz(2.412GHz~5.945GHz)	$U=1.1\text{dB}$		
		level Measurement		(-50~13) dBm(2.412GHz~5.945GHz)	$U=0.20\text{dB}$		
		Modulation level		(-50~-10) dBm(2.412GHz~5.945GHz)	$U=0.25\text{dB}$		
		EVM		(0.01~10)%	$U=1.2\%$		
		Mag error		(0.01~10)%	$U=1.2\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Frequency error Measurement	ilac-MRA	(-1~1)MHz	$U=10\text{Hz}$		
		Phase error		(0.01~15)°	$U=0.6^\circ$		
		Modulation level Measurement		(-50~10) dBm	$U=0.25\text{dB}$		
		EVM Measurement		(0.01~10)%	$U=1.2\%$		
		Frequency error Measurement		(-1~1)MHz	$U=10\text{Hz}$		
		Phase error Measurement		(0.01~15)°	$U=0.6^\circ$		
30	*Antenna Analyzers	Frequency	C. S. of Cable and Antenna Analyzers JJF 1740	25Hz~6GHz	$U_{\text{rel}}=4.3 \times 10^{-7}$		
		VSWR		1.01~2.0	$U_{\text{rel}}=6.2\%$		
		Terminal load reflection loss		(0~50) dB(10MHz~6GHz)	$U=0.22\text{dB}$		
		Power		(-50~20)dBm10MHz~6GHz	$U=0.20\text{dB}$		
31	*Range calibrator of power meter	DC Voltage	Verification regulation for range calibrator of power meter GJB 8819	0.01V~5V	$U_{\text{rel}}=4 \times 10^{-5}$		
				5V~20V	$U_{\text{rel}}=2 \times 10^{-5}$		
		Resistance		30 Ω ~ 100 Ω	$U_{\text{rel}}=4 \times 10^{-5}$		
				100 Ω ~ 10k Ω	$U_{\text{rel}}=2 \times 10^{-5}$		
32	*Road vehicles-Electrical Transient	Peak pulse voltage	Calibration specification for Road vehicles-Electrical Transient Conduction	1V~4kV	$U_{\text{rel}}=3\%$		




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Conduction Disturbance Simulator	Pulse voltage rise and fall time	Disturbance Simulator JJF(DZ) 0019	1ns~1s	$U_{rel}=5\%$		
		Pulse duration		10ns~10s	$U_{rel}=1.2\%$		
		Test voltage		1V~100V	$U_{rel}=0.1\%$		
33	*TS-109 Semi-automatic Sorting Instrument for Electrolytic Capacitors	Capacitance	Verification regulation of TS-109 semi-automatic sorting instrument for electrolytic capacitors JJG(DZ) 05009	(0.1~1000) μ F,(100Hz)	$U_{rel}=0.2\%$		
				(0.1~100) μ F,(1kHz)	$U_{rel}=0.2\%$		
				(0.1~1) μ F,(10kHz)	$U_{rel}=0.3\%$		
		Voltage		0.1V~100V	$U_{rel}=0.4\%$		
				100V~500V	$U_{rel}=0.3\%$		
		Current		1 μ A~10mA	$U_{rel}=0.3\%$		
		Loss factor		0.0001~1,(100Hz~10kHz)	$U=0.002$		
34	*Oscilloscope Calibrator	DC calibration voltage	Verification regulation of oscilloscope calibrator JJG 278	1mV~1V	$U=8.6 \times 10^{-5} V_x + 0.0003mV$		
				1V~10V	$U=8.6 \times 10^{-5} V_x + 0.0005mV$		
				10V~100V	$U=8.6 \times 10^{-5} V_x + 0.003mV$		
				100V~200V	$U=8.6 \times 10^{-5} V_x + 0.01mV$		
		Square wave calibration voltage		1mV~100mV	$U=8 \times 10^{-5} V_x + 0.002mV$		
				0.1V~10V	$U=8 \times 10^{-5} V_x + 0.02mV$		
				10V~100V	$U=0.024\% V_x + 0.2mV$		
				100V~200V	$U=0.046\% V_x + 2mV$		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Fast edge rise/fall time		150ps~1ms	$U_{rel}=2.5\%$		
		Waveform distortion overshoot		0.01%~20%	$U_{rel}=1.8\%$		
		Fast-edge pulse duty cycle		1%~99%	$U_{rel}=0.1\%$		
		Hour markers		500ps~10s	$U_{rel}=5.0 \times 10^{-8}$		
		Stabilize the amplitude of the signal		(-60~20)dBm, (9kHz~3.05GHz)	$U=0.17\text{dB}$		
				(-60~20)dBm, (3.05~6.6)GHz	$U=0.15\text{dB}$		
				(-60~20)dBm, (6.6~13.2)GHz	$U=0.19\text{dB}$		
				(-60~20)dBm, dBm, (13.2~18)GHz	$U=0.23\text{dB}$		
		Amplitude stabilization signal frequency		10Hz~18GHz	$U_{rel}=5.0 \times 10^{-8}$		
		Amplitude stabilization signal harmonics		1dB~140dB,(9kHz~3GHz)	$U=0.65\text{dB}$		
		Waveform generator amplitude		1dB~140dB,(3GHz~6GHz)	$U=1.2\text{dB}$		
				10mV~1000mV, (10Hz~20kHz)	$U_{rel}=0.035\%$		
				1V~50V, (10Hz~20kHz)	$U_{rel}=0.024\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
		Pulse period	ilac-MRA CHINA NATIONAL ACCREDITATION CENTER FOR COFFEE QUALITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	4ns~100ms	$U_{rel}=5 \times 10^{-8}$			
		Pulse width		2ns~50ms	$U_{rel}=0.1\%$			
		Resistance measurement		10 Ω ~ 100 Ω	$U_{rel}=0.12\%$			
				100 Ω ~ 1M Ω	$U_{rel}=0.024\%$			
				1M Ω ~ 15M Ω	$U_{rel}=0.24\%$			
35	*Dynamical Signal Analyzer	Frequency	V.R.of Dynamical Signal Analyzer JJG 834	10Hz~200kHz	$U_{rel}=1.1 \times 10^{-6}$			
		Amplitude		10mV~10V, (10Hz~20kHz)	$U_{rel}=0.10\%$			
					10mV~10V, (20~50)kHz	$U_{rel}=0.16\%$		
					10mV~10V, (50~200)kHz	$U_{rel}=0.3\%$		
		Dynamic range		(0~120)dB	$U=0.20\text{dB}$			
		Channel conformance		(-1~1)dB	$U=0.02\text{dB}$			
36	*Components of Lightning Testers	DC Voltage	C. S. for Components of Lightning Testers JJF(Gui) 18	(0.1~2)kV	$U_{rel}=0.12\%$	合格评定 国家认证认可 委员会 认可证书专用章		
				(1~2)kV	$U_{rel}=0.5\%$			
		DC Current		(0.1~2000) μA	$U_{rel}=0.3\%$			
		Rate Of Voltage Rise		100V/s~1000V/s	$U_{rel}=3\%$			
37	*VDR DC Parameter Testers	DC Voltage	C. S. for VDR DC Parameter Testers JJF(Zhe)1088	(0.1~2)kV	$U_{rel}=0.12\%$			
				(1~2)kV	$U_{rel}=0.5\%$			




No. CNAS L2336

第 116 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Current		(0.1~2000) μ A	$U_{rel}=0.3\%$		
		Rate Of Voltage Rise		100V/s~1000V/s	$U_{rel}=3\%$		
38	*Transformer Winding Deformation Testers	Attenuation of amplitude Sweep frequency	Calibration Specification for Transformer Winding Deformation Testers JJF(Zhe) 1138	(-80~0) dB 10Hz~1MHz	$U=0.3dB$ $U_{rel}=0.002\%$		
39	*High Voltage Dielectric Loss Tester	Electric capacity Dielectric Loss Factor	Verification Regulation of High Voltage Dielectric Loss Tester JJG 1126	100pF(50Hz) (0.0002~0.1)	$U_{rel}=0.3\%$ $U=0.3\%Dx+0.0002$		
40	Electric spark leak detectors	Pulse voltage	Calibration Specification for electric spark leak detectors JJF(Lu) 101	(0.5~20) kV	$U_{rel}=3.0\%$		
41	Artificial Mains Networks	Impedance	Calibration Specification of Artificial Mains Networks JJF 1705	1 Ω ~ 100 Ω (9kHz~108MHz) -180° ~ 180° (9kHz~108MHz)	$U_{rel}=7\%$ $U=3.9^\circ \sim 4.0^\circ$		
		Voltage divider coefficient		-20dB~0dB (9kHz~108MHz)	$U=1.24dB$		
42	Electromagnetic Clamps	Input Impedance	Calibration Specification of Electromagnetic Clamps JJF (DZ) 0031	50 Ω ~ 350 Ω (0.1MHz~100MHz)	$U_{rel}=11\%$		
		Decoupling factor		-10dB~0dB (0.1MHz~100MHz)	$U=2.2dB$		
		Coupling factor		-5dB~3dB (0.1MHz~1000MHz)	$U=2.2dB \sim 2.3dB$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
43	Coupling/Decoupling Networks of Conducted Disturbances Induced by Radio-Frequency Fields Testing	Common-Mode Impedance Modulus	Calibration Specification of Coupling/Decoupling Networks of Conducted Disturbances Induced by Radio-Frequency Fields Testing JJF 2079	5 Ω ~ 300 Ω (0.15MHz ~ 230MHz)	$U_{rel}=6.4\%$		
		Decoupling attenuation		10dB ~ 50dB (0.15MHz ~ 230MHz)	$U=2.4dB \sim 2.6dB$		
		Coupling coefficient		-5dB ~ 5dB (0.15MHz ~ 230MHz)	$U=0.54dB$		
		Adapter Insertion Loss		5dB ~ 15dB (0.15MHz ~ 230MHz)	$U=0.32dB$		
Time and Frequency							
1	*Microwave Frequency Counters	Frequency Measurement and input sensitivity	Verification Regulation of Microwave Frequency Counters JJG 841	10mV ~ 300mV(0.1Hz ~ 1MHz)	$U=1.8mV$		
				(-33 ~ -20)dBm(1MHz ~ 18GHz)	$U=1dB$		
		Frequency Measurement error		1Hz ~ 18GHz	$U_{rel}=1.5 \times 10^{-8}$		
		Crystal oscillator		10MHz	$U_{rel}=3 \times 10^{-8}$		
2	*Universal Counters	Frequency Measurement and input sensitivity	Calibration Specification for Universal Counters JJF 2196	10mV ~ 300mV(0.1Hz ~ 1MHz)	$U=1.8mV$		
				(-33 ~ -20)dBm(1MHz ~ 18GHz)	$U=1dB$		
		Frequency Measurement error		1Hz ~ 18GHz	$U_{rel}=1.5 \times 10^{-8}$		
		Period Measurement and input		(-33 ~ -20)dBm(1ps ~ 1 μs)	$U=1dB$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		sensitivity		10mV~300mV(1 μs~1s)	$U=1.8mV$		
		Frequency Measurement error		1 μs	$U_{rel}=1.5 \times 10^{-8}$		
		Crystal oscillator		10MHz	$U_{rel}=3 \times 10^{-8}$		
3	*Frequency Meters	Frequency	Verification Regulation of Frequency Meters JJG 603	10Hz~10kHz	$U_{rel}=0.03\%$		
4	*Stopwatches	time interval	C.S.for Stopwatches JJF 2195	electronic stopwatch: (1~3600)s	$U=15ms$		
				Electronic stopwatch daily deviation: (-10~10)s/d	$U=0.05s$		
				Mechanical stopwatch: (1~3600)s	$U=0.11s$		
				Electric stopwatch: (1~3600)s	$U=5ms$		
5	*Time Delay Relays	time error	Calibration Specification for Time Delay Relays JJF 1282	0.01s~10.00s	$U=7ms$	Only Calibrate electronic time relays	
				10.00s~60.00s	$U=8ms$		
				0.1s~60.0s	$U=0.06s$		
				1s~2h	$U=0.6s$		
6	*Time interval meters	Time interval	Verification Regulation of Time Interval Meters JJG 238	50ns~1000s	$U=2 \times 10^{-7}T_x+15ns$		
		frequency		5MHz,10MHz	$U_{rel}=1 \times 10^{-9}$		
7	*Online Timer	Time Interval	Calibration Specification of Online Timer JJF (Liao) 492	(1~18000)s	$U=2 \times 10^{-5}T+0.62s$		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
8	*Pulse Counters	Frequency of Internal Time Base	Calibration Specification for Pulse Counters JJF1686	5MHz,10MHz	$U_{rel}=1.2 \times 10^{-8}$		
		Pulse Counting		1~1000000	$U=0.0006\%T_x+2$		
9	*Vibrating Wire Frequency Readouts	Excitation signal	Calibration Specification for Vibrating Wire Frequency Readouts JJF1401	1mV~100V	$U_{rel}=1.8\%$		
		Frequency		300Hz~6000Hz	$U_{rel}=0.007\%$		
Optics measuring instrument							
1	Optical Power Meter in Telecommunication	Optical Power	V.R.of Optical Power Meter in Telecommunication JJG965	(-50~0)dBm (850nm, 1310nm, 1550nm)	$U=0.18dB$		
2	Stabilized Laser Sources for Optical Transmit	Output Power	V.R.of Stabilized Laser Sources for Optical Transmit JJG958	(-50~0)dBm	$U=0.18dB$		
		Light Wavelength		(1270~1580)nm	$U=0.06nm$		
		Spectral bandwidth		(0.2~10) nm	$U=0.06nm$		
3	Optical Attenuator for Telecommunications	Insertion Loss	C.S. of Optical Attenuator for Telecommunications JJF1199	(0.01~3)dB	$U=0.18dB$		
		Attenuation Indication		(0.01~50)dB (850nm, 1310nm, 1550nm)	$U=0.18dB$		
4	Illuminance Meter	Illumination	V.R.of Illuminance Meter JJG 245	(50~3000) lx	$U_{rel}=1.5\%$		
5	Color Temperature Meter	Color temperature	C.S. for Colour Temperature Meters JJF 2100	(2300~3200)K	$U=16K \sim 26K$		
				(3200~9000)K	$U=52K \sim 1.7 \times 10^2 K$		
		chroma		Full color gamut	$U=0.005$		



No. CNAS L2336

第 120 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rendering Index		0~100	$U=0.4$		
6	*Clarity Test Equipment	Illumination	C.S. for Clarity Test Equipment JJF 1287	(100~3000) lx	$U_{rel}=8\%$		
		Time		(0~30) s	$U=0.2s$		
7	Luminance Meter	luminance	V.R. of Luminance Meter JJG 211	(50~1000)cd/m ²	$U_{rel}=2.2\%$	Do not make shading cylinder type	
		chroma		Panchromatic Gamut	$U=0.005$		
8	*Whiteness Meter	Whiteness	V.R. of the Whiteness Meters JJG 512	50~100	$U(R_{457})=1.8$		
9	*Colorimeters and Color Difference Meters	Chroma	V.R. of Colorimeters and Color Difference Meters JJG 595	Y:0~100	$U(Y)=2.2$		
				Panchromatic Gamut	$U=0.0060$		
10	*Water Colorimeters	chroma	C. S. for Water Colorimeters JJF 1689	Digital Instrument: (1~70) PCU	$U=1$ PCU		
				Visual Instrument: (1~70) PCU	$U=6$ PCU		
11	Ultraviolet Radiometers	UV Irradiance	V.R. of Ultraviolet Radiometers JJG 879	(0.1~100) mW/cm ²	$U_{rel}=20\%$		
12	*Abbe Refractometer	Refraction Ratio	V.R. of Abbe Refractometer JJG 625	$n_D: 1.47000\sim 1.67000$	$U(n_D)=6\times 10^{-5}$		
		Average dispersion		$n_F-n_C: 0.0070\sim 0.0210$	$U(n_F-n_C)=1.5\times 10^{-4}$		
13	*Reflection Densitometers	optical density	C.S. for Reflection Densitometers JJF 1492	$0.00<D\leq 1.00$	$U=0.03$		
				$1.00<D\leq 2.00$	$U=0.04$		
14	*Diffuse Transmission Visual	Transmission density	V.R. of Diffuse Transmission Visual Densitometers JJG 920	$0.0<D\leq 2.0$	$U=0.020$		
				$2.0<D\leq 4.0$	$U=0.024$		



No. CNAS L2336

第 121 页 共 187 页


The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Densitometers			$4.0 < D \leq 5.0$	$U=0.032$		
15	*Specular Gloss Meters	Specular Gloss	V.R. of Specular Gloss Meters and Gloss Plates JJG 696	(0.0~120.0)GU	$U=1.4GU$		
16	*Reflectometers	reflectivity	C.S. for Reflectometers JJF 1232	(50~100)%	$U=2.3\%$		
17	*Hazemeter	Haze	C.S. for Hazemeters JJF 1303	0.1~35.0	$U=0.30$		
		Transmission		(10~100)%	$U=0.8\%$		
18	*Standard Light Sources Boxes	Illumination	C.S.(FZ) for Standard Light Source Boxes JJF(FZ) 055	(200~3000)lx	$U_{rel}=6\%$		
		Color temperature		(2300~7500)K	$U=37K \sim 1.4 \times 10^2K$		
19	*UV Analyzers	Irradiance	C.S. for UV Analyzers JJF 1936	(10~1000) $\mu W/cm^2$	$U_{rel}=22\%$		
		wavelength		(254~365)nm	$U=0.7nm$		
20	*Electrical Light Source Spectrum Test System	wavelength	C.S. for Electrical Light Source Spectrum Test System JJF(Su) 153	(250~580)nm	$U=0.13nm$		
		Luminous flux		(1 ~2000)lm	$U_{rel}=2\%$		
		chromaticity coordinates		Panchromatic Gamut	$U=0.005$		
		Correlated color temperature		(2000~3400)K	$U=26K$		
21	Cathode Ray Tubes (CRT) Color Analyzers	Brightness	C.S. for Cathode Ray Tubes (CRT) Color Analyzers JJF 1079	(20~400)cd/m ²	$U_{rel}=2.6\%$		
		Chroma		Panchromatic Gamut	$U=0.004$		
22	Retroreflectometers	Retroreflective luminance coefficient	Calibration Specification for Retroreflectometers JJF1809	(0.1~1999)mcd · m ² · lx ⁻¹	$U_{rel}=15\%$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
23	*Automatic Lamp-Inspecting Machines	Illuminance	Calibration Specification for Automatic Lamp-Inspecting Machines JJF 1824	(100~3000)lx	$U_{rel}=5\%$		
24	Ultraviolet Radiant Exposure Meters	Radiant Exposure	C.S. for Ultraviolet Radiant Exposure Meters JJF 2130	(0.01~6)J/cm ² (320 nm~390 nm)	$U_{rel}=9\%$		
Chemistry measuring instrument SCHEDULE OF ACCREDITATION CERTIFICATE							
1	*Extrusion Plastometer	Temperature	V.R. of Extrusion Plastometer JJG 878	(125~400)°C	$U=0.2^{\circ}\text{C}$		
		Rate		Melt Flow Rate:(0.1~25)g/10min	$U_{rel}=5.0\%$		
		Length		(1~3) mm	$U=0.001\text{ mm}$		
		Mass		(0~200) g	$U=0.002\text{ g}$		
		Time		(0~10) min	$U=0.2\text{ s}$		
2	*Differential Scanning Calorimeters	Heating rate	V.R. of the Differential Scanning Calorimeters JJG 936	(1~20) °C/min	$U_{rel}=6\%$		
		Temperature		Melting temperature:(100~500) °C	$U=(0.3\sim 1.0)^{\circ}\text{C}$		
				Phase transition temperature:(100~600) °C	$U=(0.8\sim 1.5)^{\circ}\text{C}$		
		Heat		Melting heat:(20~110)J/g	$U=(0.4\sim 1.5)\text{ J/g}$		
3	*Open/Closed Cup Flash Point Testers	Temperature	Calibration Specification for Open/Closed Cup Flash Point Testers JJF1385	Open Cup:(80~200) °C	$U=8^{\circ}\text{C}$		
				Open Cup:(200~300) °C	$U=(8.2\sim 11)^{\circ}\text{C}$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Closed Cup: (40~110) °C	U=6.4 °C		
				Closed Cup: (110~300) °C	U=6.8 °C		
4	*Thermogravimetric Analyzers	Temperature	V.R. of Thermogravimetric Analyzers JJG 1135	Curie point:(20~800) °C	U=(1.0~2.0) °C		
		Mass		(1~20) mg	U=(7~12) μg		
5	*Melting-point Measurement Instruments	Temperature	V.R. of Melting-point Measurement Instruments JJG 701	Capillary method:(50~300)°C	0.2 °C/min: U=0.3 °C		
				Hot stage:(50~300) °C	1.0 °C/min:U=0.4 °C		
		Heating rate		(0.2~2.0) °C/min	U _{rel} =6 %		
6	*Thermogravimetric Moisture Meters	Content	Verification Regulation of Thermogravimetric Moisture Meters JJG 658	Water content:5 %~95 %	U _{rel} =0.06 %		
		Mass		1 mg~1100 g	U= (0.06~21) mg		
7	*pH Meters	Voltage	Verification Regulation of pH Meters JJG 919	(10~1000) mV	U _{rel} =3×10 ⁻³		
		Resistance		(1~3) G Ω	U _{rel} =1.3 %		
		pH		0~14	U=0.0002		
8	*Verficing Meter for Chromatograph	DC Voltage	Verification Regulation of Verficing Meter for Chromatograph JJG 937	(0.01~10) V	U _{rel} =0.02 %		
		Resistence		(1~1000) Ω	U _{rel} =0.03 %		
		DC Current		(0.1~10) mA	U _{rel} =0.3 %		
9	*Ultraviolet, visible and near-infrared	Wavelength	V.R. of Ultraviolet, Visible, Near-Infrared Spectrophotometers JJG 178	(240~900) nm	U=0.5 nm		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	spectrophotometer	Wavelength		(1000~2000) nm	U=0.8 nm		
		Transmission		(0~35) %	U=0.5 %		
10	*Dispersive Infrared Spectrophotometer	Wave number	V.R. of Dispersive Infrared Spectrophotometers JJG 681	(3100~800)cm ⁻¹	U=0.4cm ⁻¹		
11	*Atomic Absorption Spectrophotometer	Concentration	V.R. of Atomic Absorption Spectrophotometers JJG 694	Flame method (Cu) detection limit: ≤0.02 μ g/mL	U=0.006 μ g/mL		
				Graphite furnace method (Cd) detection limit: ≤ 4pg	U=0.3 pg		
12	*Fluorescence Spectrophotometer	Concentration	V.R. of Fluorescence Spectrophotometer JJG 537	Detection Limit: ≤1×10 ⁻⁸ g/mL	U _{rel} =2.7 %		
13	*Emission Spectrometer	Wavelength	V.R. of Emission Spectrometer JJG 768	ICP spectrometer:(190~1100) nm	U=0.01 nm		
		Concentration		Detection limit of ICP spectrometer(Zn):≤0.01 mg/L	U _{rel} =7 %		
				Detection limit of ICP spectrometer(Ni):≤0.03 mg/L	U _{rel} =7 %		
				Detection limit of ICP spectrometer(Mn、Ba):≤0.005 mg/L	U _{rel} =7 %		
				Detection limit of ICP spectrometer(Cr、Cu):≤0.02 mg/L	U _{rel} =7 %		




No. CNAS L2336

第 125 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Content		Detection limit of direct reading spectrometer(C、Mn): ≤0.02 %	$U_{rel}=7\%$		
				Detection limit of direct reading spectrometer(Si、Ni): ≤0.02 %	$U_{rel}=8\%$		
				Detection limit of direct reading spectrometer(Cr、V): ≤0.01 %	$U_{rel}=8\%$		
14	*Flame photometer	Concentration	V.R. of Flame Photometer JIG 630	Detection limit(K): ≤0.004 mmol/L	$U=0.0005\text{ mmol/L}$		
				Detection limit(Na): ≤0.008 mmol/L	$U=0.0024\text{ mmol/L}$		
15	*Polarimeter	Polarimete	V.R. of Polarimeter and Saccharimeter JIG 536	(-45~45)°	$U=0.003^\circ$		
16	*Hand Saccharimeter(Content-meter) and Hand Refractometer	Saccharimeter(Content-meter)	V.R. of Hand Saccharimeter(Content-meter) and Hand Refractometer JIG 820	5%~60%	$U=0.07\%$		
		Refractivity		1.3330~1.5200	$U=0.0002$		
17	*energy dispersive X ray fluorescence spectrometer	Energy	Calibration Specification for Energy Dispersive X-Ray Fluorescence Spectrometers JJF 2024	Resolution ≤280eV	$U=13\text{eV}$		
		Content		In Polymer: Pb (60~1200) mg/kg、Hg (8~1200) mg/kg、Cr (40~1200) mg/kg、Br (1400~1500) mg/kg、Cd (5~150) mg/kg	$U_{rel}=10\%$		




No. CNAS L2336

第 126 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Cl in Polymer: (350~400) mg/kg	$U_{rel}=19\%$		
				Pb、Cd in Metal: (0.002%~0.110%)	$U_{rel}=11\%$		
				Cr、Ni in metal: (7%~27%)	$U_{rel}=0.6\%~1.0\%$		
18	*Gas Chromatograph	Flow rate	V.R. of Gas Chromatograph JIG 700	Carrier gas flow rate:(20~150) mL/min	$U_{rel}=5.2\%$		
		Detection limit		TCD Sensitivity: ≥ 800 mV·mL/mg	$U_{rel}=9\%$		
				FID Detection Limit: ≤ 0.5 ng/s	$U_{rel}=9\%$		
				FPD Detection Limit(S): ≤ 0.5 ng/s	$U_{rel}=9\%$		
				FPD Detection Limit(P): ≤ 0.1 ng/s	$U_{rel}=9\%$		
				ECD Detection Limit: ≤ 5 pg/mL	$U_{rel}=9\%$		
				NPD Detection Limit(N): ≤ 5 pg/s	$U_{rel}=9\%$		
				NPD Detection Limit(N):(P) ≤ 10 pg/s	$U_{rel}=9\%$		
19	*Liquid Chromatograph	Flow	V.R. of Liquid Chromatograph JIG 705	(0.2~10) mL/min	$U_{rel}=0.3\%$		
		Detection limit		UV-VisD and DAD Detection limit: $\leq 5 \times 10^{-8}$ g/ml	$U_{rel}=9\%$		
				FLD Detection limit: $\leq 5 \times 10^{-8}$ g/ml	$U_{rel}=9\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				RID Detection limit: $\leq 5 \times 10^{-6}$ g/m	$U_{rel}=10\%$		
				ELSD Detection limit: $\leq 5 \times 10^{-6}$ g/ml	$U_{rel}=11\%$		
20	*Ion Chromatographs	Detection limit	V.R. of Ion Chromatographs JIG 823	Detection limit of conductivity detector: $\leq 0.02 \mu\text{g/mL}$	$U_{rel}=6\%$		
				Detection limit of UV-visible light detector: $\leq 0.02 \mu\text{g/mL}$	$U_{rel}=6\%$		
				Detection limit of electrochemical detector: $\leq 0.02 \mu\text{g/mL}$	$U_{rel}=6\%$		
21	*Capillary electrophoresis apparatus	Wavelength	V. R. of Capillary electrophoresis apparatus JIG 964	(190~360) nm	$U=1\text{ nm}$		
		Detection limit		$\leq 1 \mu\text{g/mL}$	$U_{rel}=15\%$		
22	*Liquid Chromatograph-Atomic Fluorescence	Flow	V. R. of Liquid Chromatograph-Atomic Fluorescence JIG 1151	(0.1~10) mL/min	$U_{rel}=0.3\%$		
		Detection limit		Detection limit (As(V)) : $< 1.0\text{ ng}$	$U_{rel}=10\%$		
				Detection limit (MMA) : $< 0.7\text{ ng}$	$U_{rel}=10\%$		
				Detection limit (DMA) : $< 0.7\text{ ng}$	$U_{rel}=10\%$		
23	*Chromatography-Mass Spectrometer	Mass Concentration	C.S. for Gas Chromatography-Mass Spectrometers JJF 1164	(50~350) u	$U_{rel}=2\%$		
				Electron impact source (EI+) ion trap、single quadrupole、triple quadrupole S/N: $\geq 10:1$	$U_{rel}=15\%$		



No. CNAS L2336


第 128 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Electron bombardment source (EI+) flight time、electrostatic field orbitrap (octafluoronaphthalene) S/N: $\geq 50:1$	$U_{rel}=15\%$		
				Positive chemical ionization source (CI+) ion trap、single quadrupole、triple quadrupole (benzophenone) S/N: $\geq 10:1$	$U_{rel}=15\%$		
				Negative chemical ionization source (CI-) ion trap、single quadrupole (octafluoronaphthalene) S/N: $\geq 10:1$	$U_{rel}=15\%$		
		Temperature	(50~300)°C	$U_{rel}=0.2\%$			
24	*liquid chromatography-mass spectrometers	Signal to Noise Ratio	C.S. for liquid chromatography-mass spectrometers JJF 1317	three pole quadrupole (ESI+、APCI+) S/N: $\geq 30:1$	$U_{rel}=11\%$		
			Three pole quadrupole (ESI-) S/N: $\geq 10:1$	$U_{rel}=15\%$			
			Single quadrupole (ESI+、APCI+、ESI-) S/N: $\geq 10:1$	$U_{rel}=15\%$			
	Mass		(1~600) u	$U_{rel}=2\%$			




No. CNAS L2336

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(600~1000) u	$U_{rel}=3\%$		
				(1000~2500) u	$U_{rel}=5\%$		
		Time		Retention time: (0.1~30) min	$U_{rel}=3\%$		
		Abundance Ratio		Abundance ratio:(40~150) %	$U_{rel}=10\%$		
25	*Analyzers for oil content in water	Concentration	V.R. of Instrument for measuring oil concentration in water JJG 950	(0.1~1000) mg/L	$U_{rel}=4\%$		
26	*Turbidity Meter	Turbidity	V.R.of Turbidimeters JJG 880	(0.1~400) NTU	$U_{rel}=4\%$		
27	*On-line Dissolved Oxygen Meter with Covered-Membrane-Electrode	Concentration	V.R. of On-line Dissolved Oxygen Meter with Covered-Membrane-Electrode JJG(Zhe) 111	(6~15) mg/L	$U=0.12$ mg/L		
		Time		(0~90) s	$U=0.08$ s		
28	Dissolved Oxygen Meter	Concentration	V.R. of Dissolved Oxygen Meter JJG 291	(6~15) mg/L	$U=0.12$ mg/L		
		Time		(0~60) s	$U=0.08$ s		
		Temperature		(0~50) °C	$U=0.2$ °C		
29	*On-line Automatic Determinator of Chemical Oxygen Demand(COD)	Concentration	V.R. of On-line Automatic Determinator of Chemical Oxygen Demand(COD) JJG 1012	(10~1000) mg/L	$U_{rel}=2.5\%$		
30	*Chemical Oxygen Demand(COD)Meters	Temperature	V.R. of Chemical Oxygen Demand(COD) Meters JJG 975	(0~200) °C	$U=0.6$ °C		
		Time		(1~120) min	$U_{rel}=0.2\%$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Concentration		(10~2000) mg/L	$U_{rel}=2.1\%$		
31	*ON-LINE Automatic Analyzer of Total Organic Carbon	Concentration	V.R. of ON-LINE Automatic Analyzer of Total Organic Carbon JJG(Zhe) 132	(0.01~20000) mg/L	$U_{rel}=2.2\%$		
32	*Total Organic Carbon Analyzer	Concentration	V.R. of Total Organic Carbon Analyzer JJG 821	Inorganic carbon:(0.001~1000) mg/L	$U_{rel}=2.2\%$		
				Organic carbon:(0.001~1000)mg/L	$U_{rel}=2.2\%$		
33	*Ammonia-nitrogen automatic analyzers	Concentration	V. R. of Ammonia nitrogen automatic monitor JJG 631	(0.1~100) mg/L	$U_{rel}=3\%$		
		Time		(1~360) s	$U=2.0\text{ s}$		
		Voltage		(0.1~500) MΩ	$U_{rel}=16\%$		
34	*Water quality on-line analyzers of total phosphorus and total nitrogen	Concentration	V. R. of On-line water quality analyzer for total phosphorus and total nitrogen JJG 1094	Total phosphorus:(0.1~100) mg/L	$U=0.01\text{ mg/L}$		
				Total phosphorus:(0.5~100) mg/L	$U_{rel}=3\%$		
				Total nitrogen:(0.1~2) mg/L	$U=0.06\text{ mg/L}$		
				Total nitrogen:(2~100) mg/L	$U_{rel}=3\%$		
35	*Water Quality On-line Analyzers of Heavy Metals	Concentration	C. S. for On line analyzer for heavy metal water quality JJF 1565	Pb:(0.001~1000) mg/L	$U_{rel}=2.8\%$		
				Cd:(0.0001~1000) mg/L	$U_{rel}=2.8\%$		
				Hg:(0.0001~1000) mg/L	$U_{rel}=2.8\%$		
				As:(0.001~1000) mg/L	$U_{rel}=2.8\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Cr:(0.01~1000) mg/L	$U_{rel}=2.8\%$		
				Cr ⁶⁺ :(0.001~1000) mg/L	$U_{rel}=2.8\%$		
				Cu:(0.01~1000) mg/L	$U_{rel}=2.8\%$		
				Zn:(0.01~1000) mg/L	$U_{rel}=2.8\%$		
				Ni:(0.001~1000) mg/L	$U_{rel}=2.8\%$		
				Fe:(0.01~1000) mg/L	$U_{rel}=2.8\%$		
				Mn:(0.01~1000) mg/L	$U_{rel}=2.8\%$		
36	*Residual Chlorine Meters	Concentration	C.S. for Residual Chlorine Meters JJF 1609	Free residual chlorine: (0.1~50) mg/L	$U_{rel}=1.6\%$		
				Total residual chlorine: (0.1~500) mg/L	$U_{rel}=0.9\%$		
37	Wood moisture content measuring instrument	Content	V.R. of moisture content measuring instrument JJG 986	Content of water: 6%~28%	$U=0.3\%$		
38	*Instrument for KF Coulometry Titration	Content	V.R. of Carle Fischer Coulomb method micro moisture meter JJG 1044	Moisture content: (10~5000) μg	$U_{rel}=(1.6\sim0.8)\%$		
		Time		Electrolytic rate:(1~180) s	$U=4\text{ s}$		
				Zero balance:(1~120) s	$U=4\text{ s}$		
39	*Karl Fischer Volumetric Titrators for Water Content	Content	Karl V.R. of Fischer Volumetric Titrators for Water Content JJG 1154	Moisture content:0.1%~5.0%	$U_{rel}=1.6\%$		
40	*Liquid-borne Particle Counters	Count	V.R for Liquid-borne Particle Counters JJG 1061	(1~3000) Counter/mL	$U_{rel}=5\%$		



在线扫码获取验证

No. CNAS L2336

第 132 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Particle size		(1~50) μm	$U_{\text{rel}}=5\%$		
41	*Static Light Scattering Particle Size Analyzers	Particle size	C.S. for Static Light Scattering Particle Size Analyzers JJF 1211	(1~5) μm	$U_{\text{rel}}=5\%$		
				(5~20) μm	$U_{\text{rel}}=5\%$		
				(20~100) μm	$U_{\text{rel}}=4\%$		
				(100~200) μm	$U_{\text{rel}}=3\%$		
42	Flow Cups Viscosimeter	Time	V.R. of Flow Cups Viscosimeter JJG 743	Time of outflow:(15~150) s	$U_{\text{rel}}= (0.5\% \sim 0.8\%)$		
				Correction coefficient:0.95~1.05	$U_{\text{rel}}=2.9\%$		
43	Rotandarl Viscometer	Viscosity	V.R. of Rotational Visometers JJG 1002	($2 \sim 1 \times 10^5$) $\text{mPa}\cdot\text{s}$	$U_{\text{rel}}= (0.6 \sim 0.8)\%$		
44	*Laboratory pH Meters	pH	V.R. of Laboratory pH Meters JJG 119	Meter: 0~14	$U=0.001$		
				Instrument: 4~9.5	$U=0.013$		
		Potential		(-2000~2000) mV	$U=0.2\text{ mV}$		
		Temperature		(0~50) $^{\circ}\text{C}$	$U=0.2\text{ }^{\circ}\text{C}$		
45	*On-line pH Meters	pH	C.S. for On-line pH Meters JJF 1547	Meter: 0~14	$U=0.01$		
				Instrument:1.6~13.5	$U=0.02$		
		Potential		(-2000~2000) mV	$U=0.2\text{ mV}$		
		Temperature		(0~50) $^{\circ}\text{C}$	$U=0.2^{\circ}\text{C}$		
46	*Laboratory Ion Meters	Potential	V.R. Laboratory Ion meters JJG 757	(-2000~2000) mV	$U=0.2\text{ mV}$		
		pX		Meter: (0~14) pX	$U=0.001$		




No. CNAS L2336

第 133 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Instrument: pX: 2~4	U=0.02		
		Temperature		(0~50) °C	U=0.2 °C		
47	*Automatic Potentionmetric Titrator	Potential	V.R. of Automatic Potentionmetric Titrator JIG 814	(-2000~2000) mV	U=0.2 mV		
		Capacity		(1~100) mL	U= (0.003~0.030) mL		
		Concentration		0.1 mol/L	U _{rel} =0.5 %		
48	*Electrolytic Conductivity Meters	Electrolytic Conductivity	V.R. of Electrolytic Conductivity Meters JIG 376	Meter:(0.05~1×10 ⁵) μ S/cm	U=0.1%FS		
				Instrument: (100~2000) μ S/cm	U _{rel} =0.4%		
				Instrument: (10~120) mS/cm	U _{rel} =0.2%		
		Temperature		(0~50) °C	U=0.2°C		
49	*Micro oxygen analyzer	Concentration	V. R. of Micro Oxygen analyzer JIG 945	(20~1000) μ mol/mol	U _{rel} =2 %		
50	*Zirconia oxygen analyzer	Concentration	V. R. of Zirconia Oxygen analyzer JIG 535	(0.1~25)×10 ⁻² mol/mol	U _{rel} =1.5 %		
51	*Electrochemical oxygen analyzer	Concentration	V. R. of Electrochemical Oxygen analyzer JIG 365	(5.0~25)×10 ⁻² mol/mol	U _{rel} =1.5 %		
52	*Photoionization detector for volatile organic compounds	Concentration	C. S. for Photoionization detector for volatile Organic compounds JJF 1172	(0.1~2000) μ mol/mol	U _{rel} =2.3%		
53	*Carbon monoxide detection alarm	Concentration	V. R. of Carbon monoxide detection alarm JIG 915	(1~2000) μ mol/mol	U _{rel} =2.3 %		
54	*Combustible gas detection alarm	Concentration	V. R. of Combustible gas detection alarm JIG 693	CH ₄ :(10~60)%LEL	U _{rel} =2.1 %	Only carry out the calibration of	
				H ₂ :(10~60)%LEL	U _{rel} =2.1 %		



No. CNAS L2336

第 134 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				i-C ₄ H ₁₀ :(10~60)%LEL	U _{rel} =2.1 %	methane, isobutane, hydrogen, propane, and acetylene gas detectors.	
				C ₃ H ₈ :(10~60)%LEL	U _{rel} =2.1 %		
				C ₂ H ₂ :(10~60)%LEL	U _{rel} =2.1 %		
55	*Ammonia detector	Concentration	V. R. of Ammonia detector JIG 1105	(1~80) μ mol/mol	U _{rel} =2.0 %		
56	*Hydrogen sulfide gas detector	Concentration	V. R. of Hydrogen sulfide gas detector JIG695	(1~80) μ mol/mol	U _{rel} =2.2 %		
57	*Dust sampler	Flow	V. R. of Dust sampler JIG 520	(0.01~80) L/min	U _{rel} =1.2 %		
		Time		(1~300) s	U=0.3 s		
58	*samplers for stack dust	Flow	V. R. of samplers for stack dust JIG 680	(5~100) L/min	U _{rel} =1.5 %		
		Temperature		Temperature before flowmeter: (0~50) °C	U=0.4 °C		
				Flue gas temperature: (1~300) °C	U=1.3 °C		
				Pressure	Dynamic pressure: (0~2000) Pa		
				Static pressure: (-30~30) kPa	U=0.2 %FS		
Time	(1~600) s	U=0.2 s					
59	*air sampler	Flow	V. R. of air sampler JIG 956	(0.1~6) L/min	U _{rel} =1.2 %		
		time		(1~3600) s	U _{rel} =0.06 %		
		Temperature		(0~50) °C	U=0.2 °C		
60	*Tsp sampler	Flow	V. R. of Tsp sampler JIG 943	(80~150) L/min	U _{rel} =1.2 %		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		time		(1~1200) s	$U=0.3s$		
		Temperature		(0~50) °C	$U=0.2^{\circ}C$		
		Pressure		(800~1060) hPa	$U=2.4\text{ hPa}$		
		Length		(1~200) mm	$U_{rel}=0.3\%$		
61	*Planktonic bacteria sampler	Flow	C. S. for Planktonic bacteria sampler JJF(Su)188	(20~200) L/min	$U_{rel}=1.2\%$		
		Time		(1~600) s	$U=0.3\text{ s}$		
62	Dust Particles Counter	Flow	C.S. of Dust Particles Counter JJF 1190	(0.3~150) L/min	$U_{rel}=1.3\%$		
		Time		(0~600) s	$U=0.30\text{ s}$		
		Concentration		Particle size distribution: (0.3~10) μm	$U_{rel}=11\%$		
				Particle concentration: (10~100000) $\uparrow/2.83\text{ L}$	$U_{rel}=17\%$		
63	*Particulate Analyzer	Volume	C.S. for Particulate Analyzer JJF 1290	(0.01~10) mL	$U_{rel}=0.4\%$		
		Count		(10~100000) counter/mL	$U_{rel}=5\%$		
64	*flue gas samplers	flow	V. R. of flue gas samplers JJG 1169	(0.1~2) L/min	$U_{rel}=1.3\%$		
		Time		(1~1200) s	$U=0.3\text{ s}$		
		Temperature		(0~50) °C	$U=0.2\text{ }^{\circ}C$		
65	*Fourier Transform Infrared Spectrometers	Wave numbe	C.S. for Fourier Transform Infrared Spectrometers JJF 1319	(3100~800) cm^{-1}	$U=(0.09\sim0.13)\text{ cm}^{-1}$		
66	*Atomic Fluorescence	Concentration	V.R. of Atomic Fluoreacence Spectrometer JJG 939	Detection limit(As): $\leq 0.4\text{ ng}$	$U=0.015\text{ ng}$		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Spectrophotometer s			Detection limit(Sb): ≤ 0.4 ng	$U=0.04$ ng		
67	*Osmotic pressure molar concentration meter	Concentration	V.R. of Osmotic pressure molar concentration meter JJG1089	Osmotic pressure molar concentration:(100~700) mOsmol \cdot kg ⁻¹	$U=(1.7\sim 3.4)$ mOsmol \cdot kg ⁻¹		
68	*Fixed carbon sulfur analyzer	Content	V.R. of Carbon Sulfur Analyzer JJG 395	C:(0.0050~0.0100) %	$U=0.0005$ %		
				C:(0.010~0.100) %	$U=0.004$ %		
				C:(0.100~0.500) %	$U=0.007$ %		
				C:(0.500~1.000) %	$U=0.024$ %		
				C:(1.00~4.00) %	$U=0.04$ %		
				S:(0.0030~0.0100) %	$U=0.0003$ %		
				S:(0.010~0.050) %	$U=0.0013$ %		
				S:(0.050~0.100) %	$U=0.003$ %		
		S:(0.100~0.200) %	$U=0.005$ %				
		Time		(1~1200) s	$U=0.3$ s		
69	*Elemental Analyzers	Content	C.S. for Element analyzer JJF1321	Kjeldahl nitrogen analyzer: 0.1%~50%	$U_{rel}=1.4$ %		
70	*Quadrupole Inductively Coupled Plasma Mass Spectrometers	Detection limit	C.S. for Quadrupole Inductively Coupled Plasma Mass Spectrometers JJF 1159	Detection Limit(Be): ≤ 30 ng/L	$U=0.4$ ng/L		
				Detection Limit(In): ≤ 10 ng/L	$U=0.2$ ng/L		
				Detection Limit(Bi): ≤ 10 ng/L	$U=0.2$ ng/L		




No. CNAS L2336

第 137 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Sensitivity(Be): ≥ 5 Mcps/(mg/L)	$U_{rel}=10\%$		
				Sensitivity(In): ≥ 30 Mcps/(mg/L)	$U_{rel}=14\%$		
				Sensitivity(Bi): ≥ 20 Mcps/(mg/L)	$U_{rel}=12\%$		
71	*Dynamic Light Scattering Particle Size Analyzers	Particle size	V.R. of Dynamic Light Scattering Particle Size Analyzers JJG 1104	(10~1000) nm	$U_{rel}=2.6\%$		
		Temperature		(0~90) °C	$U=0.2\text{ °C}$		
72	*Micro-spectrophotometer s	Concentration	C.S. for Micro-spectrophotometers JJF 1836	Protein: (0.1~10.0)mg/mL	$U_{rel}=12\%$		
				DNA: (0.1~2500) ng/ μ L	$U_{rel}=7\%$		
73	*Colony Counters	Total number of colonies	C.S. for Colony Counters JJF 1751	(30~300) CFU	$U_{rel}=2.6\%$		
		Length		(0.10~5.10) mm	$U=0.04\text{ mm}$		
		Color temperature		(2300~9000)K	$U=(22\sim 1.9\times 10^2)\text{K}$		
74	*McFarland Bacterial Turbidity Analyzers	Bacteria turbidity	C.S. for McFarland Bacterial Turbidity Analyzers JJF 1825	(0~4) MCF	$U=(0.06\sim 0.23)\text{ MCF}$		
75	*Bacterial Endotoxin Analyzers	Temperature	C.S. for Bacterial Endotoxin Analyzers JJF 1529	37°C	$U=0.14\text{ °C}$		
76	*Oxidation-reduct Potential Meters	Potential	C.S. for Oxidation-reduct Potential Meters JJF(Liao) 474	电计: (-2000~2000) mV	$U=0.2\text{ mV}$		
				仪器: (-2000~2000) mV	$U=0.4\text{ mV}$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Resistance		Impedance: (3~10) G Ω	$U_{rel}=12\%$		
77	*Flue Gas Analyzers	Concentration	V.R. of Flue Gas Analyzers JIG 968	O ₂ : (0.1~21)×10 ⁻² mol/mol	$U_{rel}=1.0\%$		
				CO: (1~2000)×10 ⁻⁶ mol/mol	$U_{rel}=1.3\%$		
				CO ₂ : (0.1~20)×10 ⁻² mol/mol	$U_{rel}=1.4\%$		
				SO ₂ : (1~100)×10 ⁻⁶ mol/mol	$U_{rel}=1.4\%$		
				NO: (10~5000)×10 ⁻⁶ mol/mol	$U_{rel}=1.4\%$		
78	*Sulfur Dioxide Gas Detectors	Concentration	V.R. of Sulfur Dioxide Gas Detectors JIG 551	(1~80)×10 ⁻⁶ mol/mol	$U_{rel}=2.4\%$		
79	*Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer	Concentration	V.R. of Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer JIG 635	CO: (1~2000)×10 ⁻⁶ mol/mol	$U_{rel}=1.4\%$		
				CO ₂ : (0.1~20)×10 ⁻² mol/mol	$U_{rel}=1.4\%$		
80	*Nitric Oxide And Nitrogen Dioxide Gas Detectors	Concentration	V.R. of Nitric Oxide And Nitrogen Dioxide Gas Detectors JIG(Xin) 01	NO: (1~2000) μmol/mol	$U_{rel}=1.4\%$		
				NO ₂ : (1~500) μmol/mol	$U_{rel}=2.3\%$		
81	*Alarmer Detectors of Benzene	Concentration	C.S. for Alarmer Detectors of Benzene JJF 1674	(1~80) μmol/mol	$U_{rel}=2.3\%$		
82	*Hydrogen Chloride Gas Detectors and Alarms	Concentration	C.S. for Hydrogen Chloride Gas Detectors and Alarms JJF 1888	(1~100) μmol/mol	$U_{rel}=2.3\%$		




No. CNAS L2336

第 139 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
83	*Chlorine Alarm Detectors	Concentration	C.S. for Chlorine Alarm Detectors JJF 1433	(1~20) μ mol/mol	$U_{rel}=2.5\%$		
84	*The Alarmer Detector of Sulfur Hexafluoride	Concentration	C.S. for the Alarmer Detector of Sulfur Hexafluoride JJF 1263	(1~800) μ mol/mol	$U_{rel}=2.3\%$		
85	*Raman Spectrometers	Frequency shift	C.S. for Raman Spectrometers JJF 1544	(300~40000) cm^{-1}	$U=0.3 \text{ cm}^{-1}$		
		Resolution of spectrum		(0~20) cm^{-1}	$U=0.1 \text{ cm}^{-1}$		
		relative intensity		0~1	$U=0.01\sim 0.07$		
86	*Gas Detectors of Ethylene Oxide	Concentration	C.S. for Gas Detectors of Ethylene Oxide JJF2288	(20~80) μ mol/mol	$U_{rel}=2.5\%$		
87	*Methanol Gas Detection Alarm	Concentration	C.S. for Methanol Gas Detection Alarm JJF(Hei)11	(1~300) μ mol/mol	$U_{rel}=2.3\%$		
88	*Ethanol Gas Detectors and Alarms	Concentration	C.S. for Ethanol Gas Detectors and Alarms JJF(SH)081	(1~400) μ mol/mol	$U_{rel}=1.5\%$		
89	*Acetone Gas Detectors and Alarms	Concentration	C.S. for Acetone Gas Detectors and Alarms JJF(SH)082	(1~500) μ mol/mol	$U_{rel}=2.7\%$		
90	*Gas Detectors and Alarms of Hydrogen Fluoride	Concentration	C.S. for Gas Detectors and Alarms of Hydrogen Fluoride JJF(SH)047	(3~8) μ mol/mol	$U_{rel}=3.8\%$		
91	*Formaldehyde gas detection alarm	Concentration	C.S. for formaldehyde gas detection alarm JJF(SH)100	(2~8) μ mol/mol	$U_{rel}=4.2\%$		
92	Sulfur Hexafluoride Decomposition Products Detectors	Concentration	C.S. for Sulfur Hexafluoride Decomposition Products Detectors JJF 1711	SO ₂ (1~50) μ mol/mol	$U_{rel}=3.7\%$		
		Concentration		H ₂ S(1~50) μ mol/mol	$U_{rel}=3.9\%$		
		Concentration		CO(1~500) μ mol/mol	$U_{rel}=3.2\%$		



No. CNAS L2336

第 140 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
93	*Electrochemical Workstations	Potential	C.S. for Electrochemical Workstations JJF 1910	(0~2)V	$U=0.0001V$		
		Current		(0~2) μA	$U=2 \times 10^{-10}A$		
94	*Real-Time Fluorescent Quantitative Polymerase Chain Reaction Analyzer	Temperature	C.S. for Real-Time Fluorescent Quantitative Polymerase Chain Reaction Analyzer JJF(Su) 222	(30~95) $^{\circ}C$	$U=0.2^{\circ}C$		
		threshold cycle (crossing point)		10~40	$U=0.7$		
		Concentration		(1~10 ⁷)copies/ μL	$U_{rel}=7\%$		
95	*Adenosine Triphosphate (ATP) Fluorescence Detectors	Fluorescence	C.S. for Adenosine Triphosphate (ATP) Fluorescence Detectors JJF 1828	(1 $\times 10^{-16}$ ~5 $\times 10^{-10}$)mol/mL	$U_{rel}=3.6\%$		
		RLU		(1~1000000) RLU	$U_{rel}=17\%$		
96	*Total Organic Carbon Analyzer based on Conductivity Measurement	Contamination	C.S of Total Organic Carbon Analyzer based on Conductivity Measurement JJF(Jing)112	(0.1~1.0) mg/L	$U_{rel}=1.3\%$		
97	*Surfaceion contamination tester	Surfaceion contamination	C.S of Surfaceion contamination tester JJF (DZ) 0070	(0.1~20) $\mu g/cm^2$	$U_{rel}=4.8\%$		
		Electric conductivity		(50~150) $\mu S/cm$	$U_{rel}=1.3\%$		
		Temperature		(5~50) $^{\circ}C$	$U=0.1^{\circ}C$		
98	*Water Activity Meters	Water Activity	V.R of Water Activity Meters JJG(Yue) 052	0.0~1.0	$U=0.008$		
99	*Bioreactor	Temperature	C.S. for Bioreactor JJF(Xiang) 97	(0~50) $^{\circ}C$	$U=0.2^{\circ}C$		
		pH		0~14	$U=0.02$		




在线扫码获取验证

No. CNAS L2336

第 141 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rotational speed		(10~2000) rpm	$U_{rel}=0.6\%$		
		Gas Flow		(0.3~30) L/min	$U_{rel}=0.8\%$		
		Fluid Flow		(0.1~100) mL/min	$U_{rel}=0.8\%$		
		Concentration		Dissolved oxygen: (6~15) mg/L	$U=0.1$ mg/L		
100	Routine Capillary Viscometers	Kinematic viscosity	V.R. of Routine Capillary Viscometers JJG 155	(1~100000)mm ² /s	$U_{rel}=(0.34\sim0.86)\%$		
101	Zahn Viscosity Cups	Time	C.S. for Zahn Viscosity Cups JJF (BGMP) 0009	(15~150)s	$U_{rel}=(0.5\%\sim0.8\%)$		
102	DIN 53211-4 Coating Viscometers	Time	C.S. for DIN 53211-4 Coating Viscometers JJF (Ji) 114	(15~150)s	$U_{rel}=(0.5\%\sim0.8\%)$		
103	*Kinematic Viscosity Tester	Kinematic viscosity	C.S. for Kinematic Viscosity Tester JJF 1274	(1~20000)mm ² /s	$U_{rel}=2\%$	Only test the viscosity at 20°C	
		Temperature		(20~100) °C	$U=0.009^{\circ}\text{C}$		
104	*Gold Gauge Utilizing X-ray Fluorescence Spectrometry	Concentration	C.S. of Gold Gauge Utilizing X-ray Fluorescence Spectrometry JJF 1133	30%~100%	$U=0.02\%\sim0.04\%$		
105	*Rapid Detection Analyzers of Heavy Metal in Food	Concent	C.S. for Rapid Detection Analyzers of Heavy Metal in Food JJF(Jing)113	(0.001~0.2)mg/kg	$U=0.014$ mg/kg		
				(0.2~0.4)mg/kg	$U_{rel}=7\%$		
				(0.4~2.2)mg/kg	$U_{rel}=3.5\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
106	*Portable Stripping Voltammetry Heavy Metal Detectors	Concentration	C.S. for Portable Stripping Voltammetry Heavy Metal Detectors JJF2037	(0.01~1000)mg/L	$U_{rel}=2\%$		
107	*Gel Documentation Systems	Illuminance	C.S. for Gel Documentation Systems JJF1530	50 lx~50 klx	$U_{rel}=5\%$		
		Irradiance		(0.01~2) mW/cm ²	$U_{rel}=24\%$		
108	*Digital Refractometers	Content	C.S. for Digital Refractometers JJF(Jin)83	NaCl: 0%~28%	$U=0.06\%$		
				糖量: 0%~80%	$U=0.02\%$		
		Refractivity		1.30~1.66	$U=0.0002$		
109	*Total Dissolved Solids(TDS) Meters	Concentration	C.S. for Total Dissolved Solids(TDS) Meters JJF(Min)1097	(1~20000)mg/L	$U_{rel}=0.2\%~0.6\%$		
		Temperature		(0~50)°C	$U=0.2^{\circ}\text{C}$		
110	*Online Conductivity Meters for Water Quality	Conductivity	C.S. of Online Conductivity Meters for Water Quality JJF(Jin)74	Electronic unit:(0.05~1 × 10 ⁵) μ S/cm	$U=0.2\% \text{FS}$		
				Instrument:(0.5~20) μ S/cm	$U=1.2\% \text{FS}$		
				Instrument:(20~2 × 10 ⁵) μ S/cm	$U=0.3\% \text{FS}$		
		Temperature		(0~50) °C	$U=0.3^{\circ}\text{C}$		
111	*Digital Saccharimeters and Digital Refractometers	Sugar Concent	V.R. of Digital Saccharimeters and Digital Refractometers JJG(Jin)68	1%~80%	$U=0.02\%$		
		Refractivity		1.30~1.66	$U=0.00006$		
		Temperature		(10~30) °C	$U=0.13^{\circ}\text{C}$		
112	*Scanning Electronic	Length	C.S.for Scanning Electronic Microscopes(SEM) JJF 1916	100nm	$U_{rel}=1.5\%$		

No. CNAS L2336


第 143 页 共 187 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Microscopes(SEM)			200nm (500~10000) nm	$U_{rel}=1.2\%$ $U_{rel}=0.6\%$		
113	*Polymerase Chain Reaction Analyzers	Temperature Concentration	C.S. for Polymerase Chain Reaction Analyzers JJF 1527	(10~100) °C (1~10 ⁷)copies/ μ L	$U=0.20^{\circ}\text{C}$ $U_{rel}=7\%$		
114	*Disintegration Analyzers	Length Time Time limit for disintegration	Calibration Specification for Disintegration Analyzers JJF 1449	(0~100) mm 0.1s~30min 1s~30min	$U=0.2\text{mm}$ $U_{rel}=0.2\%$ $U_{rel}= (8.2\%~11\%)$		
115	*Drug Dissolution Testers	Temperature Speed Time	Calibration Specification For Drug Dissolution Testers JJF(Su) 220	(0~50) °C (0~250) r/min (0~60) min	$U=0.2^{\circ}\text{C}$ $U=0.6\text{r/min}$ $U=0.4\text{s}$		
116	*Electrospray Detector of Liquid Chromatograph	Flow minimum detectable concentration Temperature	Calibration Specification for Electrospray Detector of Liquid Chromatograph JJF(Shan) 129	(0.1~2)mL/min $\leq 5 \times 10^{-6}\text{g/mL}$ (0~50)°C	$U_{rel}=1.0\%$ $U_{rel}=9\%$ $U=0.3^{\circ}\text{C}$		
117	*Gel Permeation Chromatographs	Molecular weight	Verification Regulation of Gel Permeation Chromatographs JJG 342	organic mobile phase: (4 × 10 ³ ~6.5 × 10 ⁵)g/mol Water flow phase: (3 × 10 ³ ~6.2 × 10 ⁵)g/mol	$U_{rel}=5\%$ $U_{rel}=8\%$		
118	*Automatic Amino Acid Analyzer	Flow Detection limit	Verification Regulation of Automatic Amino Acid Analyzer JJG 1064	(0.1~2)mL/min $\leq 1\text{nmol}$	$U_{rel}=1.0\%$ $U_{rel}=8\%$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
119	*Fourier Transform Mass Spectrometers	mass charge ratio	Calibration Specification for Fourier Transform Mass Spectrometers JJF 1531	(100~2000)u	$U_{rel}=0.00003\%$		
120	*Time-of-Flight Mass Spectrometers	mass charge ratio	Calibration Specification for Time-of-Flight Mass Spectrometers JJF 1528	(100~4200)u	$U_{rel}=0.003\%$		
121	*Organic High Resolution Magnetic Sector Mass Spectrometers	Single to noise	Calibration Specification for Organic High Resolution Magnetic Sector Mass Spectrometers JJF 1930	≥ 1000	$U_{rel}=6\%$		
		mass charge ratio		(49~1020)u	$U_{rel}=0.004\%$		
122	*Superconducting Pulsed Fourier Transform Nuclear Magnetic Resonance Spectrometers	Sensitivity	Calibration Specification for Superconducting Pulsed Fourier Transform Nuclear Magnetic Resonance Spectrometers JJF 1448	$^1H \geq 15$	$U_{rel}=5\%$		
				$^{13}C \geq 2$	$U_{rel}=5\%$		
123	*Liquid Chromatography-Inductively Coupled Plasma Mass Spectrometers	Flow	Calibration Specification for Liquid Chromatography-Inductively Coupled Plasma Mass Spectrometers JJF 2115	(0.1~2)mL/min	$U_{rel}=1.0\%$		
		Minimum detectable concentration		$\leq 0.5ng/g$	$U_{rel}=8\%$		
124	*On-line Turbidimeters	Concentration	Calibration Specification of On-line Turbidimeters JJF(Meng) 120	(0.1~5)NTU	$U_{rel}=5\%$		
				(5~4000)NTU	$U_{rel}=4\%$		
125	*Water Hardness Meters	Concentration	Calibration Specification for Water Hardness Meters JJF 1949	(0.1~2000)mg/L	$U_{rel}=4\%$		
		Temperature		(10~50) $^{\circ}C$	$U=0.2^{\circ}C$		
126	*Water Quality On-line Analyzers of Hardness	Concentration	Calibration Specification of Water Quality On-line Analyzers of Hardness JJF(Jing) 5006	(0.1~100)mg/L	$U_{rel}=5\%$		




No. CNAS L2336

第 145 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
127	*Permanganate Index Analyzers	Concentration	Calibration Specification for Permanganate Index Analyzers JJF 2173	(0.1~50)mg/L	$U_{rel}=5\%$		
		Temperature		(0~100)°C	$U=0.3^{\circ}\text{C}$		
		Time		(0~30)min	$U=0.4\text{s}$		
128	*Ozone Gas Analyzers	Concentration	Verification Regulation of Ozone Gas Analyzers JJG 1077	(0.1~1) μ mol/mol	$U_{rel}=4\%$		
				(1~400) μ mol/mol	$U_{rel}=7\%$		
129	Aerosol Diluters	Dilution ratio	Calibration Specification for Aerosol Diluters JJF 2066	1~300	$U_{rel}=4\%$		
		Flow		(1~30)L/min	$U_{rel}=2\%$		
130	*Digital Polymerase Chain Reaction Analyzers	Concentration	Calibration Specification for Digital Polymerase Chain Reaction Analyzers JJF 2055	($4 \times 10^2 \sim 2 \times 10^4$)copies/μ L	$U_{rel}=10\%$		
		Temperature		(10~100)°C	$U=0.2^{\circ}\text{C}$		
131	*Protein Purification Analyzers	Flow	Calibration Specification for Protein Purification Analyzers JJF 2204	(1~500)mL/min	$U_{rel}=1.0\%$		
		acidity		pH: (4~10)	$U=0.02$		
		Electrolytic Conductivity		Instrument: (100~2000) μ S/cm	$U_{rel}=0.6\%$		
				Instrument: (10~120) mS/cm	$U_{rel}=0.4\%$		
		Wavelength		(230~318)nm	$U=0.5\text{ nm}$		
		Temperature		(0~50)°C	$U=0.4^{\circ}\text{C}$		
		Volume		(0.1~1000) μ L	$U_{rel}=0.2\%$		
132	*Plate Electrophoresis Apparatus	Current	Calibration Specification for Plate Electrophoresis Apparatus JJF 1654	(2~500)mA	$U=0.7\text{mA}$		
		Voltage		(2~500) V	$U=0.7\text{V}$		




No. CNAS L2336

第 146 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
133	*Swing Bioreactors	Flow	Calibration Specification for Performance Parameters of Swing Bioreactors JJF 2260	Gas:(0.01~20)L/min	$U_{rel}=2.0\%$		
				Liquid:(1~500)mL/min	$U_{rel}=1.0\%$		
		Frequency		(10~50)r/min	$U_{rel}=0.2\%$		
		Temperature		(0~50)°C	$U=0.4^{\circ}C$		
		Angle		(0~20)°	$U=0.02^{\circ}$		
134	*Pesticide Residue Detectors	Transmittance	Calibration Specification for Pesticide Residue Detectors JJF 1729	9%~30%	$U_{rel}=0.7\%$		
		Wavelength		(350~450)nm	$U=0.6nm$		
		Sensitivity		50%~100%	$U_{rel}=6\%$		
135	*Microplate Chemiluminescence Analyzers	Absorbance	Calibration Specification for Microplate Chemiluminescence Analyzers JJF 1849	0.1~1.5	$U=0.012$		
136	*Multifunctional blood gas analyzer	Concentration	Calibration Specification for Multifunctional blood gas analyzer JJF 2054	6~8	$U=0.02$		
				K ⁺ : (1.5~7.5)mmol/L	$U_{rel}=4\%$		
				Na ⁺ : (100~180)mmol/L	$U_{rel}=4\%$		
				Cl ⁻ : (80~160)mmol/L	$U_{rel}=4\%$		
				iCa ²⁺ : (0.50~2.50)mmol/L	$U_{rel}=5\%$		
				Hb: (5.0~18.0)g/dL	$U_{rel}=3\%$		
				p(O ₂):(0.1~200) mmHg	$U_{rel}=2\%$		
		partial pressure		p(CO ₂):(0.1~100) mmHg	$U_{rel}=2\%$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
137	*Hydrogen Peroxide Meter	Concentration	Calibration Specification for Hydrogen Peroxide Meter JJF(Zhe) 1156	(100~2000) μ mol/mol	$U_{rel}=7\%$		
138	*Dry Injection Laser Particle Size Analyzers	Particle size	C.S. for Dry Injection Laser Particle Size Analyzers JJF(Jin) 82	(35~52) μ m	$U_{rel}=(7\sim 8)\%$		
139	*Engler Viscosimeter	Time	V.R. of Engler Viscosimeter JJG 742	(50~52)s	$U=0.24s$		
140	*Acid Number Testers	Concentration	C.S. for Acid Number Testers JJF(Ji) 3030	(0.04~0.1) mg/g	$U=0.013mg/g$		
				(0.1~1.0) mg/g	$U_{rel}=6\%$		
141	*Laboratory Pure Water Machines	Resistivity	C.S. for Laboratory Pure Water Machines JJF(E) 166	(0.1~18.25)M Ω /cm	$U_{rel}=1.3\%$	accredited only for volumetric method for the outflow rate	
		Temperature		(10~90)°C	$U=0.2^{\circ}C$		
		pH		4~10	$U=0.03$		
		Flow rate		(10~50)mL/s	$U=1.4mL/s$		
142	*Carbon Dioxide Gas Detectors	Concentration	C.S. for Carbon Dioxide Gas Detectors JJF(Hei) 10	1.00%~4.00%	$U_{rel}=2.4\%$		
143	*Portable Sulphur Hexafluoride Purity Analyzers	Concentration	C.S. for Portable Sulphur Hexafluoride Purity Analyzers JJF(Yu) 253	90%~100%	$U_{rel}=0.8\%$		
144	*Rotational Rheometers	Shear Viscosity	C.S. for Rotational Rheometers JJF 2134	(0.9~900)Pa · s	$U_{rel}=8\%$		
		Temperature		(5~100)°C	$U=0.18^{\circ}C$		
145	*Rotary Evaporators	Temperature	Calibration Specification of Rotary Evaporators JJF(E) 80	(20~200) °C	$U=0.7^{\circ}C$		
		Speed		(10~280) r/min	$U=0.6r/min$		
		Pressure		(-100~0)kPa	$U=0.3kPa$		



No. CNAS L2336


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
ionizing radiation measuring instrument							
1	*X-ray Flaw Detectors	Air specific release energy rate	V.R.of X-ray Flaw Detectors JIG 40	0.1cGy/min~10Gy/min	$U_{rel}=5\%$	Accredited only for field calibration	
Acoustic measuring instrument							
1	*Ultrasonic Flaw Detectors	Amplifier frequency response	V. R. of Ultrasonic Flaw Detectors JIG 746	10mV _{pp} ~10V _{pp} (0.5~25)MHz	$U_{rel}=3\%$		
		Dynamic range		(10~40)dB	$U=1$ dB		
		Equivalent input noise		(0.1~1000) μ V	$U_{rel}=1\%$		
		Attenuator error		(0.1~96)dB	$U=0.2$ dB		
				(96~110)dB	$U=0.3$ dB		
		Amplitude linear error		0.1%~100%	$U=1.6\%$		
		Time base linear error		(0.1~10000)mm	$U_{rel}=0.15\%$		
		Launch pulse		(0.01~10)V	$U_{rel}=3\%$		
2	*Electro-acoustical Measurement Instruments	Frequency	C. S. for Electro-acoustical Measurement Instruments JJF 1339	20Hz~20kHz	$U_{rel}=0.01\%$		
		AC Voltage		(0.1~30)V,(20Hz~20kHz)	$U_{rel}=0.2\%$		
		Total distortion		0.01%~10%,(20Hz~20kHz)	$U_{rel}=0.1\%$		
		Frequency response		(0.1~30)V,(20Hz~20kHz)	$U_{rel}=0.2\%$		
Textile and leather special measuring instruments (including footwear testing instruments)							



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1	*Fabric Thickness Instruments	Length	Calibration Specification for Fabric Thickness Instruments JJF(FZ) 020	(0.5~20)mm	$U=3 \mu m$		
		Area		(100~10000)mm ²	$U_{rel}=0.2\%$		
		Force value		(10~200)cN	$U_{rel}=0.3\%$		
		Time		(5~35)s	$U=0.3s$		
2	*Circle Sample Cutters	Area	Calibration Specification for Circle Sample Cutters JJF(FZ) 061	(90~110)cm ²	$U_{rel}=0.3\%$		
3	*Vertical Combustion Tester	Mass	Calibration Specification for Vertical Combustion Tester JJF(FZ) 068	54.5 g、113.4 g、226.8 g、340.2 g、453.6 g	$U=0.4 g$		
		Time		(10~3600)s	$U=0.2 s$		
		Length		11.0 mm、12.5 mm、17.0 mm、40.0 mm	$U=0.04 mm$		
				30 mm、51 mm、329 mm、356 mm、767 mm	$U=0.3 mm$		
		Angle		25°	$U=0.5^\circ$		
		Velocity		45 mm/s	$U=1.2 mm/s$		
4	*Sliver and Roving Length Sampler	Length	Calibration Specification for Sliver and Roving Length Sampler JJF(FZ) 001	500 mm、1000 mm	$U=0.3 mm$		
		Mass		(2000~4500)g	$U=5 g$		
5	*Fabrics Water Spray Testers	Length	Calibration Specification for Fabrics Water Spray Testers JJF(FZ) 083	0.86 mm	$U=0.02 mm$		
				10 mm、21 mm	$U=0.05 mm$		
				150 mm、155 mm、195 mm	$U=1.0 mm$		
		Angle		45°	$U=0.3^\circ$		



No. CNAS L2336

第 150 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time		(25~30)s	$U=0.6$ s		
6	*Light and Weather Fastness Tester	Temperature	Calibration Specification for Light and Weather Fastness Tester JJF(FZ) 051	Test chamber: (20~63)°C	$U=0.8$ °C		
				BPT, BST: (20~63)°C	$U=1.6$ °C		
		Humidity		20%RH~95%RH	$U=1.7\%$ RH		
		Rotate speed		(10~100)r/min	$U=0.7$ r/min		
		Time		1 min、30 min	$U=0.6$ s		
		Irradiance		1.1 W/m ² 、42 W/m ²	$U_{rel}=23\%$		
7	*Ventilated Textile Oven	Temperature	Calibration Specification for Ventilated Textile Oven JJF(FZ) 059	(50~150)°C	$U=0.8$ °C		
		Velocity		(0.8~10)m/s	$U=0.05$ m/s		
		Time		(1~20)min	$U=0.5$ s		
8	*Eight-Basket Oven	Temperature	Calibration Specification for Eight-Basket Oven JJF(FZ) 011	(50~150)°C	$U=0.8$ °C		
		Time		(1~25)min	$U=0.5$ s		
9	*Colour Fastness to Washing Testers	Temperature	Calibration Specification for Colour Fastness to Washing Testers JJF(FZ) 026	(20~98)°C	$U=0.5$ °C		
		Time		(1~24)min、30 min、45 min	$U=0.5$ s		
		Rotating velocity		40 r/min	$U=0.4$ r/min		
		Length		45 mm、75 mm、90 mm、125 mm、200 mm	$U=0.04$ mm		
10	*Fabric Shrinkage Testers	Temperature	Calibration Specification for Fabric Shrinkage Testers JJF(FZ) 052	40 °C、60 °C、92 °C	$U=1.4$ °C		
		Time		3 s~22 min	$U=0.5$ s		



在线扫码获取验证

No. CNAS L2336

第 151 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rotating velocity		52 r/min、500 r/min	$U_{rel}=1.5\%$		
		Length		100 mm、130 mm	$U=0.3$ mm		
11	*Color Fastness to Seorch and Sublimation Testers	Temperature	Calibration Specification for Color Fastness to Seorch and Sublimation Testers JJF(FZ) 029	110 °C、150 °C、180 °C、200 °C、210 °C	$U=1.4$ °C		
		Force		16 N	$U=0.5$ N		
		Time		15 s、30 s	$U=0.5$ s		
12	*Perspiration Fastness Instruments	Force	Calibration Specification for Perspiration Fastness Instruments JJF(FZ) 028	(49~51)N	$U=0.1$ N		
		Length		(2~120) mm	$U=0.5$ mm		
13	*Corlour Fastness to Friction Testers	Force	Calibration Specification for Corlour Fastness to Friction Testers JJF(FZ) 027	9 N	$U=0.08$ N		
		Length		16 mm、19 mm、25.4 mm、104 mm	$U=0.5$ mm		
		Rotating velocity		60 r/min	$U=0.7$ r/min		
14	*Geotextiles Abrasion Testers	Length	Calibration Specification for Geotextiles Abrasion Testers JJF(FZ) 095	25.4 mm、50.0 mm、200.0 mm	$U=0.05$ mm		
		Frequency		90 min ⁻¹	$U=0.5$ min ⁻¹		
		Mass		6000 g	$U=0.8$ g		
15	*Thickness of Geosynthetics Determinator	Length	Calibration Specification for the Thickness of Geosynthetics Determinator JJF(Min) 1125	0 mm、0.5 mm、1 mm、2 mm、3 mm、4 mm、5 mm	$U=0.01$ mm		
		Area		25 cm ²	$U=0.04$ cm ²		
		Pressure		2 kPa、20 kPa、200 kPa	$U_{rel}=0.4\%$		




No. CNAS L2336

第 152 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
16	*Constant Temperature Water Bath Oscillator	Temperature	Calibration Specification for Constant Temperature Water Bath Oscillator JJF(FZ) 105	(20~100)°C	U=0.5°C		
		Rate		(30~120)times/min	U=0.8times/min		
		Amplitude		(20~50)mm	U=0.05mm		
		Time		(1~60)min	U=1s		
Special measuring instruments for motor vehicles							
1	*Tire Tread Depth Gauges	Length	C.S. for Tire Tread Depth Gauges JJF 1477	(0~50)mm	U=0.02mm		
2	*Transmittance Meter of Automobile	Transmittance	C.S. for Transmittance Meter of Automobile JJF 1225	(0~100) %	U=0.7%		
Special measuring instruments for Meteorology and ocean							
1	Thermo-anemoscopes	Wind speed	C. S. for Thermo-anemoscopes JJF1939	(5~25)m/s	U= (0.09~0.20) m/s		
2	Fanning Mill Anemometers	Wind speed	C. S. for Fanning Mill Anemometers JJF1971	(5~25)m/s	U= (0.09~0.27) m/s		
3	Wheel Type Digital Anemometers	Wind speed	V.R.of Wheel Type Digital Anemometers JJG 1194	(5~25)m/s	U= (0.09~0.27) m/s		
4	Wind Speed Transducer	Wind speed	C. S. for Wind Speed Transducer JJF(Zhe)1126	(5~25)m/s	U= (0.10~0.30) m/s		
5	Bucket Thermometers	Temperature	V. R. of Bucket Thermometers JJG289	(-5~40)°C	U=0.10°C		
6	Water Level Gauge	Length	V.R. of Water Level Gauge JJG(JT) 168	(0~100) mm	U=0.3mm+2×10 ⁻⁵ L		
Special measuring instrument for paper and paper							



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1	*Compression Strength Tester for Board	Force	V. R.of Compression Strength Tester for Board JJG(QG) 49	(1~3000)N	$U_{rel}=0.4\%$		
2	*Compression Strength Tester for Corrugated Box	Force	V. R.of Compression Strength Tester for Corrugated Box JJG(QG) 115	(0.1~50)kN	$U_{rel}=0.4\%$		
		Length		(0~500) mm	$U=0.4\text{mm}$		
3	*Bursting Strength Tester for Paper and Board	Pressure	C. S. for Bursting Strength Testers for Paper(Board) JJF 1811	(0.001~6)MPa	$U=0.2\%FS$		
Medical special measuring instrument							
1	*Baby Incubator	Temperature	C.S.for Baby Incubator JJF 1260	(20~50)°C	$U=0.3^{\circ}\text{C}$		
		Humidity		20%RH~90%RH	$U=2.0\%RH$		
		Noise		(30~100)dB	$U=2.0\text{dB}$		
		Concentration of Oxygen		30%~40%	$U=2.1\%$		
2	*Medical Centrifuges	Rotate speed	Calibration Specification for Medical Centrifuges JJF 2004	(20~30000)r/min	$U_{rel}=0.2\%$		
		Time		(0~3600)s	$U=0.12\text{s}$		
		Temperature		(-40~85)°C	$U=0.12^{\circ}\text{C}$		
		Rate		(0~5)°C/min	$U=0.12^{\circ}\text{C}/\text{min}$		
		Noise		(35~100)dB(A)	$U_{rel}=1.3\%$		
3	*ELISA analytical Instruments	Wavelength	V.R.of ELISA Analytical Instruments JJG 861	(400~630) nm	$U=0.5\text{ nm}$		
		Absorbance		0.200~1.600	$U=0.006$		
				Channel difference:1.000	$U=0.006$		




No. CNAS L2336

第 154 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Sensitivity		≥ 0.01 L/mg	$U_{rel}=5\%$		
4	*Blood Cell Analyzers	Concentration	Verification Regulations of Blood Cell Analyzers JJG 714	RBC:(1.0~6.0) $\times 10^{12}$ /L	$U_{rel}=3\%$		
				WBC:(2.0~19.0) $\times 10^9$ /L	$U_{rel}=4\%$		
				PLT:(80~400) $\times 10^9$ /L	$U_{rel}=4\%$		
				HGB:(40~200)g/L	$U_{rel}=3\%$		
5	*Urine Analyzers	Concentration	Calibration Specification for Urine Analyzers JJF 1129	GLU:(0.1~42)mmol/L	$U_{rel}=3\%$		
				PRO:(0.1~2.0)g/L	$U_{rel}=5\%$		
				WBC:(30~200)个/ μ L	$U_{rel}=7\%$		
				RBC:(15~200)个/ μ L	$U_{rel}=7\%$		
		Density		1.005~1.025	$U=0.005$		
		pH		5.5~7.5	$U_{rel}=5\%$		
6	*Electrolyte Analyzers	Concentration	Verification Regulations of Electrolyte Analyzers JJG 1051	K ⁺ : (1.5~7.5)mmol/L	$U_{rel}=4\%$		
				Na ⁺ : (100~180)mmol/L	$U_{rel}=4\%$		
				Cl ⁻ : (80~160)mmol/L	$U_{rel}=4\%$		
				Li ⁺ : (0.40~2.00)mmol/L	$U_{rel}=5\%$		
				iCa ²⁺ : (0.5~2.5)mmol/L	$U_{rel}=5\%$		
7	*Automatic Closed Luminescence Immunoassay Analyzers	Concentration	Calibration Specification for Automatic Closed Luminescence Immunoassay Analyzers JJF 1752	(10~150)pmol/L	$U_{rel}=5\%$	Only measure human insulin	



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
8	*Portable Blood Glucose Meters	Concentration	Calibration Specification for Portable Blood Glucose Meters JJF 1383	(6~15)mmol/L	$U_{rel}=4\%$		
9	*Multifunction Patient Monitoring Instruments	Voltage	Verification Regulations of Multifunction Patient Monitoring Instruments JJG 1163	(0.5~2.0)mV	$U_{rel}=3\%$		
		Heart Rate		(30~300)次/分	$U_{rel}=1.3\%$		
		Pressure		(0~40)kPa	$U=0.2\text{kPa}$		
		Pulse Rate		(30~250)次/分	$U=1.4\text{次/分}$		
10	*Syringe Pumps and Infusion Pumps	Flow	Calibration Specification for Syringe Pumps and Infusion Pumps JJF 1259	(5~1000)mL/h	$U_{rel}=2.5\%$		
		Pressure		(60~180)kPa	$U=2.0\text{kPa}$		
11	*Medical Liquid Microflow Syringe Pumps	Flow	Calibration Specification for Medical Liquid Microflow Syringe Pumps JJF 2263	(0.1~100)mL/h	$U_{rel}=4\%$		
		Pressure		(60~180)kPa	$U=2.0\text{kPa}$		
12	*Glycohemoglobin Analyzers	Concentration	Calibration Specification for Glycohemoglobin Analyzers JJF 1841	(4.5~10.5)%	$U_{rel}=5\%$		
13	*Flow Cytometers	Concentration	Calibration Specification for Flow Cytometers JJF 1665	(60~80) %	$U=7\%$		
14	*Automatic Urinary Sediment Analyzers	Concentration	Calibration Specification for Automatic Urinary Sediment Analyzers JJF 1823	RBC:(150~2000)/ μL	$U_{rel}=11\%$		
				WBC:(150~2000)/ μL	$U_{rel}=11\%$		
15	*Genetic Analyzers	Concentration	Calibration Specification for Genetic Analyzers JJF 1838	(0.02~0.3) $\mu\text{g/g}$	$U_{rel}=9\%$		
16	*Nucleic Acids Analyzers	Concentration	Calibration Specification for Nucleic Acids Analyzers JJF 1817	(0.1~40) ng/ μL	$U_{rel}=7\%$		
17	*Specific Protein Analyzers	Volume	Calibration Specification for Specific Protein Analyzers	(1~200) μL	$U=0.3\mu\text{L}$		




No. CNAS L2336

第 156 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Temperature	JJF 2116	(-10~50)°C	U=0.4°C		
		concentration		(1~150)mg/L	U _{rel} =10%		
18	*Hospital Plate Washer	volume	Calibration Specification for Hospital Plate Washer JJF(Liao) 144	(1~500) μ L	U _{rel} =6%		
19	*Fluorescence Meters of Nucleic Acid	Concentration	Calibration Specification for Fluorescence Meters of Nucleic Acid JJF(Jing) 3024	(10~100) ng/μ L	U _{rel} =8%		
20	* (Automatic) Nucleic Acids Extractors	Temperature	Calibration Specification for (Automatic) Nucleic Acids Extractors JJF 1874	(30~95) °C	U=0.5°C		
		Frequency		(0.1~500) Hz	U=0.6Hz		
		Volume		(1~1000) μ L	U=0.6 μ L		
		Recovery Rate		(80~100)%	U _{rel} =7%		
21	*Automatic Chemistry Analyzers	Absorbance	C.S. for Automatic Chemistry Analyzers JJF 1720	0.5、1.0	U=0.004		
		Concentration		ALT:(50~110)U/L	U _{rel} =5%		
				GLU:(6~15) mmol/L	U _{rel} =3%		
Petroleum and chemical specialized measuring instruments							
1	Wet Film Thickness Gauges	Length	Calibration Specification for Wet Film Thickness Gauges JJF 1484	The comb gauge: (5~100) μ m	U=1 μ m		
				The comb gauge : (100~3000) μ m	U=3 μ m		
				Wheel gauge : (0~125) μ m	U=1 μ m		
				Wheel gauge: (125~1500) μ m	U=4 μ m		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
2	Paint Film Scribers	Length	Calibration Specification for Paint Film Scribers JJF 2163	1 mm、2 mm、3 mm	$U=3 \mu\text{m}$		
		Angle		30°	$U=0.3^\circ$		
3	*Heat Distortion and Vicat Softening Temperature Apparatus	Temperature	Calibration Specification for Heat Distortion and Vicat Softening Temperature Apparatus JJF(Zhe) 1051	(20~300)°C	$U=0.3 \text{ }^\circ\text{C}$		
		Hesting Rate		50 °C/h、120 °C/h	$U=0.2 \text{ }^\circ\text{C/h}$		
		Length		(0~10)mm	$U=0.03 \text{ mm}$		
		Mass		(1~5000)g	$U=0.3 \text{ g}$		
4	*Pencil Hardness Tester	Angle	Calibration Specification for Pencil Hardness Testers JJF(SH) 007	45°	$U=0.06^\circ$		
		Mass		500 g、750 g、765 g、 1000 g	$U=0.7 \text{ g}$		
5	*Film Impact Testers	Mass	Calibration Specification for Film Impact Testers JJF(SH) 002	1000 g	$U=0.6 \text{ g}$		
		Length		100 mm、200 mm、300 mm、400 mm、500 mm	$U=0.1 \text{ mm}$		
6	*Rebound Resilience Machine for Vulcanized Rubber	Length	Calibration Specification for Rebound Resilience Machine for Vulcanized Rubber JJF(SH) 038	(2~300)mm	$U=0.05 \text{ mm}$		
		Mass		(1~1200)g	$U=0.06 \text{ g}$		
		Force		(50~500)N	$U_{\text{rel}}=0.4\%$		
7	*Film Pendulum Hardness Testers	Length	Calibration Specification for Film Pendulum Hardness Testers JJF(SH) 008	400 mm、408 mm、500 mm	$U=0.06 \text{ mm}$		
		Mass		120 g、200 g、500 g	$U=0.02 \text{ g}$		
		Time		(1~180)s	$U=0.4 \text{ s}$		




No. CNAS L2336

第 158 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
8	*Coating Bend Testers (Cylindrical Mandrel)	Length	Calibration Specification for Coating Bend Testers(Cylindrical Mandrel) JJF(SH) 035	2 mm、3 mm、4 mm、5 mm、6 mm、8 mm、10 mm、12 mm、16 mm、20 mm、25 mm、32 mm	U=0.004 mm		
9	*Coating Flexibility Testers	Length	Calibration Specification for Coating Flexibility Testers JJF(SH) 006	4 mm、5 mm、10 mm、15 mm	U=3 μm		
				1 mm、2 mm、3 mm	U=0.04 mm		
10	*Film Adhesion Testers (Circle-drawing Method)	Length	Calibration Specification for Film Adhesion Testers(Circle-drawing Method) JJF(SH) 036	10.50 mm	U=0.04 mm		
11	*Putty Film Flexibility Testers	Length	Calibration Specification for Putty Film Flexibility Testers JJF(SH) 003	25 mm、50 mm	U=0.1 mm		
				78 mm	U=0.2 mm		
12	Wedge Cut Method Film Thickness Instruments	Length	Calibration Specification for Wedge Cut Method Film Thickness Instruments JJF(Ji) 150	(0~10)mm	U=2 μm		
		Angle		(1~90)°	U=0.03°		
13	*Horizontal and Vertical Combustion Tester	Angle	Calibration Specification for Horizontal and Vertical Combustion Tester JJF(Gui) 76	(20~90)°	U=12'		
		Length		(5~120)mm	U=0.04 mm		
		Time		(5~600)s	U=0.5 s		
		Temperature		(100~650)°C	U=1.0 °C		
Special measuring instrument for building and traffic							
1	*Boiling testing box for soundness of the portland cement	Power	V.R.of Boiling testing box for soundness of the portland cement JJG(JC) 109	(0~4400)W	U=20W		
		Temperature		(20~100)°C	U=0.2°C		



No. CNAS L2336

第 159 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
2	*Standard Curing Chamber for Cement and Concrete	Temperature	C. S. for the Equipment of the Environmental Testing for Temperature and Humidity Parameters JJF 1101	(18~22)°C	U=0.2°C		
		Humidity		80%RH~95%RH	U=1.5%RH		
3	Rebound Test Hammer	Force	Verification Regulation of Rebound Test Hammer JJG 817	(0.4~0.8)N	U _{rel} =2.5%		
		Rigidity		(65~820)N/m	U _{rel} =2.6%		
		Rate		72~82	U=1		
		Length		Pointer length(19.8~20.2)mm	U=0.2mm		
Working length(61~76)mm	U=0.2mm						
4	Metrology Verification Device for Rebound Tester	Force	Verification Regulation of Metrology Verification Device for Rebound Tester JJG(Su) 59	(0.1~1)N	U=0.012N		
		Hardness		(58~62)HRC	U=1.8HRC		
		Mass		(1999~2001)g	U=0.18g		
		Length		Marking value of tension spring instrument(0~80)mm	U=0.1mm		
5	*Apparatus to Measure Water Permeability of Concrete	Pressure	Calibration Specifications for Apparatus to Measure Water Permeability of Concrete JJF 1812	(0.1~6)MPa	U=0.7%FS		
6	*Air Entrainment Meter of Freshly Mixed Concrete by The Volumetric Method	Pressure	Verification Regulation of Air Entrainment Meter of Freshly Mixed Concrete by The Volumetric Method JJG(JT) 094	(0~0.16)MPa	U=0.7%FS	合格评定国家认可委员会 认可证书专用章	
		Volume		(6~8)L	U=0.01L		
7	*Flow Table for	Time	Verification Regulation of	(24~26)s	U=0.24s		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Determine Cement Mortar Fluidity	Length	Flow Table for Determine Cement Mortar Fluidity JJG(JT) 096	Diameter and die size(59~301)mm	U=0.06mm		
				Fall distance(9.8~10.2)mm	U=0.01mm		
		Mass		(4200~4500)g	U=3g		
8	*Apparatus for Normal Consistency and Setting Time of Cement Paste	Mass	Verification Regulation of Apparatus for Normal Consistency and Setting Time of Cement Paste JJG(JT) 050	(290~310)g	U=12mg		
		Length		Ruler(0~80)mm	U=0.02mm		
9	*Mortar Consistency Tester	Mass	Calibration Specification for Nonmetal Building Materials Plastic Limit Measuring Instruments JJF 1090	(40~310)g	U=12mg		
		Length		(0~300)mm	U=0.05mm		
		Angle		(5~50)°	U=0.1°		
		Roughness		Ra (1~2) μm	U=0.3 μm		
10	*Consistency Testing Instrument for Concrete Mixing	Length	Verification Regulation of Consistency Testing Instrument for Concrete Mixing JJG(Su) 50	(1~301)mm	U=0.3mm		
		Mass		(2700~2800)g	U=0.6g		
				(8650~8750)g	U=0.6g		
		Frequency		(40~60)Hz	U=0.4Hz		
11	*Apparatus for Time of Setting of Concrete Mixture by Penetration Resistance	Length	Verification Regulation of Apparatus for Time of Setting of Concrete Mixture by Penetration Resistance JJG(JT) 095	Tube size(0~300)mm	U=0.1mm		
				Needle size(0~15)mm	U=0.005mm		
		Force		(0.1~2)kN	U _{rel} =0.5%		
12	*Mixer for Cement Paste	Rotate speed	C.S. for Mixer for Cement Paste JJF(JC) 104	(57~305)r/min	U=0.6r/min		



No. CNAS L2336

第 161 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time		(10~180)s	$U=0.33s$		
		Length		Gap(1.0~3.0)mm	$U=0.1mm$		
				Mixer size(0~161)mm	$U=0.1mm$		
13	*Cementmortar mixer complying with ISO 679	Rotate speed	Calibration Specification for Mixer for Mixing Mortars JJF(JC) 123	(50~300)r/min	$U=0.6r/min$		
		Time		(29~91)s	$U=0.3s$		
		Length		Gap(2~4)mm	$U=0.1mm$		
				Pan and Ye Chicun(0~203)mm	$U=0.1mm$		
14	*Cement mortars mixer	Rotate speed	Verification Regulation of cement mortars mixer JJG(JC) 102	(60~150)r/min	$U=0.6r/min$		
		Time		(0~5)min	$U=0.6s$		
		Length		Gap(1.0~2.0)mm	$U=0.05mm$		
				Mixer size(0.5~200)mm	$U=0.05mm$		
15	*Apparatus of Abrasion Resistance for Cement Mortar and Concrete	Force	Verification Regulation of Apparatus of Abrasion Resistance for Cement Mortar and Concrete JJG(JT) 097	(200~400)N	$U_{rel}=0.5\%$		
		Rotate speed		(15~650)r/min	$U=(0.14~0.5)r/min$		
		Length		Spindle stroke and flower wheel size(0~100)mm	$U=0.04mm$		
16	*Los Angeles Testing Machine	Mass	Verification Regulation of Los Angeles Testing Machine JJG(JT) 108	(300~5200)g	$U=0.3g$		
		Length		Roller size(95~715)mm	$U=0.48mm$		
				Diameter of steel ball(40~50)mm	$U=0.05mm$		
		Rotate speed		(30~33)r/min	$U=0.42r/min$		




No. CNAS L2336

第 162 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
17	*Jolting Table for Compacting Mortars Specimen	Mass	C.S. for cement Jolting Table for complying Mortar Specimen JJF(JC) 124	(13.5~20.5)kg	U=0.09kg		
		Time		(58~62)s	U=1.0s		
		Length		(14~801)mm	U=0.12mm		
		Amplitude(14.5~15.5)mm		U=0.01mm			
18	*Building Material Testing Machine of Constant Loading Speed	Force	Verification Regulation of Building Material Testing Machine of Constant Loading Speed JJG 1025	(1~5000)kN	U _{rel} =0.4%		
		Speed		(0.01~25)kN/s	U _{rel} =3.0%		
19	*Sand Equivalent Tester	Length	Verification Regulation of Sand Equivalent Tester JJG(JT) 137	Oscillating distance(202~204)mm	U=0.3mm		
				Other dimensions(0.1~421)mm	U=0.1mm		
		Oscillation frequency		(178~182)time/min	U=1.0time/min		
		Time		(29~1215)s	U=0.6s		
		Mass		(995~1005)g	U=1.2g		
20	*Cement Fineness Negative Pressure Screen Analyzers	Pressure	Calibration Specification for Cement Fineness Negative Pressure Screen Analyzers JJF 1827	(-100~0)hPa	U=0.7%FS		
		Rotate speed		(28~32)r/min	U _{rel} =0.4%		
		Noise		(30~130)dB(A)	U=1dB(A)		
		Insulation resistance		(10~30)MΩ	U=1MΩ		
21	*Marshall Test Machine for Bituminous	Speed	Verification Regulation of Marshall Test Machine for Bituminous Mixtures JJG(JT)	No load(45~55)mm/min	U _{rel} =0.5%		
				load(45~55)mm/min	U _{rel} =0.7%		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Mixtures	Force	066	(0.1~50)kN	$U_{rel}=0.4\%$		
		Length		(0~80)mm	$U=0.03\text{mm}$		
		Flow value		(0~20)mm	$U=0.02\text{mm}$		
22	*Compaction Instrument of Soil	Mass	Verification Regulation of Compaction Instrument of Soil JJG(JT) 058	(2000~5000)g	$U=0.04\text{g}$		
		Length		(1~60)mm	$U=0.06\text{mm}$		
				(290~460)mm	$U=0.4\text{mm}$		
23	*Asphalt Mixture's Marshall Compaction Test Apparatus	Mass	Verification Regulation of Asphalt Mixture's Marshall Compaction Test Apparatus JJG(JT) 065	(4000~15000)g	$U=2\text{g}$		
		Length		(50~500)mm	$U=1\text{mm}$		
		Speed		(50~70)times/min	$U=1\text{times/min}$		
24	*Instrument for Determining Location of Concrete Reinforced Bar	Length	Verification Regulation of Instrument for Determining Location of Concrete Reinforced Bar JJG(JT) 131	(0~500)mm	$U=0.7\text{mm}$		
		Diameter		(0~50)mm	$U=0.2\text{mm}$		
25	*Reinforced Concrete Covermeter and Floorslab Thickness Tester	Length	Calibration Specification for Reinforced Concrete Covermeter and Floorslab Thickness Tester JJF 1224	(0~500)mm	$U=0.7\text{mm}$		
		Diameter		(0~50)mm	$U=0.2\text{mm}$		
26	*Liquid & Plastic Limit Gauge	Mass	Verification Regulation of Liquid & Plastic Limit Gauge JJG(JT) 069	(70~110)g	$U=12\text{mg}$		
		Length		(0~60)mm	$U=0.02\text{mm}$		
		Angle		(25~35)°	$U=0.1^\circ$		
		Time		(3~6)s	$U=0.04\text{s}$		



No. CNAS L2336

第 164 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
27	*Determination Specific Surface Area of Air Permeability Apparatus	Length	Verification Regulation of Determination Specific Surface Area of Air Permeability Apparatus JJG(JC) 107	(2.7~65)mm	$U=(0.1\sim 1.0)\text{mm}$		
		Time		(0~900)s	$U=0.3\text{s}$		
		Volume		(1.5~2.5) cm^3	$U_{\text{rel}}=1.0\%$		
28	*Accelerated Polishing Machine	Length	Verification Regulation of Accelerated Polishing Machine JJG(JT) 054	(43~46)mm	$U=0.02\text{mm}$		
				(204~407)mm	$U=0.1\text{mm}$		
		Rotate speed		(315~325)r/min	$U=1.2\text{r/min}$		
		Hardness		(60~72)HA	$U=1.0\text{HA}$		
		Rate		Sand injection(2.5~35)g/min	$U=(0.1\sim 1.5)\text{g/min}$		
				Water injection (50~75)mL/min	$U=2.2\text{mL/min}$		
Force	(705~735)N	$U=1.2\text{N}$					
29	Le Chatelier Flask for Determining Density of Hydraulic Cement	Volume	Verification Regulation of Le Chatelier Flask for Determining Density of Hydraulic Cement JJG(JT) 092	(0.01~24)mL	$U_{\text{rel}}=2.0\%$		
30	*Flow Time Tester for Fine Aggregate	Angle	Verification Regulation of Flow Time Tester for Fine Aggregate JJG(JT) 109	(55~65) $^{\circ}$	$U=0.2^{\circ}$		
		Length		(10~130)mm	$U=0.05\text{mm}$		
31	*Le Chatelier	Length	Verification Regulation of Le Chatelier, Tester for Determining Expansion of Le Chatelier Needles JJG(JT) 093	(0~25)mm	$U=0.60\text{mm}$		
				(0.1~160)mm	$U=0.04\text{mm}$		



No. CNAS L2336

第 165 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
32	*Needle and Flake Gages	Length	Calibration Specification for Needle and Flake Gages JJF1593	(2~90)mm	U=0.04mm		
33	*Testing Instrument for Crushing Value	Length	Verification Regulation of Testing Instrument for Crushing Value JJG(Su) 54	(8~160)mm	U=0.06mm		
34	*The Penetration Test Apparatus of Bituminous	Length	Verification Regulation of the Penetration Test Apparatus of Bituminous JJG(JT) 067	Indication(0~50)mm	U=0.02mm		
				Standard needle size(0~55)mm	U=0.02mm		
				Mass(40~210)g	U=0.02g		
				(0.5~55) °C	U=0.04°C		
				(5~15)°	U=0.2°		
		Time		(3~65)s	U=0.3s		
35	*Bituminous materials' determining ductility	Stretching rate	Verification Regulation of Bituminous Materials Determining Ductility JJG(JT) 023	(9~11)mm/min	U=0.1mm/min		
				(40~60)mm/min	U=0.3mm/min		
		Temperature		(0~30)°C	U=0.04°C		
		Length		(0~1500)mm	U= (0.03~0.3) mm		
36	*Bitumen softening meter	Temperature	Calibration Code for bitumen softening meter SL414	(0~200)°C	U=0.2°C		
		Length		(0~30)mm	U=0.04mm		
		Mass		(2~5)g	U=0.02g		
		Speed		(1~10)°C/min	U=0.2°C/min		
37	*Eight-wheel Continuous Vialog	Length	Verification Regulation of Eight-wheel Continuous	Indication(-20~20)mm	U=0.05mm		



No. CNAS L2336

第 166 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Hardness	Vialog JJG(JT) 024	Diameter(159~161)mm	$U=0.05\text{mm}$		
				Datum length(2970~3030)mm	$U=10\text{mm}$		
				(55~70)HA	$U=0.6\text{HA}$		
38	*Benkelman Beam Pavement Deflectometer	Length	Verification Regulation of Benkelman Beam Pavement Deflectometer JJG(JT) 025	(0~300)mm	$U=(0.005\sim0.05)\text{mm}$		
39	*Pavement Texture Depth Manual Sanding Device	Length	Verification Regulation of Pavement Texture Depth Manual Sanding Device JJG(JT) 117	(1~100)mm	$U=0.02\text{mm}$		
		Roughness		Ra(0.01~100) μm	$U=1\ \mu\text{m}$		
40	Measuring Instrument for Cement Bright Degumming Equipment	Acceleration sensitivity	Verification Regulation of Measuring Instrument for Cement Bright Degumming Equipment JJG 974	(1~100)m/s ² (40~100)Hz	$U_{\text{rel}}=1.2\%$		
		Frequency		(20~2000)Hz	$U_{\text{rel}}=0.3\%$		
		Amplitude		(0.01~10)mm	$U_{\text{rel}}=1.0\%$		
		Rotate speed		(15~20000)r/min	$U_{\text{rel}}=0.3\%$		
		Time		(0~5)min	$U=0.2\text{s}$		
41	*Test Apparatus for Theoretical Maximum Specific Gravity of Asphalt Mixtures	Pressure	V.R. of Test Apparatus for Theoretical Maximum Specific Gravity of Asphalt Mixtures JJG(JT) 105	(1~120)kPa	$U=0.7\%\text{FS}$		
		Vacuum		(-100~-0.1)kPa	$U=0.7\%\text{FS}$		
		Volume		(2000~5000)mL	$U=(4\sim8)\text{mL}$		
		Time		(0~2)min	$U=0.2\text{s}$		
42	*Dynamic penetration tester	Mass	V.R. of Dynamic Penetration Tester JJG(JT)169	(0~15)kg	$U=0.04\text{kg}$		
				(60~65)kg	$U=0.2\text{kg}$		



No. CNAS L2336

第 167 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Length		(0~800)mm	U=0.2mm		
		Angle		(9~16)°	U=0.2°		
43	*California Bearing Ratio Instrument	Mass	Verification Regulation of California Bearing Ratio Instrument JJG(JT) 106	(1200~1300)g	U=0.5g		
		Length		(0~200)mm	U=0.05mm		
		Force		(0.1~50)kN	U _{rel} =0.3%		
		Speed		(0.5~1.5)mm/min	U _{rel} =3.0%		
44	*Waterproof Roll Material Impervious Instrument	Pressure	Calibration Specification for Waterproof Roll Material Impervious Instrument JJF(Jin)03	(0~1)MPa	U=0.6%FS		
45	*Rebar Gauge Punching Machine	Length	Verification Regulation of Rebar Gauge Punching Machine JJG (JT) 158	(0~400)mm	U=0.05mm		
46	*Sand-cone Density Apparatus	Length	Verification Regulation of Sand-cone Density Apparatus JJG(JT) 120	(0~400)mm	U=0.05mm		
47	*Annular sword	Length	Code for calibration of annular sword SL 110	(0~100)mm	U=0.05mm		
		Angle		(9~16)°	U=0.1°		
		Roughness		Ra(0~4) μ m	U=0.3 μ m		
48	*Carbonization Depth Measuring Instruments and Calipers	Length	Calibration Specification for Carbonization Depth Measuring Instruments and Calipers JJF 1721	Measuring Rule(0~8)mm	U=0.08mm		
				Measuring instrument(0~26)mm	U=0.01mm		
49	*Relative density meter	Length	V.R. of Relative density meter JJD1021	(40~160)mm	U=0.05mm		




在线扫码获取验证

No. CNAS L2336

第 168 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Mass		(1200~1300)g	$U=1g$		
50	*Mechanical Sieve Shaker	Length	Calibration code for mechanical sieve shaker SL411	(8~14)mm	$U=0.1mm$		
		Frequency		(100~300)BPM	$U=5BPM$		
51	Pile Static Load Measuring Instrument	Force	C.S. for Pile Static Load Measuring Instrument JJF(Hei)01	(0.1~5000)kN	$U_{rel}=0.4%$		
		Pressure		(0~70)MPa	$U=0.6\%FS$		
		Length		(0~50)mm	$U=0.02mm$		
52	*Steel Bar Tarnish Measuring Instruments	DC Voltage	Verification Regulation of Steel Bar Tarnish Measuring Instruments JJF 1341	(0.01~10)V	$U_{rel}=0.1\%$		
		DC Current		(0.001~2)A	$U_{rel}=0.2\%$		
53	*Cement Le Chatelier Needles	Length	Calibration Specification for the Expansion Tester of Cement Le Chatelier Needles JJF(JC)110	(0~25)mm	$U=0.04mm$		
		Mass		(295~305)g	$U=0.02g$		
54	*Vibrating Talle for Concrete Testing	Length	V.R. of Vibrating Talle for Concrete Testing JJG(Su)60	(0.33~0.52)mm	$U=0.012mm$		
		Force		(0.1~1000)N	$U=2N$		
		Frequency		(47~53)Hz	$U=0.2Hz$		
		Time		(0~5)s	$U=0.01s$		
55	*Methylene Blue Mixer	Rotational speed	Calibration Specification for Methylene Blue Mixer JJF(Gui)82	(100~900)r/min	$U_{rel}=0.3\%$		
		Length		(65~85)mm	$U=0.04mm$		
		Time		(1~600)s	$U_{rel}=0.5\%$		
56	*Permeameter	Stress	Calibration Specification for Permeameter JJF(Zhe)1165	(0~6)MPa	$U=0.10\%FS$		




No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
57	*Mortar Setting Time Meter	Force	Calibration Specification for Mortar Setting Time Meter JJF(JC) 208	(1~100) N	$U_{rel}=0.2\%$		
58	*Geosynthetic Hydrostatic Testers	Stress	Calibration Specification for Geosynthetic Hydrostatic Testers JJF(Ji) 178	(0~2.5) MPa	$U=0.003\text{MPa}$		
		Length		(2~4) mm	$U=0.03\text{mm}$		
				(190~210) mm	$U=0.04\text{mm}$		
59	*impermeability testing instrument of geomembranes	Length	Calibrating code for impermeability testing instrument of geomembranes SL408	(2~4) mm	$U=0.03\text{mm}$		
		Length		(160~350) mm	$U=0.03\text{mm}$		
		Stress		(0~4) MPa	$U=0.3\%\text{FS}$		
60	*Tester for Separating Emulsified Bitumen Particle Ions	Length	V. R. of Tester for Separating Emulsified Bitumen Particle Ions JJG(JT) 115	(0.5~150) mm	$U=0.04\text{mm}$		
		Voltage		(5~7) V	$U=0.006\text{V}$		
61	*Measurer for Liner Removal Performance of Retroreflective Sheeting	Length	Verification Regulation of Measurer for Liner Removal Performance of Retroreflective Sheeting JJG(JT) 085	20 mm、25 mm、63 mm、80 mm、150 mm	$U=0.04\text{ mm}$		
		Mass		6600 g	$U=1.6\text{ g}$		
62	*Asphalt Mixture's Mixing Machine	Volume	Verification Regulation of Asphalt Mixture's Mixing Machine JJG(JT) 064	(10~15)L	$U=0.006\text{ L}$		
		Rotating velocity		45 r/min、75 r/min	$U_{rel}=0.2\%$		
		Time		(10~1800) s	$U=0.6\text{ s}$		
		Temperature		(20~250)°C	$U=0.3\text{ }^{\circ}\text{C}$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
63	*Concrete Resistivity Meter	Resistivity	Verification Regulation of Concrete Resistivity Meter JJG(JT) 159	10.4k $\Omega \cdot \text{cm}$ 、16.0k $\Omega \cdot \text{cm}$ 、31.4k $\Omega \cdot \text{cm}$ 、62.8k $\Omega \cdot \text{cm}$ 、314.0k $\Omega \cdot \text{cm}$	$U_{\text{rel}}=0.6\%$		
		DC resistance		100 $\Omega \sim 15 \text{ k}\Omega$	$U_{\text{rel}}=0.06\%$		
64	*Sliding Line Resistance Displacement Meter	Length	Verification Regulation of Sliding Line Resistance Displacement Meter JJG(JT) 040	(0~300)mm	$U=0.03 \text{ mm}$		
65	*Vibration Wire Multipoint Displacement Meter	Length	Verification Regulation of Vibration Wire Multipoint Displacement Meter JJG(JT) 039	(0~300)mm	$U=0.03 \text{ mm}$		
66	*Shape Selector of Glass Beads	Vibration displacement	Verification Regulation for Shape Selector of Glass Beads JJG(JT) 073	(0.1~0.5)mm, 50Hz	$U=0.03 \text{ mm}$		
		Length		150 mm、380 mm	$U=0.2 \text{ mm}$		
		Angle		(0~5) $^{\circ}$	$U=0.05^{\circ}$		
67	Container	Length	Verification Regulation of Container JJG(JT) 191	108 mm、109 mm、155 mm、160 mm、186 mm、204 mm、208 mm、255 mm、294 mm、295 mm、360 mm、467 mm	$U=0.07 \text{ mm}$		
		Capacity		1 L、3 L、5 L、10 L、20 L、30 L、80 L	$U_{\text{rel}}=0.3\%$		
68	*Rebar Bending Testing Machines	Angle	Calibration Specification for Rebar Bending Testing Machines JJF(E) 77	Horizontal:(0~360) $^{\circ}$	$U=0.010^{\circ}$		
				Vertical:(-90~90) $^{\circ}$	$U=0.047^{\circ}$		



在线扫码获取验证

No. CNAS L2336

第 171 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
69	*Flexibility Test Device of Plastic Bellows	Length	Verification Regulation of Flexibility Test Device of Plastic Bellows JJG(JT) 164	(8~150)mm	$U=0.05$ mm		
				1500 mm、1800 mm	$U=0.5$ mm		
		Angle		90°	$U=0.08$ °		
70	*Measurer for Flexibility of Retroreflective Sheeting	Length	Verification Regulation of Measurer for Flexibility of Retroreflective Sheeting JJG(JT) 098	3.20 mm	$U=0.003$ mm		
71	Standard Testing Sieve for Cement Fineness(80 μ m)	Length	Verification Regulation of Standard Testing Sieve for Cement Fineness(80 μ m) JJG(JT) 049	56 μ m、(75~85) μ m	$U=4$ μ m		
				10 mm、25 mm、38 mm、50 mm、80 mm、90 mm、125 mm、135 mm、140 mm、142 mm、150 mm、160 mm	$U=0.2$ mm		
72	Standard Block of Reinforced Covermeter and Floorslab Thickness Tester	Length	Calibration Specification for Standard Block of Reinforced Covermeter and Floorslab Thickness Tester JJF(Jin) 76	(1~600)mm	$U=0.05$ mm		
73	*Cement Concrete Mixers	Rotating velocity	Verification Regulation of Cement Concrete Mixers JJG(JT) 187	48 r/min、55 r/min	$U_{rel}=0.2\%$		
		Time		120 s	$U_{rel}=0.2\%$		
74	*Ultrasonic Hole Forming Quality Detector	Length	Verification Regulation of the Ultrasonic Hole Forming Quality Detector JJG(JT) 171	0.5 m、0.6 m、0.8 m、1.0 m、1.4 m、2.0 m	$U_{rel}=0.6\%$		
75	Servo Inclinometer	Angle	Verification Regulation of Servo Inclinometer JJG(JT) 038	(-20~20)°	$U=6''$		



No. CNAS L2336

第 172 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
76	Borehole Clinometers	Angle	Calibration Specification for Borehole Clinometers JJF 1550	(-30~30)°	U=6"		
77	*Automated Rut-Depth Measuring System	Length	Verification Regulation of Automated Rut-Depth Measuring System JJG(JT) 051	1 mm、5 mm、10 mm、20 mm、50 mm、100 mm、200 mm	U=0.5 mm		
78	*Load Wheel Tester for Asphalt Emulsion Slurry Mixture	Frequency	Verification Regulation of Load Wheel Tester for Asphalt Emulsion Slurry Mixture JJG(JT) 091	44 min ⁻¹	U=1 min ⁻¹		
		Mass		56.7 kg	U=0.10 kg		
		Hardness		65 HA	U=1.2 HA		
		Length		3.2 mm、6.4 mm、12.7 mm、26.0 mm、38.0 mm、50.0 mm、76.5 mm、152 mm、355 mm、380 mm	U=0.06 mm		
79	*Emulsified Bitumen Stability Test-tube	Length	Verification Regulation of the Emulsified Bitumen Stability Test-tube JJG(JT) 116	12mm、32mm	U=0.018mm		
				30mm、50mm、62mm、174mm	U=0.05mm		
				310mm	U=0.5mm		
80	Vibrating Wire Rebar Stress Meter	Force	Verification Regulation of Vibrating Wire Rebar Stress Meter JJG(JT) 035	(10~600)kN	U _{rel} =0.8%		
81	*Testing Instrument for Crushing Value	Length	Verification Regulation of Testing Instrument for Crushing Value JJG(JT) 190	(10~150)mm	U=0.04 mm		
82	*Mortar layering Tester	Length	Calibration Specification of Mortar layering Tester JJF(Xiang) 36	100 mm、150 mm、200 mm	U=0.04 mm		

No. CNAS L2336


第 173 页 共 187 页




在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
83	*Measurer for Resistance to Impact of Raised Pavement Markers	Length	Verification Regulation of Measurer for Resistance to Impact of Raised Pavement Markers JJG(JT) 080	1000 mm	U=0.3 mm		
		Mass		1040 g	U=0.2 g		
84	*Consistency Tester for Asphalt Emulsion Slurry Mixture	Length	Verification Regulation of Consistency Tester for Asphalt Emulsion Slurry Mixture JJG(JT) 114	38 mm、76 mm、89 mm、109 mm、129 mm、149 mm、169 mm、189 mm、209 mm	U=0.04 mm		
85	*Rapid Freezing and Thawing Test Machine for Concrete	Temperature	Calibration Specification of Rapid Freezing and Thawing Test Machine for Concrete JJF(Jin) 45	(-18~5)°C	U=0.15 °C		
86	*Standard Viscometer of Bituminous Materials	Temperature	Verification Regulation of Standard Viscometer of Bituminous Materials JJG(JT) 055	40 °C、60 °C	U=0.08 °C		
		Length		6.35 mm、12.70 mm	U=4 μm		
				5.0 mm、90.3 mm、92.0 mm	U=0.05 mm		
		Capacity		25 mL、50 mL、75 mL、100 mL	U=0.8 mL		
87	*Asphalt Mixture Rut-depth Testing Instrument	Temperature	Calibration Specification for Asphalt Mixture Rut-depth Testing Instrument JJF(Zhe) 1094	60 °C	U=0.15 °C		
		Length		Deformation measurement device: (0~30)mm	U=3 μm		
				Size: 15 mm、50 mm、200 mm	U=0.05 mm		
				230 mm	U=0.8 mm		
		Hardness		78 IRHD	U=0.4 IRHD		
		Frequency		42 min ⁻¹	U=0.3 min ⁻¹		




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pressure		0.70 MPa	$U=0.005$ MPa		
88	*Measurer for Impact Resistance of Retroreflective Sheeting	Length	V.R. of Measurer for Impact Resistance of Retroreflective Sheeting JJG(JT) 084	54 mm、60 mm、250 mm	$U=0.05$ mm		
		Mass		450 g	$U=0.009$ g		
89	*Digital Sticking Strength Testers	Force	Calibration Specification for Digital Sticking Strength Testers JJF(Ji) 224	(1~10)kN	$U_{rel}=0.3\%$		
		Length		8 mm、40 mm、45 mm、50mm、95 mm	$U=0.05$ mm		
90	*Unconfined Thickness Measurement Instrument	Length	Verification Regulation of Unconfined Thickness Measurement Instrument JJG(JT) 177	1 mm、2 mm、3 mm、4 mm、5 mm、6 mm、7 mm、8 mm、9 mm、10 mm	$U=4$ μm		
		Area		(50~2600)mm ²	$U=7$ mm ²		
		Force		100 cN、200 cN	$U=0.004$ cN		
		Time		10 s、30 s	$U=0.2$ s		
91	Wedge-Shape Filler Gauges	Length	Calibration Specification for Wedge-Shape Filler Gauges JJF 1548	(0~60)mm	$U=0.01$ mm		
92	Test Sieves	Length	Calibration Specification for Test Sieves JJF 1175	(0.04~4)mm	$U=2$ μm		
				(4~125)mm	$U=0.04$ mm		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
93	*Molds	Length	Calibration Specification for Molds JJF 1307	40.0 mm、40.1 mm、100 mm、150 mm、160 mm、175 mm、185 mm、200 mm、300 mm、400 mm、450 mm、515 mm、550 mm、600 mm	U=0.04 mm		
94	Multifunction Slope Scale	Angle	Calibration Specification for Multifunction Slope Scale JJF(Ji) 140	(-90~90)°	U=0.3°		
95	*Insulating Glass Dew Point Analyzer	Temperature	Calibration Specification for Insulating Glass Dew Point Analyzer JJF(Min) 1112	semiconductor type: (-60~0)°C	U=0.3°C	Dry ice refrigeration type equipment is not calibrated on-site	
				dry ice cooling type: (-60~0)°C	U=0.2°C		
96	*Road Traffic Marking Thickness Measuring Instrument	Length	V.R.of Road Traffic Marking Thickness Measuring Instrument JJG(JT) 203	pointer type: (0.5~10)mm	U=0.005mm		
				displayed: (0.5~12)mm	U=0.01mm		
97	Road Quality Test Ruler	Parallelism	C.S. for Road Quality Test Ruler JJF(Xin) 67	(0~5)mm/3m	U=0.06mm/3m		
98	Construction Quality Tester Sets	angle	C.S.for Construction Quality Tester Sets JJF 1110	verticality measuring ruler: (0~15)mm/2m	U=0.2mm/2m		
				Inside and outside measuring square: (0~7)mm/150mm	U=0.2mm/150mm		
		Length		wedge-shaped feeler gauge: (0~15)mm	U=0.06mm		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				bai grid network: (100~250)mm	$U_{rel}=0.1\%$		
				diagonal inspection ruler: (-100~100)mm	$U=0.2\text{mm}$		
99	Cement Length Comparator	Length	C.S. for Cement Length Comparator JJF(Su) 217	(0~10)mm	$U=5\ \mu\text{m}$		
				(156~540)mm	$U=6\ \mu\text{m}$		
100	*Crack Width Measuring Instrument	Length	Verification Regulation of Crack Width Measuring Instrument JJG(JT) 135	(0.02~10)mm	$U=6\ \mu\text{m}$		
101	*Measurer for Adhesion Performance of Retroreflective Sheeting	Length	V.R. of Measurer for Adhesion Performance of Retroreflective Sheeting JJG(JT) 083	(35~500)mm	$U=0.07\text{mm}$		
		Mass		(790~810)g	$U=0.02\text{g}$		
102	*Load Cell of static loading test of foundation pile with self-balanced method	Force value	C.S. for Load Cell of static loading test of foundation pile with self-balanced method JJF(Zhe) 1159	(10~600)kN	$U_{rel}=0.3\%$		
				(600~10000)kN	$U_{rel}=0.5\%$		
103	Test-tube for Determining Density of Asphalt	Length	Verification Regulation of Test-tube for Determining Density of Asphalt JJG(JT) 119	(1~72)mm	$U=0.04\text{mm}$		
		Mass		(1~40)g	$U=0.003\text{g}$		
		Capacity		(20~30)mL	$U=0.003\text{mL}$		
104	Steel Anvils	Mass	Calibration Specification for Steel Anvils JJF(Min) 1133	(15~46)kg	$U=0.12\text{kg}$		
		Length		(31~200)mm	$U=0.05\text{mm}$		
		Hardness		(58~62)HRC	$U=1.8\text{HRC}$		
105	*Testing	Force value	C.S. for Testing Apparatus	(1~300)kN	$U_{rel}=0.2\%$		




No. CNAS L2336

第 177 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Apparatus for Pavement Material Strength	Length	for Pavement Material Strength JJF(Wan) 108	Displacement : (0.5~50)mm	$U_{rel}=0.03\%$		
				Work schedule: (60~500)mm	$U=0.5\text{mm}$		
		Rate		(0.9~55)mm/min	$U_{rel}=0.4\%$		
106	*Hydraulic Static Pile Pressing Machine	Force value	C.S. for Hydraulic Static Pile Pressing Machine JJF(Zhe) 1158	(10~600)kN	$U_{rel}=0.3\%$		
				(600~10000)kN	$U_{rel}=0.5\%$		
Special measuring instruments for civil aviation and aviation							
1	PRO360/PRO3600 /DP-45 Protactor	Angle	C.S.for PRO360/ PRO3600 / DP-45 Protactor JJF(MH)0103	(0~90)°	$U=0.01^\circ$		
2	Dynamometer of The Engine Bootstrap Equipment	Force	C.S for Dynamometer of The Engine Bootstrap Equipment JJF(MH)0122	(0.1~200)kN	$U=0.2\%FS$		
3	Cable Tensiometer of Direct Reading Type	Force	C.S.for Cable Tensiometer of Direct Reading Type JJF(MH)0087	(0.01~5)kN	$U_{rel}=1.0\%$		
Special measuring instruments for electrical and electronic appliances							
1	*Single Insulated Wire and Cable Vertical burning Test Device	Temperature	Verification Procedure for Test Equipment of Rubber Plastic Wire and Cable -Part 5:Single Insulated Wire and Cable Vertical Burning Test Device JB/T4278.5	(0~1300) °C	$U=0.8^\circ\text{C}$		
		Time		(0~60) s	$U=0.5\text{s}$		
		Length		(0~150) mm	$U=0.5\text{mm}$		
2	*Glow-wire Apparatus	Temperature	C. S. for Glow-wire Apparatus JJF(Zhe)1050	(30~1000)°C	$U=1.6^\circ\text{C}$		




No. CNAS L2336

第 178 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time		(0~60)s	$U=0.5s$		
		Length		(0~150)mm	$U=0.05mm$		
		Force value		(0~10)N	$U=0.1N$		
3	*charge & Discharge of battery tester	Charge Voltage	Calibration Specification for Battery Charge & Discharge Testers JJF 2039	(0.01~1000) V	$U_{rel}=0.03\%$		
		Charge Current		10 μ A~200A	$U_{rel}=0.06\%$		
				200A~2000A	$U_{rel}=0.07\%$		
		Discharge Voltage		(0.01~1000) V	$U_{rel}=0.03\%$		
		Discharge Current		10 μ A~200A	$U_{rel}=0.06\%$		
				200A~2000A	$U_{rel}=0.07\%$		
		DC power		0.1W~100W	$U_{rel}=0.10\%$		
				100W~10kW	$U_{rel}=0.15\%$		
		Discharge capacity		0.1Ah~10Ah	$U_{rel}=0.09\%$		
				10Ah~1kAh	$U_{rel}=0.05\%$		
		Constant current charging current rise time		10 μ s~10ms	$U_{rel}=0.3\%$		
		Temperature		-20 $^{\circ}$ C~100 $^{\circ}$ C	$U=0.2^{\circ}$ C		
4	*Electrical safety analyzer	Insulation Resistance	Calibration specification for comprehensive safety tester JJF(DZ)0004	(1k Ω ~10M Ω)	$U_{rel}=0.3\%$		
				(10~100) M Ω	$U_{rel}=0.6\%$		



在线扫码获取验证

No. CNAS L2336

第 179 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA INTERNATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(100~1000) M Ω	$U_{rel}=1.3\%$		
				(1000~5000) M Ω	$U_{rel}=3\%$		
		Earth Continuity		(1~600) m Ω	$U_{rel}=0.6\%$		
		Earth Current		(1~40) A	$U_{rel}=0.3\%$		
		DC Voltage		(0.1~10) kV	$U_{rel}=0.7\%$		
		AC voltage		(0.01~6) kV (50Hz, 60Hz)	$U_{rel}=0.7\%$		
		DC current		(0.01~200) mA	$U_{rel}=0.8\%$		
		AC current		(0.01~200) mA (50Hz,60Hz)	$U_{rel}=0.8\%$		
Time	(0.01~999.99)s	$U=0.8\%T+0.03s$					
5	*Voltage Dips ,Short Interruptions and Voltage Variations Generator	Output voltage	C.S. for Voltage Dips ,Short Interruptions and Voltage Variations Generator JJF(DZ) 30802	(0.01~1)kV	$U_{rel}=1\%$		
		Time		0.01 s~50s	$U_{rel}=0.6\%$		
6	*Cable Tester	Cable DC resistance	Calibration for Cable Tester JJF1457	0.1 Ω ~ 1 Ω	$U_{rel}=2\%$		
				1 Ω ~ 10 Ω	$U_{rel}=1.2\%$		
				10 Ω ~ 100k Ω	$U_{rel}=1.0\%$		
		Insulation alarm resistance		1M Ω ~ 100M Ω	$U_{rel}=1.2\%$		
		100M Ω ~ 1000M Ω		$U_{rel}=1.5\%$			
		AC output voltage		(50~2000)V,(45Hz~65Hz)	$U_{rel}=1.3\%$		




在线扫码获取验证

No. CNAS L2336

第 180 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC output voltage Total harmonic distortion	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	0.1%~10%	$U_{rel}=3\%$		
		Capacitance between wires		100pF~100 μ F	$U_{rel}=0.3\%$		
		Leakage current		(0.01~20)mA,(45Hz~65Hz)	$U_{rel}=1.3\%$		
		Charging Voltage		50V~2000V	$U_{rel}=1.2\%$		
7	*Transformer On-load Tap-changer Testers	Transition time	Calibration Specification of Transformer On-load Tap-changer Testers JJF(Zhe) 1084	(10~300)ms	$U=0.15\%Tx+0.2ms$		
		resistance		(0.1~80)Ω	$U=0.8\%Rx+0.003\Omega$		
8	Electrostatic Discharge Simulators	Output voltage	C.S. for Electrostatic Discharge Simulators JJF 1397	(1~40) kV	$U_{rel}=0.8\%$		
		Contact discharge current		(1~100)A	$U_{rel}=6.6\%$		
		Rise Time		0.1 ns~100 ns	$U_{rel}=13\%$		
9	*Impulse Voltage Testers for Winding Interturn Insulation	Impulse Peak Voltage	C. S. for Impulse Voltage Testers for Winding Interturn Insulation JJF 1691	(0.01~15)kV	$U_{rel}=1.2\%$		
		Front time		(0.1~10) μ s	$U_{rel}=5\%$		
10	*Electrical Surge Generator	Output voltage	C.S. for Electrical Surge Generator JJF(DZ) 30803	(0.1~20) kV	$U_{rel}=0.8\%$		
		Half peak time		1 ns~1s	$U_{rel}=2.5\%$		
		Front Time		1 ns~1s	$U_{rel}=2.5\%$		




No. CNAS L2336

第 181 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Surge current peak		500A~8kA	$U_{rel}=2.5\%$		
11	*Current probes of Oscilloscope	DC Current	Calibration specification for Current probes of Oscilloscope JJF(DZ) 0036	0.01A~1000A	$U_{rel}=0.7\%$		
		AC Current		0.01A~1000A(40Hz~1kHz)	$U_{rel}=0.8\%$		
		Attenuation coefficient		1~1000	$U_{rel}=0.7\%$		
		Band bandwidth		50kHz~120MHz	$U_{rel}=4.2\%$		
		Rise Time		(3.6~100) ns	$U_{rel}=6\%$		
12	*Oscilloscope Voltage Probe	Attenuation	Calibration Specification for Oscilloscope Voltage Probe JJF 1437	1:1~1000:1	$U_{rel}=0.7\%$		
		Frequency Response		10kHz~300MHz	$U=0.8dB$		
		Rise time		(1~100)ns	$U_{rel}=6\%$		
		Input impedance		50 Ω ~ 100M Ω	$U_{rel}=0.04\%$		
13	*Higt Voltage probes of Oscilloscope	Attenuation ratio	Calibration specification for Differential probes of Oscilloscope JJF((DZ) 30306	1~1000	$U_{rel}=0.7\%$		
		frequency response		10Hz~300MHz	$U_{rel}=0.8\%$		
		Rise Time		(1.0~100) ns	$U_{rel}=6\%$		
		Input Resistance		50 Ω ~ 100M Ω	$U_{rel}=0.7\%$		
		CMRR		18dB~90dB	$U=1.9dB$		
14	*Higt Voltage probes of Oscilloscope	DC voltage attenuation ratio	Calibration specification for Higt Voltage probes of Oscilloscope JJF(DZ) 30304	DC:1~1000 (100V~20kV)	$U_{rel}=0.7\%$		



No. CNAS L2336

第 182 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC voltage attenuation ratio		AC:1~1000 (50V~20kV, 45Hz~65Hz)	$U_{rel}=0.8\%$		
15	*Impulse Voltage Tester with Respect to Eareh	Open circuit voltage peak	Calibration Specification for Impulse Voltage Tester with Respect to Eareh JJF(Zhe)1069	0.05kV~20kV	$U_{rel}=3\%$		
		pulse wavefront time		0.5 μ s ~ 20 μ s	$U_{rel}=5\%$		
		Half wave time		20 μ s ~ 1ms	$U_{rel}=4\%$		
16	*Electrical Fast Transient Generator	Impulse Peak Voltage	C.S.for Electrical Fast Transient Generator JJF(DZ)30804	(0.01~5)kV	$U_{rel}=3\%$		
		Burst Duration		(0.1~200)ms	$U_{rel}=1.5\%$		
		Rise time of pulse voltage		(2~10)ns	$U_{rel}=4.3\%$		
		Pulse voltage duration		(10~200)ns	$U_{rel}=4\%$		
		Pulse repetition rate		0.1kHz~1MHz	$U_{rel}=0.7\%$		
		Pulse group period		10ms~1000ms	$U_{rel}=0.6\%$		
17	*Data acquisition device	DC Voltage	Calibration Specification for Multi-Parameter Data Acquisition Instrument JJF (TX) 041	(10~100)mV	$U_{rel}=9.0 \times 10^{-5}$		
				(100mV~1V)	$U_{rel}=2.0 \times 10^{-5}$		
				(1~10)V	$U_{rel}=1.8 \times 10^{-5}$		
				(10~1000)V	$U_{rel}=3.0 \times 10^{-5}$		
		AC Voltage		(10~100)mV	$U_{rel}=8.6 \times 10^{-4}$		




No. CNAS L2336

第 183 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				100mV~1V(45Hz~10kHz)	$U_{rel}=2.6 \times 10^{-4}$		
				1V~10V(45Hz~10kHz)	$U_{rel}=2.2 \times 10^{-4}$		
				10V~100V(45Hz~10kHz)	$U_{rel}=2.4 \times 10^{-4}$		
				100~1000V(45Hz~10kHz)	$U_{rel}=2.4 \times 10^{-4}$		
		DC Current		(100μA~1mA)	$U_{rel}=6.5 \times 10^{-4}$		
				(1~100)mA	$U_{rel}=2.0 \times 10^{-4}$		
				(100mA~3A)	$U_{rel}=4.0 \times 10^{-4}$		
				(3~20)A	$U_{rel}=1.0 \times 10^{-3}$		
		AC Current		100μA~1mA(45Hz~5kHz)	$U_{rel}=2.6 \times 10^{-3}$		
				1mA~100mA(45Hz~5kHz)	$U_{rel}=7.0 \times 10^{-4}$		
				100mA~3A(45Hz~5kHz)	$U_{rel}=6.0 \times 10^{-4}$		
				3A~20A(45Hz~5kHz)	$U_{rel}=1.2 \times 10^{-3}$		
		DC Resistance		(10~100)Ω	$U_{rel}=7.8 \times 10^{-5}$		
				100Ω~10kΩ	$U_{rel}=3.2 \times 10^{-5}$		
				10kΩ~1MΩ	$U_{rel}=4.2 \times 10^{-5}$		
				(1~100)MΩ	$U_{rel}=5.0 \times 10^{-4}$		
		Temperature (TC)		(-100~1300)℃	$U=0.3^{\circ}\text{C}$		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Temperature (RTD)		(0~800) °C	$U_{rel}=0.2\%$		
		Average relative frequency		10Hz~10MHz	$U_{rel}=1.4 \times 10^{-5}$		
18	*Spark Testers	AC Voltage	C.S. for Spark Testers JJF 2239	AC: (0.1kV~20kV) 50Hz/60Hz	$U_{rel}=0.7\%$		
		DC Voltage		DC: (0.1kV~30kV)	$U_{rel}=0.6\%$		
19	*Online Winding Temperature Rise Tester	Resistance	Calibration Specification For Online Winding Temperature Rise Tester JJF 1540	(0.1~1) Ω	$U_{rel}=0.6\%$		
				(1~10) Ω	$U_{rel}=0.07\%$		
				(10 Ω ~10k Ω)	$U_{rel}=0.06\%$		
20	*Battery Resistance Test Instrument	Resistance	Calibration Specification for Battery Internal Resistance Testers JJF 1620	1m Ω ~10m Ω	$U_{rel}=0.08\%$		
				10m Ω ~3k Ω	$U_{rel}=0.06\%$		
		Voltage		(0.1~10)V	$U_{rel}=0.009\%$		
				(10~800)V	$U_{rel}=0.006\%$		
21	*Proof Tracking Index Testers	Test voltage	Calibration Specification for Proof Tracking Index Testers JJF(Zhe) 1087	(0.1~750)V(45Hz~60Hz)	$U_{rel}=0.3\%$		
		Test current		(0.01~1)A(45Hz~60Hz)	$U_{rel}=0.2\%$		
		Time		(0.01~999.9)s	$U_{rel}=0.3\%$		
22	*DC Electronic Load	DC Voltage	C.S. for DC Electronic Load JJF 1462	0.1V~100V	$U_{rel}=0.02\%$		
				100V~1000V	$U_{rel}=0.04\%$		
		DC Current		0.01A~20A	$U_{rel}=0.05\%$		




No. CNAS L2336

第 185 页 共 187 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				20A~120A	$U_{rel}=0.07\%$		
				120A ~200A	$U_{rel}=0.12\%$		
		DC Resistance DC Power		1 Ω ~1k Ω	$U_{rel}=0.10\%$		
				1W ~100W	$U_{rel}=0.12\%$		
		Current rise and fall rate		100W ~2kW	$U_{rel}=0.10\%$		
				0.1mA/ μ s ~10A/ μ s	$U_{rel}=1.6\%$		
		Current loading time and unloading time		10 μ s ~10s	$U_{rel}=0.7\%$		
		Constant voltage mode load regulation rate		0.01%~10%	$U_{rel}=0.04\%$		
Constant current mode load regulation rate	0.01%~10%	$U_{rel}=0.12\%$					
23	*Loop tester	loops	C.S. for Coil Number Testing Instructing JJF(Zhe)1065	(1~11110) loops	$U_{rel}=0.3\%$		
24	9kHz~30MHz whip antenna	Antenna coefficient	Calibration Specification of 9kHz~30MHz whip antenna JJF 1706	-30dB/m~70dB/m (9kHz~30MHz)	$U=1.2$ dB		
25	*Triple-Loop Antenna Systems	Validation Factor	Calibration Specification of Triple-Loop Antenna Systems JJF (DZ) 0035-2	(20~100) dB Ω (9kHz~30MHz)	$U=1.3$ dB		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
26	Current Probes and Injection Probes	Insertion loss	Calibration Specification of Current Probes and Injection Probes JJF(TX) 030	(0~60) dB (9kHz~30MHz)	U=1dB		
		Transfer impedance		(-40~20) dB Ω (10Hz~1GHz)	U=1dB		
		Transfer admittance		(-26~34) dBS (10Hz~1GHz)	U=1dB		



No. CNAS L2336

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证